



GREAT LAKES SCALE COMPANY  
 29540 CALAHAN RD.  
 ROSEVILLE, MI 48066  
 Phone: (586) 776-0517

## Certificate of Calibration

Customer: SUPERIOR MATERIALS 102  
 SUP1112 44922 GRAND RIVER AVE  
 NOV, MI 48375

Certificate ID: 105134 CEMENT180117024123

Test Date: 01/17/2018

ISO Number: MS-FM-CEM-001

(\*Make: ALKON Model: COMMAND SN#: 105134 CEMENT Capacity: 10000 lb Graduation: 5 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
01/17/2018			105134 CEMENT	III
SCALE LOCATION			EQUIPMENT CONDITIONS	
			Working	Dirty
SELECT NUMBER OF PM DAYS:		CAPACITY	10000 UOM lb	INCREMENT 5

### AS FOUND TEST USING TEST WEIGHTS

### AS LEFT TEST USING TEST WEIGHTS

Weight Applied	Reading	Error	Weight Applied	Reading	Error
0	0	0	0	0	0
6000	5985	-15	6000	5995	-5
0	0	0	0	0	0
			1000	995	-5
			2000	1995	-5
			3000	2995	-5
			4000	3995	-5
			5000	4995	-5
			6000	6000	0
			0	0	0

### AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
6000	3000	9000	0

OVERALL TEST RESULTS PASS



GREAT LAKES SCALE COMPANY  
 29540 CALAHAN RD.  
 ROSEVILLE, MI 48066  
 Phone: (586) 776-0517

## Certificate of Calibration

**Customer:** SUPERIOR MATERIALS 102  
 SUP1112 44922 GRAND RIVER AVE

**Certificate ID:** 105134 AGG180117025718

**Test Date:** 01/17/2018

NOV, MI 48375

**ISO Number:** MS-FM-CEM-001

(\*)Make: ALKON Model : COMMAND SN#: 105134 AGG Capacity: 40000 lb Graduation: 20 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
01/17/2018			105134 AGG	II
SCALE LOCATION		EQUIPMENT CONDITIONS		
		Working		
SELECT NUMBER OF FM DAYS:		CAPACITY	40000 UOM lb	INCREMENT 20

### AS FOUND TEST USING TEST WEIGHTS

### AS LEFT TEST USING TEST WEIGHTS

Weight Applied	Reading	Error	Weight Applied	Reading	Error
0	lb 0	lb 0		lb 0	lb 0
6000	lb 6000	lb 0		lb 0	lb 0
0	lb 0	lb 0		lb 0	lb 0
1000	lb 1000	lb 0		lb 0	lb 0
2000	lb 2000	lb 0		lb 0	lb 0
3000	lb 3000	lb 0		lb 0	lb 0
4000	lb 4000	lb 0		lb 0	lb 0
5000	lb 5000	lb 0		lb 0	lb 0
6000	lb 6000	lb 0		lb 0	lb 0
0	lb 0	lb 0		lb 0	lb 0

### AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
2000	lb 6000	lb 8020	lb 20
4000	lb 6000	lb 10020	lb 20
6000	lb 6000	lb 12020	lb 20
2000	lb 12000	lb 14020	lb 20
4000	lb 12000	lb 16020	lb 20
6000	lb 12000	lb 18020	lb 20
2000	lb 18000	lb 20020	lb 20
4000	lb 18000	lb 22020	lb 20
6000	lb 18000	lb 24020	lb 20
2000	lb 24000	lb 26020	lb 20
4000	lb 24000	lb 28020	lb 20
6000	lb 24000	lb 30020	lb 20
2000	lb 30000	lb 32020	lb 20
4000	lb 30000	lb 34020	lb 20
6000	lb 30000	lb 36020	lb 20
	lb	lb	lb
	lb	lb	lb
	lb	lb	lb
	lb	lb	lb

OVERALL TEST RESULTS PASS



GREAT LAKES SCALE COMPANY  
 29540 CALAHAN RD.  
 ROSEVILLE, MI 48066  
 Phone: (586) 776-0517

## Certificate of Calibration

Customer: SUPERIOR MATERIALS 102  
 SUP1112 44922 GRAND RIVER AVE

Certificate ID: 18020010518020001

Test Date: 1/17/2018

NOV, MI 48375

ISO Number: MS-FM-CEM-001

(\*Make: RICELAKE Model : 920 SN#: 1802300105 Capacity: 20000 lb Graduation: 20lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
1/17/2018			1802300105	II
SCALE LOCATION		EQUIPMENT CONDITIONS		
		Working	Clean	
SELECT NUMBER OF PM DAYS:	CAPACITY	20000 UOM	lb	INCREMENT
				20

### AS FOUND TEST USING TEST WEIGHTS

### AS LEFT TEST USING TEST WEIGHTS

Weight Applied	Reading	Error		Weight Applied	Reading	Error	
0	lb 0	lb 0	lb		lb 0	lb 0	lb
6000	lb 6000	lb 0	lb		lb 0	lb 0	lb
0	lb 0	lb 0	lb		lb 0	lb 0	lb
2000	lb 2000	lb 0	lb		lb 0	lb 0	lb
4000	lb 4000	lb 0	lb		lb 0	lb 0	lb
6000	lb 6000	lb 0	lb		lb 0	lb 0	lb
4000	lb 4000	lb 0	lb		lb 0	lb 0	lb
2000	lb 2000	lb 0	lb		lb 0	lb 0	lb
0	lb 0	lb 0	lb		lb 0	lb 0	lb
	lb	lb	lb		lb 0	lb 0	lb

### AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
lb		lb	lb 0
lb		lb	lb 0
lb		lb	lb 0
lb		lb	lb 0
lb		lb	lb 0
lb		lb	lb 0
lb		lb	lb 0
lb		lb	lb 0
lb		lb	lb 0
lb		lb	lb 0
lb		lb	lb 0
lb		lb	lb 0
lb		lb	lb 0
lb		lb	lb 0
lb		lb	lb 0
lb		lb	lb 0
lb		lb	lb 0
lb		lb	lb 0
lb		lb	lb 0
lb		lb	lb 0
lb		lb	lb 0
lb		lb	lb 0
lb		lb	lb 0
lb		lb	lb 0
lb		lb	lb 0
lb		lb	lb 0
lb		lb	lb 0
lb		lb	lb 0
lb		lb	lb 0
lb		lb	lb 0
lb		lb	lb 0

**OVERALL TEST RESULTS PASS**



GREAT LAKES SCALE COMPANY  
 29540 CALAHAN RD.  
 ROSEVILLE, MI 48066  
 Phone: (586) 776-0517

## Certificate of Calibration

**Customer:** SUPERIOR MATERIALS 102  
 SUP1112 44922 GRAND RIVER AVE

**Certificate ID:** 105134 WATER 180117025130

**Test Date:** 01/17/2018

NOV, MI 48375

**ISO Number:** MS-FM-CEM-001

(\*Make: ALKON Model: COMMAND SN#: 105134 WATER Capacity: 3000 lb Graduation: 5 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
01/17/2018			105134 WATER	II
SCALE LOCATION		EQUIPMENT CONDITIONS		
		Working Dirty		
SELECT NUMBER OF FM DAYS:		CAPACITY	3000 UOM	lb INCREMENT
				5

**AS FOUND TEST USING TEST WEIGHTS**

**AS LEFT TEST USING TEST WEIGHTS**

Weight Applied				Reading				Error					
0	lb	0	lb	0	lb	0	lb	0	lb	0	lb	0	lb
3000	lb	2990	lb	-10	lb	3000	lb	3000	lb	0	lb	0	lb
0	lb	0	lb	0	lb	0	lb	0	lb	0	lb	0	lb
	lb		lb		lb	1000	lb	1000	lb	0	lb	0	lb
	lb		lb		lb	2000	lb	2000	lb	0	lb	0	lb
	lb		lb		lb	3000	lb	3000	lb	0	lb	0	lb
	lb		lb		lb	0	lb	0	lb	0	lb	0	lb
	lb		lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb		lb

**AS LEFT BUILD TEST**

Test Weight Applied		Material Weight Applied		Reading		Error	
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb

OVERALL TEST RESULTS N/A