



GREAT LAKES SCALE COMPANY
29540 CALAHAN RD.
ROSEVILLE, MI 48066
Phone: (586) 776-0517

Certificate of Calibration

Customer: SUPERIOR MATERIALS 32
SUP1173 8911 WEST JEFFERSON

Certificate ID: 62821 CEMENT190119121832

Test Date: 01/19/2018

DETROIT, MI

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model: COMMAND SN#: 62821 CEMENT Capacity: 10000 lb Graduation: 5 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
01/19/2018			62821 CEMENT	II
SCALE LOCATION		EQUIPMENT CONDITIONS		
		Working	Dirty	
SELECT NUMBER OF FM DAYS:	CAPACITY	10000 UOM	lb	INCREMENT
				5

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

Weight Applied	Reading	Error	Weight Applied	Reading	Error
0	lb 0	lb 0	0	lb 0	lb 0
5000	lb 5360	lb -40	5000	lb 5000	lb 0
0	lb 0	lb 0	0	lb 0	lb 0
	lb	lb	1000	lb 1000	lb 0
	lb	lb	2000	lb 2000	lb 0
	lb	lb	3000	lb 3000	lb 0
	lb	lb	4000	lb 4000	lb 0
	lb	lb	5000	lb 5000	lb 0
	lb	lb	6000	lb 6000	lb 0
	lb	lb	0	lb 0	lb 0

AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
6000	lb 3000	lb 9000	lb 0
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb

OVERALL TEST RESULTS PASS



GREAT LAKES SCALE COMPANY
 29540 CALAHAN RD.
 ROSEVILLE, MI 48066
 Phone: (586) 776-0517

Certificate of Calibration

Customer: SUPERIOR MATERIALS 32
 SUP1173 8911 WEST JEFFERSON

Certificate ID: 002021 AGG A100119123150

Test Date: 01/19/2018

DETROIT, MI

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: 662821 AGG A Capacity: 40000 LB Graduation: 20 LB

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
01/19/2018			662821 AGG A	II
SCALE LOCATION		EQUIPMENT CONDITIONS		
AGG A		Working Dirty		
SELECT NUMBER OF FM DAYS:		CAPACITY	40000 UOM LB	INCREMENT 20

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

Weight Applied	Reading	Error	Weight Applied	Reading	Error
0	LB 0	LB 0		LB 0	LB 0
6000	LB 6000	LB 0		LB 0	LB 0
0	LB 0	LB 0		LB 0	LB 0
1000	LB 1000	LB 0		LB 0	LB 0
2000	LB 2000	LB 0		LB 0	LB 0
3000	LB 3000	LB 0		LB 0	LB 0
4000	LB 4000	LB 0		LB 0	LB 0
5000	LB 5000	LB 0		LB 0	LB 0
6000	LB 6000	LB 0		LB 0	LB 0
0	LB 0	LB 0		LB 0	LB 0

AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
2000	LB 6000	LB 8000	LB 0
4000	LB 6000	LB 10000	LB 0
6000	LB 6000	LB 12000	LB 0
2000	LB 12000	LB 14000	LB 0
4000	LB 12000	LB 16000	LB 0
6000	LB 12000	LB 18000	LB 0
2000	LB 18000	LB 20000	LB 0
4000	LB 18000	LB 22000	LB 0
6000	LB 18000	LB 24000	LB 0
2000	LB 24000	LB 26000	LB 0
4000	LB 24000	LB 28000	LB 0
6000	LB 24000	LB 30000	LB 0
2000	LB 30000	LB 32000	LB 0
4000	LB 30000	LB 34000	LB 0
6000	LB 30000	LB 36000	LB 0
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB

OVERALL TEST RESULTS PASS



GREAT LAKES SCALE COMPANY
29540 CALAHAN RD.
ROSEVILLE, MI 48066
Phone: (586) 776-0517

Certificate of Calibration

Customer: SUPERIOR MATERIALS 32
SUP1173 8911 WEST JEFFERSON

Certificate ID: 002021 WATER100119122025

Test Date: 01/19/2018

DETROIT, MI

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: 662821 WATER Capacity: 6000 lb Graduation: 1 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
01/19/2018			662821 WATER	II
SCALE LOCATION		EQUIPMENT CONDITIONS		
		Working	Dirty	
SELECT NUMBER OF FM DAYS:	CAPACITY	6000 UOM	lb	INCREMENT
				1

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

Weight Applied	Reading	Error	Weight Applied	Reading	Error
0	lb 0	lb 0	0	lb 0	lb 0
4000	lb 3994	lb -6	1000	lb 1000	lb 0
0	lb 0	lb 0	2000	lb 2000	lb 0
	lb	lb	3000	lb 3000	lb 0
	lb	lb	4000	lb 4000	lb 0
	lb	lb	0	lb 0	lb 0
	lb	lb		lb	lb
	lb	lb		lb	lb
	lb	lb		lb	lb
	lb	lb		lb	lb

AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
lb		lb	0
lb		lb	0
lb		lb	0
lb		lb	0
lb		lb	0
lb		lb	0
lb		lb	0
lb		lb	0
lb		lb	0
lb		lb	0
lb		lb	0
lb		lb	0
lb		lb	0
lb		lb	0
lb		lb	0
lb		lb	0
lb		lb	0
lb		lb	0
lb		lb	0
lb		lb	0

OVERALL TEST RESULTS PASS