



15231 Ten Mile Road • Eastpointe, MI 48021
 888.776.0517 • Fax 586.776.3169
 www.greatlakescale.com

SALES | SERVICE | RENTALS

MASTER DISTRIBUTORS OF:
 CARDINAL • SETRA • GSE • OHAUS
 AVERY WEIGHTRONIX • RICELAKE

ISO/IEC 17025 ACCREDITED
 STATE REGISTERED SERVICE AGENCY

4211



CALIBRATION CERTIFICATE / PROPORTIONING REPORT

CUSTOMER Superior Plant
 PHONE _____ DATE 1/8/16

LOCATION 20965 Hoover St, Detroit
 PO NUMBER 48205

SCALE #1		SCALE #2	
MAKE <u>AKLON</u>	TEMPERATURE <u>35°</u>	MAKE <u>AKLON</u>	TEMPERATURE <u>55°</u>
MODEL <u>SPECTRUM</u>	S/N WEIGHTS USED:	MODEL <u>SPECTRUM</u>	S/N WEIGHTS USED:
SERIAL # <u>72113</u>	<u>GLSK 1-6</u>	SERIAL # <u>72113</u>	<u>GLSK 1-6</u>
CAPACITY <u>10,000'S</u>		CAPACITY <u>40000'S</u>	
RESOLUTION <u>5'S</u>		RESOLUTION <u>40'S</u>	
LOCATION <u>Cement</u>		LOCATION <u>AGGREGATE</u>	

WT APPLIED	AS FOUND	WT APPLIED	AS LEFT	WT APPLIED	AS FOUND	WT APPLIED	AS LEFT
<u>0'S</u>	<u>0'S</u>			<u>0'S</u>	<u>0'S</u>	<u>MAT 12000'S</u>	
<u>1000'S</u>	<u>1000'S</u>			<u>1000'S</u>	<u>1000'S</u>	<u>WLT 6000'S</u>	<u>18000'S</u>
<u>2000'S</u>	<u>2000'S</u>			<u>2000'S</u>	<u>2000'S</u>	<u>MAT 18000'S</u>	
<u>3000'S</u>	<u>3000'S</u>			<u>3000'S</u>	<u>3000'S</u>	<u>WLT 6000'S</u>	<u>24000'S</u>
<u>4000'S</u>	<u>4000'S</u>			<u>4000'S</u>	<u>4000'S</u>	<u>MAT 24000'S</u>	
<u>5000'S</u>	<u>5000'S</u>			<u>5000'S</u>	<u>5000'S</u>	<u>WLT 6000'S</u>	<u>29960'S</u>
<u>6000'S</u>	<u>6000'S</u>			<u>6000'S</u>	<u>6000'S</u>	<u>MAT 30000</u>	
				<u>MAT 6000'S</u>		<u>WLT 6000'S</u>	<u>35560'S</u>
				<u>WLT 6000'S</u>	<u>12,000'S</u>		

SHIFT TEST		SHIFT TEST		SHIFT TEST		SHIFT TEST	
WT APPLIED	AS FOUND	WT APPLIED	AS LEFT	WT APPLIED	AS FOUND	WT APPLIED	AS LEFT
<u>1</u>		<u>1</u>		<u>1</u>		<u>1</u>	
<u>2</u>	<u>N/A</u>	<u>2</u>		<u>2</u>		<u>2</u>	
<u>3</u>		<u>3</u>		<u>3</u>		<u>3</u>	
<u>4</u>		<u>4</u>		<u>4</u>		<u>4</u>	
<u>5</u>		<u>5</u>		<u>5</u>		<u>5</u>	

MEASUREMENT UNCERTAINTY	MEASUREMENT UNCERTAINTY	MEASUREMENT UNCERTAINTY	MEASUREMENT UNCERTAINTY
\pm (<u>+0.6R</u>)	<u>R=RESOLUTION</u>	\pm (<u>+0.6R</u>)	<u>R=RESOLUTION</u>

TECH REMARKS	TECH REMARKS
<u>Scale on 10 load tested sealed</u>	<u>Scale on 10 load tested build tested to 36000'S sealed</u>

PARTS SUPPLIED
<u>Minimum - 2 hours Labor - 2 Men</u>
Tr. In - <u>7:00 - 7:15</u>
Work - <u>7:15</u>
Tr. Out _____
Mileage <u>10</u> Truck # <u>22</u>

These Results Relate Only to Items Tested
 Scales meet NIST standard/Handbook 44
 Scales fail to meet NIST standards

1 2

WORK ORDERED BY
[Signature]

TECHNICIAN SIGNATURE
[Signature]