



GREAT LAKES SCALE COMPANY
 29540 CALAHAN RD.
 ROSEVILLE, MI 48066
 Phone: (586) 776-0517

Customer: SUPERIOR MATERIALS 103
 SUP0992 10147 GRAND RIVER

BRIGHTON, MI 48116

Certificate ID: 23635-c180906113336

Test Date: 09/06/2018

ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	3.39
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	70
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass adjustment needed

Comments:

Test scale within maint tolerance

TRACEABILITY CERTIFICATE NUMBERS:

Kit	NIST	Date	Date Due	Description
GL0601	MI-06-17-13260	6/15/2017		
		6/15/2019		TRUCK 6 - 1000 LB
GL0602	MI-06-17-13260	6/15/2017		
		6/15/2019		TRUCK 6 - 1000 LB
GL0603	MI-06-17-13260	6/15/2017		
		6/15/2019		TRUCK 6 - 1000 LB
GL0604	MI-06-17-13260	6/15/2017		
		6/15/2019		TRUCK 6 - 1000 LB

Technician(s): RENTE EVANS;Joe bielarz

Signature:

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.



GREAT LAKES SCALE COMPANY
29540 CALAHAN RD.
ROSEVILLE, MI 48066
Phone: (586) 776-0517

Customer: SUPERIOR MATERIALS 103
SUP0992 10147 GRAND RIVER

BRIGHTON, MI 48116

Certificate ID: 23635-a180906112646

Test Date: 09/06/2018

ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	3.68
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	70
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

Comments:

Test scale within maint tolerance

TRACEABILITY CERTIFICATE NUMBERS:

Kit	NIST	Date	Date Due
Description			
GL0601	MI-06-17-13260	6/15/2017	6/15/2019
TRUCK 6 - 1000 LB			
GL0602	MI-06-17-13260	6/15/2017	6/15/2019
TRUCK 6 - 1000 LB			
GL0603	MI-06-17-13260	6/15/2017	6/15/2019
TRUCK 6 - 1000 LB			
GL0604	MI-06-17-13260	6/15/2017	6/15/2019
TRUCK 6 - 1000 LB			
GL0605	MI-06-17-13260	6/15/2017	6/15/2019
TRUCK 6 - 1000 LB			
GL0606	MI-06-17-13260	6/15/2017	6/15/2019
TRUCK 6 - 1000 LB			

Technician(s): RENTE EVANS; Joe bielarz

Signature:

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.



GREAT LAKES SCALE COMPANY
29540 CALAHAN RD.
ROSEVILLE, MI 48066
Phone: (586) 776-0517

Customer: SUPERIOR MATERIALS 103
SUP0992 10147 GRAND RIVER

BRIGHTON, MI 48116

Certificate ID: 23635-b180906113033

Test Date: 09/06/2018

ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION EXPANDED UNCERTAINTY
F 15
ENVIRONMENTAL CONDITIONS TEMPERATURE
Acceptable 70
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?
Yes
TEST LOCATION OVERALL RESULT
Onsite Pass w/o adjustment

Comments:

Test scale within maint tolerance

TRACEABILITY CERTIFICATE NUMBERS:

Kit	NIST	Date	Date Due
Description			
GL0601	MI-06-17-13260	6/15/2017	6/15/2019
TRUCK 6 - 1000 LB			
GL0602	MI-06-17-13260	6/15/2017	6/15/2019
TRUCK 6 - 1000 LB			
GL0603	MI-06-17-13260	6/15/2017	6/15/2019
TRUCK 6 - 1000 LB			
GL0604	MI-06-17-13260	6/15/2017	6/15/2019
TRUCK 6 - 1000 LB			
GL0605	MI-06-17-13260	6/15/2017	6/15/2019
TRUCK 6 - 1000 LB			
GL0606	MI-06-17-13260	6/15/2017	6/15/2019
TRUCK 6 - 1000 LB			

Technician(s): RENTE EVANS; Joe bielarz

Signature:

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.