



29540 Calahan • Roseville, MI 48066
888.776.0517 • Fax 586.776.3169
www.greatlakescale.com

SALES | SERVICE | RENTALS

MASTER DISTRIBUTORS OF:
CARDINAL • SETRA • GSE • OHAUS
AVERY WEIGHTRONIX • RICELAKE

ISO/IEC 17025 ACCREDITED
STATE REGISTERED SERVICE AGENCY

6595



CALIBRATION CERTIFICATE / PROPORTIONING REPORT

CUSTOMER Superior #103 LOCATION 10147 Grand River Brighton MI
PHONE 810-560-3539 DATE 8-2-17 PO NUMBER

| SCALE #1 | | SCALE #2 | |
|-----------------------|-----------------------|------------|-------------------|
| MAKE <u>Command</u> | TEMPERATURE <u>70</u> | MAKE | TEMPERATURE |
| MODEL <u>AL 1000</u> | S/N WEIGHTS USED: | MODEL | S/N WEIGHTS USED: |
| SERIAL # <u>23635</u> | <u>talk to</u> | SERIAL # | |
| CAPACITY <u>4000</u> | | CAPACITY | |
| RESOLUTION <u>25</u> | | RESOLUTION | |
| LOCATION <u>water</u> | | LOCATION | |

| WT APPLIED | AS FOUND | WT APPLIED | AS LEFT | WT APPLIED | AS FOUND | WT APPLIED | AS LEFT |
|------------|----------|------------|---------|------------|----------|------------|---------|
| 0 | 0 | | | | | | |
| 1000 | 1000 | | | | | | |
| 2000 | 2000 | | | | | | |
| 3000 | 3000 | | | | | | |
| 4000 | 4000 | | | | | | |
| 0 | 0 | | | | | | |

| SHIFT TEST | | SHIFT TEST | | SHIFT TEST | | SHIFT TEST | |
|------------|----------|------------|---------|------------|----------|------------|---------|
| WT APPLIED | AS FOUND | WT APPLIED | AS LEFT | WT APPLIED | AS FOUND | WT APPLIED | AS LEFT |
| 1 | | 1 | | 1 | | 1 | |
| 2 | | 2 | | 2 | | 2 | |
| 3 | | 3 | | 3 | | 3 | |
| 4 | | 4 | | 4 | | 4 | |
| 5 | | 5 | | 5 | | 5 | |

MEASUREMENT UNCERTAINTY $\pm(+0.6R)$ R=RESOLUTION

TECH REMARKS

test and cal scale

PARTS SUPPLIED

[Signature] Minimum - 2 hours Labor - 2 Men
Tr. In - _____
Work - _____
Tr. Out - _____
Mileage _____ Truck # 6

These Results Relate Only to Items Tested
Scales meet NIST standard Handbook 44
Scales fail to meet NIST standards

TECHNICIAN SIGNATURE [Signature] CUSTOMER SIGNATURE [Signature]

GLS-QMS-FM-004-REV1 Refer to State Test Report for weights applied
This report may not be reproduced except in full, without the written approval of Great Lakes Scale Company.
Best measurement or uncertainty for weights used during calibration is on our scope of accreditation, which we will provide upon request.



29540 Calahan • Roseville, MI 48066
888.776.0517 • Fax 586.776.3169
www.greatlakescale.com

SALES | SERVICE | RENTALS

MASTER DISTRIBUTORS OF:
CARDINAL • SETRA • GSE • OHAUS
AVERY WEIGHTRONIX • RICELAKE

ISO/IEC 17025 ACCREDITED
STATE REGISTERED SERVICE AGENCY

6594



CALIBRATION CERTIFICATE / PROPORTIONING REPORT

CUSTOMER SUPERION #103 LOCATION 10147 GRAND RIVER MI
PHONE 810-560-3539 DATE 10-2-17 PO NUMBER

| SCALE #1 | | SCALE #2 | |
|-------------------------|-----------------------|------------------------|-----------------------|
| MAKE <u>Command</u> | TEMPERATURE <u>70</u> | MAKE <u>Command</u> | TEMPERATURE <u>70</u> |
| MODEL <u>Alkon</u> | S/N WEIGHTS USED: | MODEL <u>Alkon</u> | S/N WEIGHTS USED: |
| SERIAL # <u>23635</u> | <u>TRK 6</u> | SERIAL # <u>23635</u> | <u>TRK 6</u> |
| CAPACITY <u>10,000</u> | | CAPACITY <u>45,000</u> | |
| RESOLUTION <u>15%</u> | | RESOLUTION <u>X 20</u> | |
| LOCATION <u>Command</u> | | LOCATION <u>Agg</u> | |

| WT APPLIED | AS FOUND | WT APPLIED | AS LEFT | WT APPLIED | AS FOUND | WT APPLIED | AS LEFT |
|------------|----------|------------|---------|------------|----------|------------|---------|
| 0 | 0 | | | 0 | 0 | wt 6000 | 38000 |
| 1000 | 1000 | | | 1000 | 1000 | wt 38000 | 38000 |
| 2000 | 2000 | | | 2000 | 2000 | wt 6000 | 44000 |
| 3000 | 3000 | | | 3000 | 3000 | | |
| 4000 | 4000 | | | 4000 | 4000 | | |
| 5000 | 5000 | | | 5000 | 5000 | | |
| 6000 | 6000 | | | 6000 | 6000 | | |
| Mat 3300 | 3300 | | | Mat 8000 | 8000 | | |
| wt 6000 | 9300 | | | wt 6000 | 14000 | | |

| SHIFT TEST | | SHIFT TEST | | SHIFT TEST | | SHIFT TEST | |
|------------|----------|------------|---------|-------------|----------|------------|---------|
| WT APPLIED | AS FOUND | WT APPLIED | AS LEFT | WT APPLIED | AS FOUND | WT APPLIED | AS LEFT |
| 1 | | 1 | | 1 Mat 16000 | 16000 | 1 | |
| 2 | | 2 | | 2 wt 6000 | 22000 | 2 | |
| 3 | | 3 | | 3 Mat 24000 | 24000 | 3 | |
| 4 | | 4 | | 4 wt 6000 | 30000 | 4 | |
| 5 | | 5 | | 5 Mat 32000 | 32000 | 5 | |

MEASUREMENT UNCERTAINTY MEASUREMENT UNCERTAINTY MEASUREMENT UNCERTAINTY MEASUREMENT UNCERTAINTY
±(+0.6R) R=RESOLUTION ±(+0.6R) R=RESOLUTION

TECH REMARKS
test scale test scale

PARTS SUPPLIED
[Signature]
Minimum - 2 hours Labor / 2 Men
Tr. In - _____
Work - _____
Tr. Out _____
Mileage _____ Truck # 6

These Results Relate Only to Items Tested
Scales meet NIST standard Handbook-44
Scales fail to meet NIST standards
1 2 WORK ORDERED BY Dave T
TECH SIGNATURE [Signature] CUSTOMER SIGNATURE [Signature]

GLS-QMS-FM-004-REV1 Refer to State Test Report for weights applied.
This report may not be reproduced except in full, without the written approval of Great Lakes Scale Company.
Best measurement or uncertainty for weights used during calibration is on our scope of accreditation, which we will provide upon request.