



15231 Ten Mile Road • Eastpointe, MI 48021
 888.776.0517 • Fax 586.776.3169
 www.greatlakescale.com

SALES | SERVICE | RENTALS

MASTER DISTRIBUTORS OF:
 CARDINAL • SETRA • GSE • OHAUS
 AVERY WEIGHTRONIX • RICELAKE

ISO/IEC 17025 ACCREDITED
 STATE REGISTERED SERVICE AGENCY

4266



CALIBRATION CERTIFICATE / PROPORTIONING REPORT

CUSTOMER SUPERIOR # 38 8911 WEST
 PHONE 810-217-4314 DATE 8-11-16

LOCATION JEFFERSON DETROIT MI # 42209
 PO NUMBER

SCALE #1

SCALE #2

MAKE <u>ALCON</u>	TEMPERATURE <u>11°</u>	MAKE <u>ALCON</u>	TEMPERATURE <u>11°</u>
MODEL <u>SPECTRUM</u>	S/N WEIGHTS USED:	MODEL <u>SPECTRUM</u>	S/N WEIGHTS USED:
SERIAL # <u>61821</u>	<u>661501-661506</u>	SERIAL # <u>61821</u>	<u>661501-661504</u>
CAPACITY <u>10,000 LBS</u>		CAPACITY <u>6000 LBS</u>	
RESOLUTION <u>X 5 LBS</u>		RESOLUTION <u>X 1 LBS</u>	
LOCATION <u>CEMENT</u>		LOCATION <u>WATER</u>	

WT APPLIED	AS FOUND	WT APPLIED	AS LEFT	WT APPLIED	AS FOUND	WT APPLIED	AS LEFT
0	0			0	0	0	0
1000	1000			4000	3994	1000	1000
1000	1000			0	0	1000	1000
2000	2000					3000	3000
4000	4000					4000	4000
5000	5000					0	0
6000	6000						
HAT 3000	3000						
WT 6000	7000						

SHIFT TEST		SHIFT TEST		SHIFT TEST		SHIFT TEST	
WT APPLIED	AS FOUND	WT APPLIED	AS LEFT	WT APPLIED	AS FOUND	WT APPLIED	AS LEFT
1		1		1		1	
2		2		2		2	
3	<u>M/A</u>	3		3	<u>M/A</u>	3	
4		4		4		4	
5		5		5		5	

MEASUREMENT UNCERTAINTY $\pm(.68 LB + 0.6R)$	MEASUREMENT UNCERTAINTY $R=RESOLUTION$	MEASUREMENT UNCERTAINTY $\pm(.68 LB + 0.6R)$	MEASUREMENT UNCERTAINTY $R=RESOLUTION$
---	--	---	--

TECH REMARKS	TECH REMARKS
<u>TEST SCALE SCALE WORKING 6000. APPROVED BY WAYNE COUNTY.</u>	<u>TEST AND CALIBRATE SCALE SAME WORKING 6000. APPROVED BY WAYNE COUNTY.</u>

PARTS SUPPLIED	Minimum - 2 hours Labor - 2 Men
	Tr. In - <u>6:45 - 7:50</u>
	Work - <u>7:30 - 10:45</u>
	Tr. Out
	Mileage <u>35</u> Truck # <u>02</u>

These Results Relate Only to Items Tested	1 <input checked="" type="checkbox"/>	2 <input checked="" type="checkbox"/>	WORK ORDERED BY
Scales meet NIST standard/Handbook 44	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<u>GARY LOWELL</u>
Scales fail to meet NIST standards	<input type="checkbox"/>	<input type="checkbox"/>	
TECHNICIAN SIGNATURE <u>[Signature]</u>	CUSTOMER SIGNATURE <u>[Signature]</u>		



15231 Ten Mile Road • Eastpointe, MI 48021
 888.776.0517 • Fax 586.776.3169
 www.greatlakescale.com

SALES | SERVICE | RENTALS

MASTER DISTRIBUTORS OF:
 CARDINAL • SETRA • GSE • OHAUS
 AVERY WEIGHTRONIX • RICELAKE

ISO/IEC 17025 ACCREDITED
 STATE REGISTERED SERVICE AGENCY

4267



CALIBRATION CERTIFICATE / PROPORTIONING REPORT

CUSTOMER SUPERIOR #12 8TH WEST JEFFERSON LOCATION DETROIT MICHIGAN 48209
 PHONE 810-277-4314 DATE 8-11-16 PO NUMBER _____

SCALE #1

SCALE #2

MAKE <u>ALCON</u>	TEMPERATURE <u>110</u>	MAKE _____	TEMPERATURE _____
MODEL <u>SPECTRUM</u>	S/N WEIGHTS USED: _____	MODEL _____	S/N WEIGHTS USED: _____
SERIAL # <u>63831</u>	<u>661501 - 661506</u>	SERIAL # _____	_____
CAPACITY <u>40,000 LBS</u>	_____	CAPACITY _____	_____
RESOLUTION <u>1/20 LBS</u>	_____	RESOLUTION _____	_____
LOCATION <u>AGG</u>	_____	LOCATION _____	_____

WT APPLIED	AS FOUND	WT APPLIED	AS LEFT	WT APPLIED	AS FOUND	WT APPLIED	AS LEFT
0	0	WT 6000	6000	WT 18000	18000	WT 30000	30000
6000	6000	WT 2000	8000	WT 2000	32000	WT 2000	39000
0	0	4000	10000	4000	37000	4000	34000
1000	1000	6000	12000	6000	34000	6000	36000
2000	2000	WT 13000	13000	WT 14000	34000		
3000	3000	WT 2000	14000	WT 2000	36000		
4000	4000	4000	16000	4000	38000		
5000	5000	6000	18000	6000	37000		
6000	6000						

SHIFT TEST		SHIFT TEST		SHIFT TEST		SHIFT TEST	
WT APPLIED	AS FOUND	WT APPLIED	AS LEFT	WT APPLIED	AS FOUND	WT APPLIED	AS LEFT
1		1		1		1	
2	<u>M/A</u>	2		2	<u>M/A</u>	2	
3		3		3		3	
4		4		4		4	
5		5		5		5	

MEASUREMENT UNCERTAINTY ±(<u>3.0LB +0.6R</u>)	MEASUREMENT UNCERTAINTY <u>R=RESOLUTION</u>	MEASUREMENT UNCERTAINTY ±(<u>+0.6R</u>)	MEASUREMENT UNCERTAINTY <u>R=RESOLUTION</u>
---	---	---	---

TECH REMARKS
TEST SCALE, FOUND SMALL PIECE OF SWAP STEEL WEDGED BETWEEN SCALE AND
SAFETY ROLL. REMOVED STEEL. SCALE WORKING GOOD. SCALE APPROVED BY
WAYNE COUNTY.

PARTS SUPPLIED

Minimum - 2 hours Labor - 2 Men
 Tr. In - _____
 Work - SEE REPORT 4266
 Tr. Out - _____
 Mileage 35 Truck # 22

These Results Relate Only to Items Tested
 Scales meet NIST standard/Handbook 44 1 2
 Scales fail to meet NIST standards 1 2

TECHNICIAN SIGNATURE [Signature] MIKE - MATTEP.

WORK ORDERED BY
GARY LOWELL

CUSTOMER SIGNATURE [Signature]