



GREAT LAKES SCALE

29540 Calahan • Roseville, MI 48066
888.776.0517 • Fax 586.776.3169
www.greatlakescale.com

SALES | SERVICE | RENTALS

MASTER DISTRIBUTORS OF:
CARDINAL • SETRA • GSE • OHAUS
AVERY WEIGHTRONIX • RICELAKE

ISO/IEC 17025 ACCREDITED
STATE REGISTERED SERVICE AGENCY

5627



CALIBRATION CERTIFICATE / PROPORTIONING REPORT

CUSTOMER SUPERIOR PLAST 5103 10147 GRABO LOCATION BRIDGE BRIGHTON MICHIGAN 48116
PHONE 248-788-8048 DATE 10-7-16 PO NUMBER

SCALE #1

SCALE #2

MAKE <u>ALCON</u>	TEMPERATURE <u>75°</u>	MAKE	TEMPERATURE
MODEL <u>COMMAND</u>	S/N WEIGHTS USED:	MODEL	S/N WEIGHTS USED:
SERIAL # <u>A3635</u>	<u>661501-661506</u>	SERIAL #	
CAPACITY <u>40000 LBS</u>		CAPACITY	
RESOLUTION <u>X 10 LBS</u>		RESOLUTION	
LOCATION <u>AGGREGATE</u>		LOCATION	

WT APPLIED	AS FOUND	WT APPLIED	AS LEFT	WT APPLIED	AS FOUND	WT APPLIED	AS LEFT
0	0	MAT 10000	10000				
6000	6000	WT 6000	16000				
0	0	MAT 20000	20000				
2000	2000	WT 6000	26000				
4000	4000	MAT 30000	30000				
6000	6000	WT 6000	36000				
4000	4000						
2000	2000						
0	0						

SHIFT TEST		SHIFT TEST		SHIFT TEST		SHIFT TEST	
WT APPLIED	AS FOUND	WT APPLIED	AS LEFT	WT APPLIED	AS FOUND	WT APPLIED	AS LEFT
1		1		1		1	
2		2		2		2	
3		3		3		3	
4		4		4		4	
5		5		5		5	

MEASUREMENT UNCERTAINTY $\pm(3.0LB +0.6R)$	MEASUREMENT UNCERTAINTY R=RESOLUTION	MEASUREMENT UNCERTAINTY $\pm(+0.6R)$	MEASUREMENT UNCERTAINTY R=RESOLUTION
---	--------------------------------------	---------------------------------------	--------------------------------------

TECH REMARKS	TECH REMARKS
<u>TRST SCALE, WORKING GOOD.</u>	

PARTS SUPPLIED

Ø
 Minimum - 2 hours Labor - 2 Men
 Tr. In -
 Work - SEE REPORT 5618
 Tr. Out
 Mileage _____ Truck # 02

These Results Relate Only to Items Tested
Scales meet NIST standard/Handbook 44
Scales fail to meet NIST standards

1 2
 WORK ORDERED BY
DAV THOMAS
 CUSTOMER SIGNATURE
[Signature]

TECHNICIAN SIGNATURE
[Signature] MIKE-JOE



GREAT LAKES SCALE

29540 Calahan • Roseville, MI 48066
888.776.0517 • Fax 586.776.3169
www.greatlakescale.com

SALES | SERVICE | RENTALS

MASTER DISTRIBUTORS OF:
CARDINAL • SETRA • GSE • OHAUS
AVERY WEIGHTRONIX • RICELAKE

ISO/IEC 17025 ACCREDITED
STATE REGISTERED SERVICE AGENCY

5628



CALIBRATION CERTIFICATE / PROPORTIONING REPORT

CUSTOMER SUPERIOR PUMP 5103 10147 GRAND RIVER LOCATION BRIGHTON MICHIGAN 48116

PHONE 248-788-8048 DATE 10-7-16 PO NUMBER

SCALE #1

SCALE #2

MAKE ALKOP	TEMPERATURE 75°	MAKE ALKOP	TEMPERATURE 75°
MODEL COMMAND	S/N WEIGHTS USED:	MODEL COMMAND	S/N WEIGHTS USED:
SERIAL # 23635	GL1501-GL1506	SERIAL # 23635	GL1501-GL1504
CAPACITY 10000 LBS		CAPACITY 4000 LBS	
RESOLUTION X 5 LBS		RESOLUTION X 5 LBS	
LOCATION CEMENT		LOCATION WATER	

WT APPLIED	AS FOUND	WT APPLIED	AS LEFT	WT APPLIED	AS FOUND	WT APPLIED	AS LEFT
0	0	0	0	0	0	0	0
6000	5990	6000	6000	4000	4030	4000	4000
0	0	0	0	0	0	0	0
		2000	2000			0	0
		4000	4000			1000	1000
		6000	6000			2000	2000
		4000	4000			4000	4000
		2000	2000			2000	2000
		0	0			1000	1000

SHIFT TEST		SHIFT TEST		SHIFT TEST		SHIFT TEST	
WT APPLIED	AS FOUND	WT APPLIED	AS LEFT	WT APPLIED	AS FOUND	WT APPLIED	AS LEFT
1		1 WT 3000	3000	1		1	0
2		2 WT 6000	6000	2		2	
3		3		3		3	
4		4		4		4	
5		5		5		5	

MEASUREMENT UNCERTAINTY ±(.68 LBS +0.6R)	MEASUREMENT UNCERTAINTY R=RESOLUTION	MEASUREMENT UNCERTAINTY ±(.39 LB +0.6R)	MEASUREMENT UNCERTAINTY R=RESOLUTION
---	--------------------------------------	--	--------------------------------------

TECH REMARKS	TECH REMARKS
TEST AND CALIBRATE CEMENT SCALE.	TEST AND CALIBRATE WATER SCALE.

PARTS SUPPLIED	Minimum - 2 hours (Labor - 2 Men)
Ø	Tr. In - 2:15 - 3:30
	Work - 3:30 - 5:00
	Tr. Out - 5:00 - 6:00
	Mileage 88 Truck # 11

These Results Relate Only to Items Tested Scales meet NIST standard/Handbook 44 Scales fail to meet NIST standards	1 <input type="checkbox"/> 2 <input checked="" type="checkbox"/>	WORK ORDERED BY DAN THOMAS
TECHNICIAN SIGNATURE <i>Mike - 50E</i>	CUSTOMER SIGNATURE <i>[Signature]</i>	