



29540 Calahan • Roseville, MI 48066
 888.776.0517 • Fax 586.776.3169
 www.greatlakescale.com

SALES | SERVICE | RENTALS

MASTER DISTRIBUTORS OF:
 CARDINAL • SETRA • GSE • OHAUS
 AVERY WEIGHTRONIX • RICELAKE

ISO/IEC 17025 ACCREDITED
 STATE REGISTERED SERVICE AGENCY

6186



CALIBRATION CERTIFICATE / PROPORTIONING REPORT

CUSTOMER Superior #14
 PHONE 810-560-3539 DATE 1-23-17

LOCATION Lansing Plant
 PO NUMBER

SCALE #1

SCALE #2

MAKE Alkon TEMPERATURE 400F
 MODEL Command S/N WEIGHTS USED:
 SERIAL # 9678 TRUCK 22
 CAPACITY 10,000 LB
 RESOLUTION X10
 LOCATION Cement

MAKE Alkon TEMPERATURE 400F
 MODEL Command S/N WEIGHTS USED:
 SERIAL # 9678 TRUCK 22
 CAPACITY 40,000 LB
 RESOLUTION X 70 LB
 LOCATION AG6

WT APPLIED	AS FOUND	WT APPLIED	AS LEFT	WT APPLIED	AS FOUND	WT APPLIED	AS LEFT
0	0			0	0	WT 6000	38000
2000	2000			2000	2000		
4000	4000			4000	4000		
6000	6000			6000	6000		
9000	4000			4000	4000		
2000	2000			2000	2000		
0	0			0	0		
WT 3200	3200			WT 8000	8000		
WT 6000	9200			WT 6000	14000		

SHIFT TEST		SHIFT TEST		SHIFT TEST		SHIFT TEST	
WT APPLIED	AS FOUND	WT APPLIED	AS LEFT	WT APPLIED	AS FOUND	WT APPLIED	AS LEFT
1		1		1	WT 1600	1600	1
2		2		2	WT 16000	22000	2
3		3		3	WT 24000	24000	3
4		4		4	WT 16000	30000	4
5		5		5	WT 32000	32000	5

MEASUREMENT UNCERTAINTY $\pm (.68 LB + 0.6R)$ R=RESOLUTION
 MEASUREMENT UNCERTAINTY $\pm (3.0 LB + 0.6R)$ R=RESOLUTION

TECH REMARKS
Tested scale tested scale

PARTS SUPPLIED

Minimum - 2 hours (Labor - 2 Men)

Tr. In -
 Work - see tablet
 Tr. Out
 Mileage _____ Truck # 22

These Results Relate Only to Items Tested
 Scales meet NIST standard/Handbook 44
 Scales fail to meet NIST standards

1
 2

WORK ORDERED BY
Dan Thomas

TECHNICIAN SIGNATURE
James Moore Joe

CUSTOMER SIGNATURE
[Signature]

GLS-QMS-FM-004-REV1 Refer to State Test Report for weights applied
 This report may not be reproduced except in full, without the written approval of Great Lakes Scale Company.
 Best measurement or uncertainty for weights used during calibration is on our scope of accreditation, which we will provide upon request.