



29540 Calahan • Roseville, MI 48066  
888.776.0517 • Fax 586.776.3169  
www.greatlakescale.com

### SALES | SERVICE | RENTALS

MASTER DISTRIBUTORS OF:  
CARDINAL • SETRA • GSE • OHAUS  
AVERY WEIGHTRONIX • RICELAKE

ISO/IEC 17025 ACCREDITED  
STATE REGISTERED SERVICE AGENCY

6089



### CALIBRATION CERTIFICATE / PROPORTIONING REPORT

CUSTOMER Superior plant 202  
PHONE 248-345-1528 DATE 12-8-16

LOCATION S300 N Dort Hwy Flat MI 48305  
PO NUMBER

#### SCALE #1 SCALE #2

MAKE <u>Alkon</u>	TEMPERATURE <u>30</u>	MAKE <u>Alkon</u>	TEMPERATURE <u>30</u>
MODEL <u>Command</u>	S/N WEIGHTS USED:	MODEL <u>Command</u>	S/N WEIGHTS USED:
SERIAL # <u>62822</u>	<u>Truck 22</u>	SERIAL # <u>62822</u>	<u>Truck 22</u>
CAPACITY <u>10,000 LB</u>		CAPACITY <u>40,000 LB</u>	
RESOLUTION <u>5 LB</u>		RESOLUTION <u>20 LB</u>	
LOCATION <u>Cement</u>		LOCATION <u>AGG</u>	

WT APPLIED	AS FOUND	WT APPLIED	AS LEFT	WT APPLIED	AS FOUND	WT APPLIED	AS LEFT
0	0			0	0		
2000	2000			2000	2000		
4000	4000			4000	4000		
6000	6000			6000	6000		
8000	8000			8000	8000		
10000	10000			10000	10000		
0	0			0	0		
WT 3400	3400		WT	10,000	10,000		
WT 6000	6000		WT	6,000	6,000		

SHIFT TEST		SHIFT TEST		SHIFT TEST		SHIFT TEST	
WT APPLIED	AS FOUND	WT APPLIED	AS LEFT	WT APPLIED	AS FOUND	WT APPLIED	AS LEFT
1		1		1	70000	20000LB	1
2		2		2	10000	26000LB	2
3		3		3	30000	20000LB	3
4		4		4	10000	36000LB	4
5		5		5			5

MEASUREMENT UNCERTAINTY  $\pm(0.06 + 0.6R)$  R=RESOLUTION  
 MEASUREMENT UNCERTAINTY  $\pm(3.0 + 0.6R)$  R=RESOLUTION

TECH REMARKS Test scale

#### PARTS SUPPLIED

Minimum - 2 hours Labor - 2 Men  
 Tr. In -  
 Work - See tablet  
 Tr. Out  
 Mileage \_\_\_\_\_ Truck # 72

These Results Relate Only to Items Tested  
 Scales meet NIST standard/Handbook 44  
 Scales fail to meet NIST standards

1  2   
 WORK ORDERED BY Dan Thomas

TECHNICIAN SIGNATURE James Moore Joe James CUSTOMER SIGNATURE William Barber