



GREAT LAKES SCALE COMPANY
 35380 UNION LAKE RD.
 HARRISON TOWNSHIP, MI 48045
 Phone: (586) 776-0517

Certificate of Calibration

Customer: SUPERIOR MATERIALS 36
 SUP1192 2470 AUBURN RD

Certificate ID: CBS62803B191126045749

Test Date: 11/26/2019

AUBURN HILLS, MI

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: CBS62803B Capacity: 38000 lb Graduation: 20 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
11/26/2019	05/27/2020	Semi-Annually	CBS62803B	III L
SCALE LOCATION		EQUIPMENT CONDITIONS		
AGG B		Working	Clean	
SELECT NUMBER OF PM DAYS:		CAPACITY	38000 UOM lb	INCREMENT 20

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS						AS LEFT TEST USING TEST WEIGHTS					
Weight Applied		Reading		Error		Weight Applied		Reading		Error	
0	lb	0	lb	0	lb		lb		lb		lb
1000	lb	1000	lb	0	lb		lb		lb		lb
2000	lb	2000	lb	0	lb		lb		lb		lb
3000	lb	3000	lb	0	lb		lb		lb		lb
4000	lb	4000	lb	0	lb		lb		lb		lb
5000	lb	5000	lb	0	lb		lb		lb		lb
6000	lb	6000	lb	0	lb		lb		lb		lb
0	lb	0	lb	0	lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb

AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
0	lb	0	lb
2000	lb	8000	lb
4000	lb	8000	lb
6000	lb	8000	lb
2000	lb	16000	lb
4000	lb	16000	lb
6000	lb	16000	lb
2000	lb	24000	lb
4000	lb	24000	lb
6000	lb	24000	lb
2000	lb	30000	lb
4000	lb	30000	lb
6000	lb	30000	lb
	lb		lb
	lb		lb
	lb		lb
	lb		lb
	lb		lb
	lb		lb

OVERALL TEST RESULTS PASS



GREAT LAKES SCALE COMPANY
 35380 UNION LAKE RD.
 HARRISON TOWNSHIP, MI 48045
 Phone: (586) 776-0517

Certificate of Calibration

Customer: SUPERIOR MATERIALS 36
 SUP1192 2470 AUBURN RD

Certificate ID: CBS62803B191126045749

Test Date: 11/26/2019

AUBURN HILLS, MI

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: CBS62803B Capacity: 38000 lb Graduation: 20 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
11/26/2019	05/27/2020	Semi-Annually	CBS62803B	III L
SCALE LOCATION		EQUIPMENT CONDITIONS		
AGG B		Working	Clean	
SELECT NUMBER OF PM DAYS:	CAPACITY	38000	UOM	lb
			INCREMENT	20

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS						AS LEFT TEST USING TEST WEIGHTS					
Weight Applied		Reading		Error		Weight Applied		Reading		Error	
0	lb	0	lb	0	lb		lb		lb		lb
1000	lb	1000	lb	0	lb		lb		lb		lb
2000	lb	2000	lb	0	lb		lb		lb		lb
3000	lb	3000	lb	0	lb		lb		lb		lb
4000	lb	4000	lb	0	lb		lb		lb		lb
5000	lb	5000	lb	0	lb		lb		lb		lb
6000	lb	6000	lb	0	lb		lb		lb		lb
0	lb	0	lb	0	lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb

AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
0	lb	0	lb
2000	lb	8000	lb
4000	lb	8000	lb
6000	lb	8000	lb
2000	lb	16000	lb
4000	lb	16000	lb
6000	lb	16000	lb
2000	lb	24000	lb
4000	lb	24000	lb
6000	lb	24000	lb
2000	lb	30000	lb
4000	lb	30000	lb
6000	lb	30000	lb
	lb		lb
	lb		lb
	lb		lb
	lb		lb
	lb		lb
	lb		lb

OVERALL TEST RESULTS PASS



GREAT LAKES SCALE COMPANY
 35380 UNION LAKE RD.
 HARRISON TOWNSHIP, MI 48045
 Phone: (586) 776-0517

Certificate of Calibration

Customer: SUPERIOR MATERIALS 36
 SUP1192 2470 AUBURN RD

Certificate ID: CBS62803C191126045844

Test Date: 11/26/2019

AUBURN HILLS, MI

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: CBS62803C Capacity: 6700 lb Graduation: 5 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
11/26/2019	05/27/2020	Semi-Annually	CBS62803C	III
SCALE LOCATION		EQUIPMENT CONDITIONS		
CEMENT		Working	Clean	
SELECT NUMBER OF PM DAYS:		CAPACITY	6700 UOM lb	INCREMENT 5

AS FOUND TEST USING TEST WEIGHTS

Weight Applied	Reading	Error
0	0	0
1000	1000	0
2000	1995	-5
3000	2990	-10
4000	3985	-15
5000	4980	-20
6000	5975	-25
0	0	0

AS LEFT TEST USING TEST WEIGHTS

Weight Applied	Reading	Error
0	0	0
1000	1000	0
2000	2000	0
3000	3000	0
4000	4000	0
5000	5000	0
6000	6000	0
0	0	0

AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb

OVERALL TEST RESULTS N/A

Customer: SUPERIOR MATERIALS 36
 SUP1192 2470 AUBURN RD

AUBURN HILLS, MI

Certificate ID: CBS62803C191126045844

Test Date: 11/26/2019

ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	3.68
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	75
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

Comments:

Scale on zero, load tested, calibrated, retested and sealed.

TRACEABILITY CERTIFICATE NUMBERS:

Kit	NIST	Date	Date Due
Description			
GL1502	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1503	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1505	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1506	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1504	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1517	MI-06-19-14306	6/26/2019	6/26/2021
TRUCK 15 - 1000 LB			

Technician(s): RAY MCCREARY; Edward G Gorham



Signature:

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.

Customer: SUPERIOR MATERIALS 36
 SUP1192 2470 AUBURN RD

AUBURN HILLS, MI

Certificate ID: CBS62803B191126045749

Test Date: 11/26/2019

ISO Number: MS-FM-CEM-001


TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	15
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	38
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

Comments:

TRACEABILITY CERTIFICATE NUMBERS:			
Kit	NIST	Date	Date Due
Description			
GL1502	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1503	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1505	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1506	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1504	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1517	MI-06-19-14306	6/26/2019	6/26/2021
TRUCK 15 - 1000 LB			

Technician(s): RAY MCCREARY; Ed G.



Signature:

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.

Customer: SUPERIOR MATERIALS 36
 SUP1192 2470 AUBURN RD

AUBURN HILLS, MI

Certificate ID: CBS62803B191126045749

Test Date: 11/26/2019

ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	15
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	38
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

Comments:

TRACEABILITY CERTIFICATE NUMBERS:			
Kit	NIST	Date	Date Due
Description			
GL1502	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1503	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1505	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1506	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1504	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1517	MI-06-19-14306	6/26/2019	6/26/2021
TRUCK 15 - 1000 LB			

Technician(s): RAY MCCREARY; Ed G.



Signature:

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.