

WORK ORDER

Great Lakes Scale

35380 UNION LAKE RD
 HARRISON TOWNSHIP, MI 48045
 Phone: (586)776-0517
 Fax: (586)776-3169
 Fed Tax ID: 38-3206225

Work Order# 0027424-1
Request Date 11/14/2019 9:20:00 AM
Schedule Date 11/19/2019 7:50:00 AM

Bill To SUP0896
 SUP0896
 SUPERIOR MATERIALS HOLDINGS LLC
 PO BOX 2900
 FARMINGTON HILLS, MI, 48336
 (248) 788-8000

Ship To SUP1193
 SUP1193
 SUPERIOR MATERIALS 38
 901 S WAGNER RD
 ANN ARBOR, MI, 48103
 (248) 788-8000

Customer P.O
Requester DAN THOMAS
Phone (248) 788-8048
Email dehthomas@superiormaterials.net

CONTRACT TYPE: No Contract

DESCRIPTION:

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK - 8am	

Service Date 11/19/2019

2 hours minimum portal to portal

WORK PERFORMED
 CALL COMPLETE Test and calibrate cement scale, test aggregate scale.

TECHNICIAN(S)
 MATT DIMOND, MIKE GLADSTONE

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
Agg	102559agg	102559agg	ALKON SCALE/INDICATOR / COMMAND	
10000 lbs x 5 lbs	102559cem	102559cem	ALKON SCALE/INDICATOR / COMMAND	

PARTS LIST

--

Customer Signature

 Name:



GREAT LAKES SCALE COMPANY
 35380 UNION LAKE RD.
 HARRISON TOWNSHIP, MI 48045
 Phone: (586) 776-0517

Certificate of Calibration

Customer: SUPERIOR MATERIALS 38
 SUP1193 901 S WAGNER RD

Certificate ID: 102559cem191119092136

Test Date: 11/19/2019

ANN ARBOR, MI 48103

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: 102559cem Capacity: 10000 lb Graduation: 5 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
11/19/2019	05/20/2020	Semi-Annually	102559cem	III
SCALE LOCATION		EQUIPMENT CONDITIONS		
		Working	Clean	
SELECT NUMBER OF PM DAYS:		CAPACITY	10000 UOM lb	INCREMENT 5

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS						AS LEFT TEST USING TEST WEIGHTS					
Weight Applied	lb	Reading	lb	Error	lb	Weight Applied	lb	Reading	lb	Error	lb
0	lb	0	lb	0	lb	0	lb	0	lb	0	lb
1000	lb	1005	lb	5	lb	1000	lb	1000	lb	0	lb
2000	lb	2005	lb	5	lb	2000	lb	2005	lb	5	lb
3000	lb	3005	lb	5	lb	3000	lb	3005	lb	5	lb
4000	lb	4010	lb	10	lb	4000	lb	4005	lb	5	lb
5000	lb	5010	lb	10	lb	5000	lb	5000	lb	0	lb
6000	lb	6010	lb	10	lb	6000	lb	6000	lb	0	lb
0	lb	0	lb	0	lb	0	lb	0	lb	0	lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb

AS LEFT BUILD TEST

Test Weight Applied	lb	Material Weight Applied	lb	Reading	lb	Error	lb
6000	lb	3700	lb	9700	lb	0	lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb

OVERALL TEST RESULTS PASS



GREAT LAKES SCALE COMPANY
 35380 UNION LAKE RD.
 HARRISON TOWNSHIP, MI 48045
 Phone: (586) 776-0517

Certificate of Calibration

Customer: SUPERIOR MATERIALS 38
 SUP1193 901 S WAGNER RD

Certificate ID: 102559agg191119092438

Test Date: 11/19/2019

ANN ARBOR, MI 48103

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: 102559agg Capacity: 40000 lb Graduation: 20 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
11/19/2019	05/20/2020	Semi-Annually	102559agg	III L
SCALE LOCATION		EQUIPMENT CONDITIONS		
		Working	Clean	
SELECT NUMBER OF PM DAYS:	CAPACITY	40000 UOM	lb	INCREMENT 20

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS						AS LEFT TEST USING TEST WEIGHTS					
Weight Applied	Reading		Error			Weight Applied	Reading		Error		
0	lb	0	lb	0	lb	0	lb	0	lb	0	lb
1000	lb	1000	lb	0	lb	1000	lb	1000	lb	0	lb
2000	lb	2000	lb	0	lb	2000	lb	2000	lb	0	lb
3000	lb	3000	lb	0	lb	3000	lb	3000	lb	0	lb
4000	lb	4000	lb	0	lb	4000	lb	4000	lb	0	lb
5000	lb	5000	lb	0	lb	5000	lb	5000	lb	0	lb
6000	lb	6000	lb	0	lb	6000	lb	6000	lb	0	lb
0	lb	0	lb	0	lb	0	lb	0	lb	0	lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb

AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
0	lb 0	lb 0	lb 0
2000	lb 10000	lb 12000	lb 0
4000	lb 10000	lb 14000	lb 0
6000	lb 10000	lb 16000	lb 0
2000	lb 20000	lb 22000	lb 0
4000	lb 20000	lb 24000	lb 0
6000	lb 20000	lb 26000	lb 0
2000	lb 30000	lb 32000	lb 0
4000	lb 30000	lb 34000	lb 0
6000	lb 30000	lb 36000	lb 0
	lb	lb	lb
	lb	lb	lb
	lb	lb	lb
	lb	lb	lb
	lb	lb	lb
	lb	lb	lb
	lb	lb	lb
	lb	lb	lb
	lb	lb	lb
	lb	lb	lb

OVERALL TEST RESULTS PASS

WORK ORDER

Great Lakes Scale

35380 UNION LAKE RD
 HARRISON TOWNSHIP, MI 48045
 Phone: (586)776-0517
 Fax: (586)776-3169
 Fed Tax ID: 38-3206225

Work Order# 0027424-1
Request Date 11/14/2019 9:20:00 AM
Schedule Date 11/19/2019 7:50:00 AM

Bill To SUP0896
 SUP0896
 SUPERIOR MATERIALS HOLDINGS LLC
 PO BOX 2900
 FARMINGTON HILLS, MI, 48336
 (248) 788-8000

Ship To SUP1193
 SUP1193
 SUPERIOR MATERIALS 38
 901 S WAGNER RD
 ANN ARBOR, MI, 48103
 (248) 788-8000

Customer P.O
Requester DAN THOMAS
Phone (248) 788-8048
Email dehthomas@superiormaterials.net

CONTRACT TYPE: No Contract

DESCRIPTION:

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK - 8am	

Service Date 11/19/2019

2 hours minimum portal to portal

WORK PERFORMED
 CALL COMPLETE Test and calibrate cement scale, test aggregate scale.

TECHNICIAN(S)
 MATT DIMOND, MIKE GLADSTONE

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
Agg	102559agg	102559agg	ALKON SCALE/INDICATOR / COMMAND	
10000 lbs x 5 lbs	102559cem	102559cem	ALKON SCALE/INDICATOR / COMMAND	

PARTS LIST

--

Customer Signature



Steve M Ratliff

Customer: SUPERIOR MATERIALS 38
 SUP1193 901 S WAGNER RD

Certificate ID: 102559agg191119092438

Test Date: 11/19/2019

ANN ARBOR, MI 48103

ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	15
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	65
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment


Comments:

Test scale,sealed.

TRACEABILITY CERTIFICATE NUMBERS:			
Kit	NIST	Date	Date Due
Description			
GL1502	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1503	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1504	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1505	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1506	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1517	MI-06-19-14306	6/26/2019	6/26/2021
TRUCK 15 - 1000 LB			

Technician(s): MIKE GLADSTONE;Matt Dimond

Signature:



Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.

Customer: SUPERIOR MATERIALS 38
 SUP1193 901 S WAGNER RD

ANN ARBOR, MI 48103

Certificate ID: 102559cem191119092136

Test Date: 11/19/2019

ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	3.68
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	65
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass adjustment needed

Comments:

Test and calibrate scale,sealed.

TRACEABILITY CERTIFICATE NUMBERS:				
Kit	NIST	Date	Date Due	
Description				
GL1502	MI-03-19-14172	3/19/2019	3/19/2021	
TRUCK 15 - 1000 LB				
GL1503	MI-03-19-14172	3/19/2019	3/19/2021	
TRUCK 15 - 1000 LB				
GL1504	MI-03-19-14172	3/19/2019	3/19/2021	
TRUCK 15 - 1000 LB				
GL1505	MI-03-19-14172	3/19/2019	3/19/2021	
TRUCK 15 - 1000 LB				
GL1506	MI-03-19-14172	3/19/2019	3/19/2021	
TRUCK 15 - 1000 LB				
GL1517	MI-06-19-14306	6/26/2019	6/26/2021	
TRUCK 15 - 1000 LB				

Technician(s): MIKE GLADSTONE;Matt Dimond

Signature:



Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.