

WORK ORDER

Great Lakes Scale

35380 UNION LAKE RD
 HARRISON TOWNSHIP, MI 48045
 Phone: (586)776-0517
 Fax: (586)776-3169
 Fed Tax ID: 38-3206225

Work Order# 0027354-1
Request Date 11/1/2019 2:08:00 PM
Schedule Date 11/4/2019 9:00:00 AM

Bill To SUP0896
 SUP0896
 SUPERIOR MATERIALS HOLDINGS LLC
 PO BOX 2900
 FARMINGTON HILLS, MI, 48336
 (248) 788-8000

Ship To SUP1173
 SUP1173
 SUPERIOR MATERIALS 32
 8911 WEST JEFFERSON
 DETROIT, MI,
 (248) 788-8000

Customer P.O
Requester DAN THOMAS
Phone (248) 788-8048
Email dehthomas@superiormaterials.net

CONTRACT TYPE: No Contract

DESCRIPTION:

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK WAYNE COUNTY 9 AM	

Service Date 11/4/2019

2 hours minimum portal to portal

WORK PERFORMED
 CALL COMPLETE
 CALIBRATIONS COMPLETE. Test scales with county inspector. Couldn't start right away due to customer loading trucks. Then Hilo ran out of gas. Floors had to clean to move weighs around. Agg was bad. Maint helped clean floor for agg. Setup on cement scale to check scale and belt guard was in way to drop chains. Customer told us to move to water so they could remove guard. Moved all equipment to water. Customer loaded trucks between scale checks. Had to wait to remove and install equipment. Cement was heavy customer recalibrated scale.

TECHNICIAN(S)
 JOE BIELARZ, RAY MCCREARY

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
10,000 x 5 lb	62821 CEMENT	62821 CEMENT	ALKON SCALE/INDICATOR / COMMAND	
6000 x 1	662821 WATER	662821 WATER	ALKON SCALE/INDICATOR / COMMAND	
40,000 x 20 lb	662821 AGG A	662821 AGG A	ALKON SCALE/INDICATOR / COMMAND	AGG A

PARTS LIST

Customer Signature

 Name:



GREAT LAKES SCALE COMPANY
 35380 UNION LAKE RD.
 HARRISON TOWNSHIP, MI 48045
 Phone: (586) 776-0517

Certificate of Calibration

Customer: SUPERIOR MATERIALS 32
 SUP1173 8911 WEST JEFFERSON

Certificate ID: 62821 CEMENT191104121113

Test Date: 11/04/2019

DETROIT, MI

ISO Number: MS-FM-CEM-001

(*Make: ALKON Model : COMMAND SN#: 62821 CEMENT Capacity: 10000 lb Graduation: 5 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
11/04/2019	05/05/2020	Semi-Annually	62821 CEMENT	III
SCALE LOCATION		EQUIPMENT CONDITIONS		
		Working	Clean	
SELECT NUMBER OF PM DAYS:	CAPACITY	10000 UOM	lb	INCREMENT
				5

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS					AS LEFT TEST USING TEST WEIGHTS				
Weight Applied		Reading		Error	Weight Applied		Reading		Error
0	lb	0	lb	0	0	lb	0	lb	0
2000	lb	2005	lb	5	6000	lb	6000	lb	0
4000	lb	4010	lb	10	0	lb	0	lb	0
6000	lb	6015	lb	15	1000	lb	1000	lb	0
	lb		lb		2000	lb	2000	lb	0
	lb		lb		3000	lb	3000	lb	0
	lb		lb		4000	lb	4000	lb	0
	lb		lb		5000	lb	5000	lb	0
	lb		lb		6000	lb	6000	lb	0
	lb		lb		0	lb	0	lb	0

AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
6000	3300	9300	0

OVERALL TEST RESULTS PASS



Certificate of Calibration

Customer: SUPERIOR MATERIALS 32
 SUP1173 8911 WEST JEFFERSON

Certificate ID: 662821 WATER191104120829

Test Date: 11/04/2019

DETROIT, MI

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: 662821 WATER Capacity: 6000 lb Graduation: 1 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS		
11/04/2019	05/05/2020	Semi-Annually	662821 WATER	III		
SCALE LOCATION			EQUIPMENT CONDITIONS			
			Working	Clean		
SELECT NUMBER OF PM DAYS:		CAPACITY	6000 UOM	lb	INCREMENT	1

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS						AS LEFT TEST USING TEST WEIGHTS					
Weight Applied		Reading		Error		Weight Applied		Reading		Error	
0	lb	0	lb	0	lb	0	lb	0	lb	0	lb
6000	lb	6000	lb	0	lb	1000	lb	1000	lb	0	lb
0	lb	0	lb	0	lb	2000	lb	2000	lb	0	lb
	lb		lb		lb	3000	lb	3000	lb	0	lb
	lb		lb		lb	4000	lb	4000	lb	0	lb
	lb		lb		lb	5000	lb	5000	lb	0	lb
	lb		lb		lb	6000	lb	6000	lb	0	lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb

AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
	lb		lb
	lb		lb
	lb		lb
	lb		lb
	lb		lb
	lb		lb
	lb		lb
	lb		lb
	lb		lb
	lb		lb
	lb		lb
	lb		lb
	lb		lb
	lb		lb
	lb		lb
	lb		lb
	lb		lb
	lb		lb
	lb		lb
	lb		lb
	lb		lb
	lb		lb
	lb		lb

OVERALL TEST RESULTS N/A



GREAT LAKES SCALE COMPANY
 35380 UNION LAKE RD.
 HARRISON TOWNSHIP, MI 48045
 Phone: (586) 776-0517

Certificate of Calibration

Customer: SUPERIOR MATERIALS 32
 SUP1173 8911 WEST JEFFERSON

Certificate ID: 662821 AGG A191104120603

Test Date: 11/04/2019

DETROIT, MI

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: 662821 AGG A Capacity: 40000 LB Graduation: 20 LB

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
11/04/2019	05/05/2020	Semi-Annually	662821 AGG A	III L
SCALE LOCATION		EQUIPMENT CONDITIONS		
AGG A		Working	Clean	
SELECT NUMBER OF PM DAYS:		CAPACITY	40000 UOM	LB INCREMENT 20

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

Weight Applied	Reading	Error	Weight Applied	Reading	Error
0	LB 0	LB 0		LB	LB
6000	LB 6000	LB 0		LB	LB
0	LB 0	LB 0		LB	LB
1000	LB 1000	LB 0		LB	LB
2000	LB 2000	LB 0		LB	LB
3000	LB 3000	LB 0		LB	LB
4000	LB 4000	LB 0		LB	LB
5000	LB 5000	LB 0		LB	LB
6000	LB 6000	LB 0		LB	LB
0	LB 0	LB 0		LB	LB

AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
2000	LB 6000	LB 8000	LB 0
4000	LB 6000	LB 10000	LB 0
6000	LB 6000	LB 12000	LB 0
2000	LB 12000	LB 14000	LB 0
4000	LB 12000	LB 16000	LB 0
6000	LB 12000	LB 18000	LB 0
2000	LB 18000	LB 20000	LB 0
4000	LB 18000	LB 22000	LB 0
6000	LB 18000	LB 24000	LB 0
2000	LB 24000	LB 26000	LB 0
4000	LB 24000	LB 28000	LB 0
6000	LB 24000	LB 30000	LB 0
2000	LB 32000	LB 34000	LB 0
4000	LB 32000	LB 36000	LB 0
6000	LB 32000	LB 38000	LB 0
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB

OVERALL TEST RESULTS PASS

WORK ORDER

Great Lakes Scale

35380 UNION LAKE RD
 HARRISON TOWNSHIP, MI 48045
 Phone: (586)776-0517
 Fax: (586)776-3169
 Fed Tax ID: 38-3206225

Work Order# 0027354-1

Request Date 11/1/2019 2:08:00 PM

Schedule Date 11/4/2019 9:00:00 AM

Bill To SUP0896

SUP0896
 SUPERIOR MATERIALS HOLDINGS LLC
 PO BOX 2900
 FARMINGTON HILLS, MI, 48336
 (248) 788-8000

Ship To SUP1173

SUP1173
 SUPERIOR MATERIALS 32
 8911 WEST JEFFERSON
 DETROIT, MI,
 (248) 788-8000

Customer P.O

Requester DAN THOMAS

Phone (248) 788-8048

Email dehthomas@superiormaterials.net

CONTRACT TYPE: No Contract

DESCRIPTION:

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK WAYNE COUNTY 9 AM	

Service Date 11/4/2019

2 hours minimum portal to portal


WORK PERFORMED
 CALL COMPLETE
 CALIBRATIONS COMPLETE. Test scales with county inspector. Couldn't start right away due to customer loading trucks. Then Hilo ran out of gas. Floors had to clean to move weighs around. Agg was bad. Maint helped clean floor for agg. Setup on cement scale to check scale and belt guard was in way to drop chains. Customer told us to move to water so they could remove guard. Moved all equipment to water. Customer loaded trucks between scale checks. Had to wait to remove and install equipment. Cement was heavy customer recalibrated scale.

TECHNICIAN(S)
 JOE BIELARZ, RAY MCCREARY

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
10,000 x 5 lb	62821 CEMENT	62821 CEMENT	ALKON SCALE/INDICATOR / COMMAND	
6000 x 1	662821 WATER	662821 WATER	ALKON SCALE/INDICATOR / COMMAND	
40,000 x 20 lb	662821 AGG A	662821 AGG A	ALKON SCALE/INDICATOR / COMMAND	AGG A

PARTS LIST

Customer Signature



Andrew

Customer: SUPERIOR MATERIALS 32
 SUP1173 8911 WEST JEFFERSON

DETROIT, MI

Certificate ID: 662821 AGG A191104120603

Test Date: 11/04/2019

ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	15
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	55
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

Comments:

TRACEABILITY CERTIFICATE NUMBERS:			
Kit	NIST	Date	Date Due
Description			
GL0601	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0602	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0603	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0604	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0605	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0606	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			

Technician(s): RAY MCCREARY; Joe bielarz

Signature:



Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.

Customer: SUPERIOR MATERIALS 32
 SUP1173 8911 WEST JEFFERSON

DETROIT, MI

Certificate ID: 662821 WATER191104120829

Test Date: 11/04/2019

ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	1.28
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	55
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

Comments:

TRACEABILITY CERTIFICATE NUMBERS:			
Kit	NIST	Date	Date Due
Description			
GL0601	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0602	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0603	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0604	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0605	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0606	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			

Technician(s): RAY MCCREARY; Joe B.

Signature:



Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.

Customer: SUPERIOR MATERIALS 32
 SUP1173 8911 WEST JEFFERSON

DETROIT, MI

Certificate ID: 62821 CEMENT191104121113

Test Date: 11/04/2019

ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	3.39
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	55
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass adjustment needed

Comments:

Test and cal scale with county inspector

TRACEABILITY CERTIFICATE NUMBERS:

Kit	NIST	Date	Date Due
Description			
GL0601	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0602	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0603	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0604	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0605	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0606	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			

Technician(s): RAY MCCREARY; Joe B

Signature:



Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.