



**Customer:** SUPERIOR MATERIALS 103  
 SUP0992 10147 GRAND RIVER

**Certificate ID:** 23635-a200218021910

**Test Date:** 02/18/2020

BRIGHTON, MI 48116

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	3.68
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	64
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass adjustment needed

**Comments:**

Found scale on zero load tested and sealed.

**TRACEABILITY CERTIFICATE NUMBERS:**

Kit	NIST	Date	Date Due
Description			
GL1502	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1503	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1504	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1505	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1506	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1517	MI-06-19-14306	6/26/2019	6/26/2021
TRUCK 15 - 1000 LB			

**Technician(s):** ED GORHAM;RAY MCCREARY;



**Signature:**

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.