



GREAT LAKES SCALE COMPANY
 35380 UNION LAKE RD.
 HARRISON TOWNSHIP, MI 48045
 Phone: (586) 776-0517

Certificate of Calibration

Customer: SUPERIOR MATERIALS 2
 SUP1048 20565 HOOVER

Certificate ID: 72113-a200806111806

Test Date: 08/06/2020

DETROIT, MI

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: 72113-a Capacity: 10000 LB Graduation: 5 LB

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
08/06/2020	02/05/2021	Semi-Annually	72113-a	III
SCALE LOCATION		EQUIPMENT CONDITIONS		
CEMENT SCALE		Working	Clean	
SELECT NUMBER OF PM DAYS:	CAPACITY	10000 UOM	LB	INCREMENT 5

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS			AS LEFT TEST USING TEST WEIGHTS		
Weight Applied	Reading	Error	Weight Applied	Reading	Error
0	LB 0	LB 0	0	LB 0	LB 0
6000	LB 6000	LB 0	1000	LB 1000	LB 0
0	LB 0	LB 0	2000	LB 2000	LB 0
	LB	LB	3000	LB 3000	LB 0
	LB	LB	4000	LB 4000	LB 0
	LB	LB	5000	LB 5000	LB 0
	LB	LB	6000	LB 6000	LB 0
	LB	LB		LB	LB
	LB	LB		LB	LB
	LB	LB		LB	LB
	LB	LB		LB	LB

AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
6000	LB 3600	LB 9600	LB 0
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB

OVERALL TEST RESULTS PASS

Customer: SUPERIOR MATERIALS 2
 SUP1048 20565 HOOVER

DETROIT, MI

Certificate ID: 72113-a200806111806

Test Date: 08/06/2020

ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	3.68
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	75
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass adjustment needed

Comments:

TRACEABILITY CERTIFICATE NUMBERS:			
Kit	NIST	Date	Date Due
Description			
GL1502	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1503	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1504	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1505	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1506	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1517	MI-06-19-14306	6/26/2019	6/26/2021
TRUCK 15 - 1000 LB			

Technician(s): JOE BIELARZ;MIKE GLADSTONE;

Signature:



Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.



GREAT LAKES SCALE COMPANY
 35380 UNION LAKE RD.
 HARRISON TOWNSHIP, MI 48045
 Phone: (586) 776-0517

Certificate of Calibration

Customer: SUPERIOR MATERIALS 2
 SUP1048 20565 HOOVER

Certificate ID: 72113-b200806111936

Test Date: 08/06/2020

DETROIT, MI

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: 72113-b Capacity: 40000 LB Graduation: 40 LB

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
08/06/2020	02/05/2021	Semi-Annually	72113-b	III L
SCALE LOCATION		EQUIPMENT CONDITIONS		
AGG SCALE		Working	Clean	
SELECT NUMBER OF PM DAYS:		CAPACITY	40000 UOM	LB INCREMENT 40

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS						AS LEFT TEST USING TEST WEIGHTS					
Weight Applied		Reading		Error		Weight Applied		Reading		Error	
0	LB	0	LB	0	LB		LB		LB		LB
1000	LB	1000	LB	0	LB		LB		LB		LB
2000	LB	2000	LB	0	LB		LB		LB		LB
3000	LB	3000	LB	0	LB		LB		LB		LB
4000	LB	4000	LB	0	LB		LB		LB		LB
5000	LB	5000	LB	0	LB		LB		LB		LB
6000	LB	6000	LB	0	LB		LB		LB		LB
0	LB	0	LB	0	LB		LB		LB		LB
	LB		LB		LB		LB		LB		LB
	LB		LB		LB		LB		LB		LB

AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
2000	LB 6000	LB 8000	LB 0
4000	LB 6000	LB 10000	LB 0
6000	LB 6000	LB 12000	LB 0
2000	LB 12000	LB 14000	LB 0
4000	LB 12000	LB 16000	LB 0
6000	LB 12000	LB 18000	LB 0
2000	LB 18000	LB 20000	LB 0
4000	LB 18000	LB 22000	LB 0
6000	LB 18000	LB 24000	LB 0
2000	LB 24000	LB 26000	LB 0
4000	LB 24000	LB 28000	LB 0
6000	LB 24000	LB 30000	LB 0
2000	LB 32000	LB 34000	LB 0
4000	LB 32000	LB 36000	LB 0
6000	LB 32000	LB 38000	LB 0
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB

OVERALL TEST RESULTS PASS

Customer: SUPERIOR MATERIALS 2
 SUP1048 20565 HOOVER

DETROIT, MI

Certificate ID: 72113-b200806111936

Test Date: 08/06/2020

ISO Number: MS-FM-CEM-001

TEST INFORMATION


TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	27
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	75
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

Comments:

TRACEABILITY CERTIFICATE NUMBERS:			
Kit	NIST	Date	Date Due
Description			
GL1502	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1503	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1504	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1505	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1506	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1517	MI-06-19-14306	6/26/2019	6/26/2021
TRUCK 15 - 1000 LB			

Technician(s): JOE BIELARZ;MIKE GLADSTONE;

Signature:



Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.

WORK ORDER

Great Lakes Scale

35380 UNION LAKE RD
 HARRISON TOWNSHIP, MI 48045
 Phone: (586)776-0517
 Fax: (586)776-3169
 Fed Tax ID: 38-3206225

Work Order# 0028780-1

Request Date 7/24/2020 10:32:00 AM

Schedule Date 8/6/2020 9:01:00 AM

Bill To SUP0896

SUP0896
 SUPERIOR MATERIALS HOLDINGS LLC
 PO BOX 2900
 FARMINGTON HILLS, MI, 48336
 (248) 788-8000

Ship To SUP1048

SUP1048
 SUPERIOR MATERIALS 2
 20565 HOOVER
 DETROIT, MI,
 (248) 788-8000

Customer P.O

Requester DAN THOMAS

Phone (248) 788-8048

Email dehthomas@superiormaterials.net

CONTRACT TYPE: No Contract

DESCRIPTION:

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK WAYNE COUNTY - BE THERE 9 AM	

Service Date 8/6/2020

2 hours minimum portal to portal

WORK PERFORMED
 CALL COMPLETE Test scales with Wayne county. Calibrated cement scale. Both scales approved by Wayne county

TECHNICIAN(S)
 JOE BIELARZ, MIKE GLADSTONE

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
10,000 x 5 LB	72113-a	72113-a	ALKON SCALE/INDICATOR / COMMAND	CEMENT SCALE
40,000 x 40 LB	72113-b	72113-b	ALKON SCALE/INDICATOR / COMMAND	AGG SCALE

PARTS LIST

--

Customer Signature



Tom Peters no signature due to virus