



**Customer:** SUPERIOR MATERIALS 20  
 SUP1142 G 5300 NORTH DORT HWY

FLINT , MI 48505

**Certificate ID:** 62822C201027031753

**Test Date:** 10/27/2020

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	3.68
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	45
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass adjustment needed

**Comments:**  
 Test scale,sealed.

TRACEABILITY CERTIFICATE NUMBERS:			
Kit	NIST	Date	Date Due
Description			
GL0601	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0602	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0603	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0604	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0605	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0606	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			

**Technician(s):** JOE AGY;MIKE GLADSTONE;

**Signature:**



Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.



**Customer:** SUPERIOR MATERIALS 20  
 SUP1142 G 5300 NORTH DORT HWY

FLINT , MI 48505

**Certificate ID:** 62822A201027031940

**Test Date:** 10/27/2020

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	15
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	45
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

**Comments:**  
 Test scale,sealed.

TRACEABILITY CERTIFICATE NUMBERS:			
Kit	NIST	Date	Date Due
Description			
GL0601	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0602	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0603	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0604	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0605	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0606	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			

**Technician(s):** JOE AGY;MIKE GLADSTONE;

**Signature:**



Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.