

Customer: SUPERIOR MATERIALS 32
 SUP1173 8911 WEST JEFFERSON

DETROIT, MI

Certificate ID: 62821 CEMENT201027120805

Test Date: 10/27/2020

ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	3.39
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	45
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass adjustment needed

Comments:

Test and calibrate scale,sealed.

TRACEABILITY CERTIFICATE NUMBERS:			
Kit	NIST	Date	Date Due
Description			
GL0601	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0602	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0603	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0604	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0605	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0606	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			

Technician(s): JOE AGY;MIKE GLADSTONE;

Signature:



Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.

Customer: SUPERIOR MATERIALS 32
 SUP1173 8911 WEST JEFFERSON

DETROIT, MI

Certificate ID: 662821 WATER201027122333

Test Date: 10/27/2020

ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	1.28
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	45
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

Comments:

Scale on zero load tested and sealed.

TRACEABILITY CERTIFICATE NUMBERS:

Kit	NIST	Date	Date Due
Description			
GL0601	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0602	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0603	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0604	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0605	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0606	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			

Technician(s): JOE AGY;MIKE GLADSTONE;

Signature:



Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.



GREAT LAKES SCALE COMPANY
 35380 UNION LAKE RD.
 HARRISON TOWNSHIP, MI 48045
 Phone: (586) 776-0517

Certificate of Calibration

Customer: SUPERIOR MATERIALS 32
 SUP1173 8911 WEST JEFFERSON

Certificate ID: 662821 AGG A201027121851

Test Date: 10/27/2020

DETROIT, MI

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: 662821 AGG A Capacity: 40000 LB Graduation: 20 LB

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
10/27/2020	04/28/2021	Semi-Annually	662821 AGG A	III L
SCALE LOCATION		EQUIPMENT CONDITIONS		
AGG A		Working	Clean	
SELECT NUMBER OF PM DAYS:		CAPACITY	40000 UOM	LB INCREMENT 20

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS			AS LEFT TEST USING TEST WEIGHTS		
Weight Applied	Reading	Error	Weight Applied	Reading	Error
0	LB 0	LB 0	0	LB 0	LB 0
6000	LB 6020	LB 20	6000	LB 6000	LB 0
0	LB 0	LB 0	0	LB 0	LB 0
	LB	LB	1000	LB 1000	LB 0
	LB	LB	2000	LB 2000	LB 0
	LB	LB	3000	LB 3000	LB 0
	LB	LB	4000	LB 4000	LB 0
	LB	LB	5000	LB 5000	LB 0
	LB	LB	6000	LB 6000	LB 0
	LB	LB	0	LB 0	LB 0

AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
2000	LB 6000	LB 8000	LB 0
4000	LB 6000	LB 10000	LB 0
6000	LB 6000	LB 12000	LB 0
2000	LB 12000	LB 14000	LB 0
4000	LB 12000	LB 16000	LB 0
6000	LB 12000	LB 18000	LB 0
2000	LB 18000	LB 20000	LB 0
4000	LB 18000	LB 22000	LB 0
6000	LB 18000	LB 24000	LB 0
2000	LB 24000	LB 26000	LB 0
4000	LB 24000	LB 28000	LB 0
6000	LB 24000	LB 30000	LB 0
2000	LB 32000	LB 34000	LB 0
4000	LB 32000	LB 36000	LB 0
6000	LB 32000	LB 38000	LB 0
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB

OVERALL TEST RESULTS PASS

Customer: SUPERIOR MATERIALS 32
 SUP1173 8911 WEST JEFFERSON

DETROIT, MI

Certificate ID: 662821 AGG A201027121851

Test Date: 10/27/2020

ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	15
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	45
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment


Comments:

Test and calibrate scale,sealed.

TRACEABILITY CERTIFICATE NUMBERS:			
Kit	NIST	Date	Date Due
Description			
GL0601	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0602	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0603	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0604	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0605	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0606	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			

Technician(s): JOE AGY;MIKE GLADSTONE;

Signature:



Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.