

Customer: SUPERIOR MATERIALS 39
 SUP1148 6750 SIMS DR

Certificate ID: 23635-cc200810105802

Test Date: 08/10/2020

ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	1.59
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	82
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

Comments:

Test scale,sealed.

TRACEABILITY CERTIFICATE NUMBERS:			
Kit	NIST	Date	Date Due
Description			
GL1502	MI-03-19-14172	3/19/2019	
3/19/2021	TRUCK 15 - 1000 LB		
GL1503	MI-03-19-14172	3/19/2019	
3/19/2021	TRUCK 15 - 1000 LB		
GL1504	MI-03-19-14172	3/19/2019	
3/19/2021	TRUCK 15 - 1000 LB		
GL1505	MI-03-19-14172	3/19/2019	
3/19/2021	TRUCK 15 - 1000 LB		

Technician(s): JOE BIELARZ;MIKE GLADSTONE;

Signature:



Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.

WORK ORDER

Great Lakes Scale

35380 UNION LAKE RD
 HARRISON TOWNSHIP, MI 48045
 Phone: (586)776-0517
 Fax: (586)776-3169
 Fed Tax ID: 38-3206225

Work Order# 0028856-1

Request Date 8/7/2020 1:45:00 PM

Schedule Date 8/10/2020 8:00:00 AM

Bill To SUP0896

SUP0896
 SUPERIOR MATERIALS HOLDINGS LLC
 PO BOX 2900
 FARMINGTON HILLS, MI, 48336
 (248) 788-8000

Ship To SUP1148

SUP1148
 SUPERIOR MATERIALS 39
 6750 SIMS DR
 STERLING HEIGHTS, MI, 48313
 (248) 788-8000

Customer P.O

Requester DAN THOMAS

Phone (248) 788-8048

Email dehthomas@superiormaterials.net

CONTRACT TYPE: No Contract

DESCRIPTION:

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK	

Service Date 8/10/2020

2 hours minimum portal to portal

WORK PERFORMED	TECHNICIAN(S)
CALL COMPLETE Test and calibrate cement scale, test aggregate and water scales.	JOE BIELARZ, MIKE GLADSTONE

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
10,000 x 5 lb	23635-aa	23635-aa	ALKON SCALE/INDICATOR / COMMAND	CEMENT
40,000 x 20 lb	23635-bb	23635-bb	ALKON SCALE/INDICATOR / COMMAND	AGG
4000 lbs x 2 lbs	23635-cc	23635-cc	ALKON SCALE/INDICATOR / COMMAND	WATER

PARTS LIST

Customer Signature

 Dan Thomas no signature due to virus