

WORK ORDER

Great Lakes Scale

35380 UNION LAKE RD
 HARRISON TOWNSHIP, MI 48045
 Phone: (586)776-0517
 Fax: (586)776-3169
 Fed Tax ID: 38-3206225

Work Order# 0034156-1
Request Date 2/24/2023 3:49:00 PM
Schedule Date 2/27/2023 2:00:00 PM

Bill To SUP0896
 SUP0896
 SUPERIOR MATERIALS HOLDINGS LLC
 PO BOX 2900
 FARMINGTON HILLS, MI, 48336
 (248) 788-8000

Ship To SUP0992
 SUP0992
 SUPERIOR MATERIALS 4912
 10147 GRAND RIVER
 BRIGHTON, MI, 48116
 (248) 788-8000

Customer P.O
Requester DAN THOMAS
Phone
Email dan.hargrave-
 thomas@vcimentos.com

CONTRACT TYPE: No Contract

DESCRIPTION:

| TASK DESCRIPTION | SERVICE TYPE: CAL |
|--|-------------------|
| SCALE CHECK 2pm | |
| WORK PERFORMED | |
| CALL COMPLETE | |
| Did cement scale inspection, calibrated the cement scale that was 10 lbs off. Stickers signed, calibration reports produced. | |

Service Date 2/27/2023

2 hours minimum portal to portal

| TECHNICIAN(S) |
|---------------------------|
| LUCAS SULLIVAN, ROB DUBAY |

| EQUIPMENT SERVICED | ASSET/UNIT # | SERIAL # | MODEL | LOCATION |
|--------------------|--------------|----------|------------------------------------|----------|
| 4,000 x 5 LB | 23635-c | 23625-c | ALKON SCALE/INDICATOR / COMMAND | |
| 10,000 x 5 LB | 23635-a | 23635-a | ALKON SCALE/INDICATOR / COMMAND | |
| 45,000 x 20 LB | 23635-b | 23635-b | ALKON SCALE/INDICATOR / COMMAND | |

| PARTS LIST |
|------------|
| |

Customer Signature

 Name:



WORK ORDER

35380 UNION LAKE RD
HARRISON TOWNSHIP, MI 48045
Phone: (586)776-0517
Fax: (586)776-3169
Fed Tax ID: 38-3206225

Work Order# 0034156-1
Request Date 2/24/2023 3:49:00 PM
Schedule Date 2/27/2023 2:00:00 PM

Bill To SUP0896
SUP0896
SUPERIOR MATERIALS HOLDINGS LLC
PO BOX 2900
FARMINGTON HILLS, MI, 48336
(248) 788-8000

Ship To SUP0992
SUP0992
SUPERIOR MATERIALS 4912
10147 GRAND RIVER
BRIGHTON, MI, 48116
(248) 788-8000

Customer P.O
Requester DAN THOMAS
Phone
Email dan.hargrave-
thomas@vcimentos.com

CONTRACT TYPE: No Contract

DESCRIPTION:

| TASK DESCRIPTION | SERVICE TYPE: CAL |
|------------------|-------------------|
| SCALE CHECK 2pm | |

Service Date 2/27/2023

2 hours minimum portal to portal

| WORK PERFORMED | TECHNICIAN(S) |
|--|---------------------------|
| CALL COMPLETE | LUCAS SULLIVAN, ROB DUBAY |
| Did cement scale inspection, calibrated the cement scale that was 10 lbs off. Stickers signed, calibration reports produced. | |

| EQUIPMENT SERVICED | ASSET/UNIT # | SERIAL # | MODEL | LOCATION |
|--------------------|--------------|----------|------------------------------------|----------|
| 4,000 x 5 LB | 23635-c | 23625-c | ALKON SCALE/INDICATOR / COMMAND | |
| 10,000 x 5 LB | 23635-a | 23635-a | ALKON SCALE/INDICATOR / COMMAND | |
| 45,000 x 20 LB | 23635-b | 23635-b | ALKON SCALE/INDICATOR / COMMAND | |

| PARTS LIST |
|------------|
| |

Customer Signature

Kyle



GREAT LAKES SCALE COMPANY
 35380 UNION LAKE RD.
 HARRISON TOWNSHIP, MI 48045
 Phone: (586) 776-0517

Certificate of Calibration

Certificate ID: 23635-c230308081222

Customer: SUPERIOR MATERIALS 4912
 SUP0992 10147 GRAND RIVER

Test Date: 02/27/2023

BRIGHTON, MI 48116

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: 23625-c Capacity: 4000 lb Graduation: 5 lb

| CALIBRATION DATE | DUE DATE | TEST INTERVAL | SCALE ID | SCALE CLASS |
|---------------------------|------------|----------------------|----------|-------------|
| 02/27/2023 | 08/29/2023 | Semi-Annually | 23635-c | III |
| SCALE LOCATION | | EQUIPMENT CONDITIONS | | |
| WATER SCALE | | Working | Clean | |
| SELECT NUMBER OF PM DAYS: | CAPACITY | 4000 UOM | lb | INCREMENT |
| | | | | 5 |

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

| Weight Applied | Reading | Error | Weight Applied | Reading | Error |
|----------------|---------|-------|----------------|---------|-------|
| 0 | 0 | 0 | 0 | 0 | 0 |
| 1000 | 1000 | 0 | 1000 | 1000 | 0 |
| 2000 | 2000 | 0 | 2000 | 2000 | 0 |
| 3000 | 3000 | 0 | 3000 | 3000 | 0 |
| 4000 | 4000 | 0 | 4000 | 4000 | 0 |
| 0 | 0 | 0 | 0 | 0 | 0 |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |

AS LEFT BUILD TEST

| Test Weight Applied | Material Weight Applied | Reading | Error |
|---------------------|-------------------------|---------|-------|
| 0 | 0 | 0 | 0 |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |

OVERALL TEST RESULTS PASS

Customer: SUPERIOR MATERIALS 4912
 SUP0992 10147 GRAND RIVER

Certificate ID: 23635-c230308081222

Test Date: 02/27/2023

BRIGHTON, MI 48116

ISO Number: MS-FM-CEM-001

TEST INFORMATION

| | |
|---|----------------------|
| TEST WEIGHT CLASSIFICATION | EXPANDED UNCERTAINTY |
| F | 8.7 |
| ENVIRONMENTAL CONDITIONS | TEMPERATURE |
| Acceptable | 33 |
| WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY? | |
| Yes | |
| TEST LOCATION | OVERALL RESULT |
| Onsite | Pass w/o adjustment |

Comments:

Scale on zero, load tested and sealed.

| TRACEABILITY CERTIFICATE NUMBERS: | | | | |
|-----------------------------------|----------------|----------|----------|--|
| Kit | NIST | Date | Date Due | |
| Description | | | | |
| GL1503 | MI-03-21-15156 | 3/5/2021 | 3/5/2023 | |
| TRUCK 15 - 1000 LB | | | | |
| GL1504 | MI-03-21-15156 | 3/5/2021 | 3/5/2023 | |
| TRUCK 15 - 1000 LB | | | | |
| GL1505 | MI-03-21-15156 | 3/5/2021 | 3/5/2023 | |
| TRUCK 15 - 1000 LB | | | | |
| GL1506 | MI-03-21-15156 | 3/5/2021 | 3/5/2023 | |
| TRUCK 15 - 1000 LB | | | | |

Technician(s): LUCAS SULLIVAN;ROB DUBAY;

Signature: **Lucas Sullivan** Digitally signed by Lucas Sullivan

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.



GREAT LAKES SCALE COMPANY
 35380 UNION LAKE RD.
 HARRISON TOWNSHIP, MI 48045
 Phone: (586) 776-0517

Certificate of Calibration

Customer: SUPERIOR MATERIALS 4912
 SUP0992 10147 GRAND RIVER

Certificate ID: 23635-a230308081344

Test Date: 02/27/2023

BRIGHTON, MI 48116

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: 23635-a Capacity: 10000 lb Graduation: 5 lb

| CALIBRATION DATE | DUE DATE | TEST INTERVAL | SCALE ID | SCALE CLASS |
|---------------------------|------------|----------------------|-----------|-------------|
| 02/27/2023 | 08/29/2023 | Semi-Annually | 23635-a | III |
| SCALE LOCATION | | EQUIPMENT CONDITIONS | | |
| CEMENT SCALE | | Working | Clean | |
| SELECT NUMBER OF PM DAYS: | | CAPACITY | 10000 UOM | INCREMENT |
| | | | lb | 5 |

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

| AS FOUND TEST USING TEST WEIGHTS | | | | | | AS LEFT TEST USING TEST WEIGHTS | | | | | |
|----------------------------------|---------|------|-------|-----|----|---------------------------------|---------|------|-------|---|----|
| Weight Applied | Reading | | Error | | | Weight Applied | Reading | | Error | | |
| 0 | lb | 0 | lb | 0 | lb | 0 | lb | 0 | lb | 0 | lb |
| 2000 | lb | 1995 | lb | -5 | lb | 2000 | lb | 2000 | lb | 0 | lb |
| 4000 | lb | 3995 | lb | -5 | lb | 4000 | lb | 4000 | lb | 0 | lb |
| 6000 | lb | 5990 | lb | -10 | lb | 6000 | lb | 6000 | lb | 0 | lb |
| 0 | lb | 0 | lb | 0 | lb | 0 | lb | 0 | lb | 0 | lb |
| | lb | | lb | | lb | | lb | | lb | | lb |
| | lb | | lb | | lb | | lb | | lb | | lb |
| | lb | | lb | | lb | | lb | | lb | | lb |
| | lb | | lb | | lb | | lb | | lb | | lb |
| | lb | | lb | | lb | | lb | | lb | | lb |

AS LEFT BUILD TEST

| Test Weight Applied | Material Weight Applied | | Reading | | Error | |
|---------------------|-------------------------|------|---------|------|-------|---|
| 6000 | lb | 3125 | lb | 9125 | lb | 0 |
| | lb | | lb | | lb | |
| | lb | | lb | | lb | |
| | lb | | lb | | lb | |
| | lb | | lb | | lb | |
| | lb | | lb | | lb | |
| | lb | | lb | | lb | |
| | lb | | lb | | lb | |
| | lb | | lb | | lb | |
| | lb | | lb | | lb | |
| | lb | | lb | | lb | |
| | lb | | lb | | lb | |
| | lb | | lb | | lb | |
| | lb | | lb | | lb | |
| | lb | | lb | | lb | |
| | lb | | lb | | lb | |
| | lb | | lb | | lb | |
| | lb | | lb | | lb | |

OVERALL TEST RESULTS PASS

Customer: SUPERIOR MATERIALS 4912
 SUP0992 10147 GRAND RIVER

Certificate ID: 23635-a230308081344

Test Date: 02/27/2023

BRIGHTON, MI 48116

ISO Number: MS-FM-CEM-001

TEST INFORMATION

| | |
|---|----------------------|
| TEST WEIGHT CLASSIFICATION | EXPANDED UNCERTAINTY |
| F | 8.7 |
| ENVIRONMENTAL CONDITIONS | TEMPERATURE |
| Acceptable | 33 |
| WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY? | |
| Yes | |
| TEST LOCATION | OVERALL RESULT |
| Onsite | Pass w/o adjustment |

Comments:

Scale on zero, load tested, build tested and sealed.

TRACEABILITY CERTIFICATE NUMBERS:

| Kit | NIST | Date | Date Due |
|--------------------|----------------|----------|----------|
| Description | | | |
| GL1501 | MI-03-21-15156 | 3/5/2021 | 3/5/2023 |
| TRUCK 15 - 1000 LB | | | |
| GL1502 | MI-03-21-15156 | 3/5/2021 | 3/5/2023 |
| TRUCK 15 - 1000 LB | | | |
| GL1503 | MI-03-21-15156 | 3/5/2021 | 3/5/2023 |
| TRUCK 15 - 1000 LB | | | |
| GL1504 | MI-03-21-15156 | 3/5/2021 | 3/5/2023 |
| TRUCK 15 - 1000 LB | | | |
| GL1505 | MI-03-21-15156 | 3/5/2021 | 3/5/2023 |
| TRUCK 15 - 1000 LB | | | |
| GL1506 | MI-03-21-15156 | 3/5/2021 | 3/5/2023 |
| TRUCK 15 - 1000 LB | | | |

Technician(s): LUCAS SULLIVAN;ROB DUBAY;

Signature: **Lucas Sullivan** Digitally signed by Lucas Sullivan

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.



GREAT LAKES SCALE COMPANY
 35380 UNION LAKE RD.
 HARRISON TOWNSHIP, MI 48045
 Phone: (586) 776-0517

Certificate of Calibration

Customer: SUPERIOR MATERIALS 4912
 SUP0992 10147 GRAND RIVER

Certificate ID: 23635-b230308081435

Test Date: 02/27/2023

BRIGHTON, MI 48116

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: 23635-b Capacity: 45000 lb Graduation: 20 lb

| CALIBRATION DATE | DUE DATE | TEST INTERVAL | SCALE ID | SCALE CLASS |
|---------------------------|------------|----------------------|--------------|--------------|
| 02/27/2023 | 08/29/2023 | Semi-Annually | 23635-b | III L |
| SCALE LOCATION | | EQUIPMENT CONDITIONS | | |
| AGG SCALE | | Working | Clean | |
| SELECT NUMBER OF PM DAYS: | | CAPACITY | 45000 UOM lb | INCREMENT 20 |

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

| AS FOUND TEST USING TEST WEIGHTS | | | | | | AS LEFT TEST USING TEST WEIGHTS | | | | | |
|----------------------------------|---------|------|-------|---|----|---------------------------------|---------|------|-------|---|----|
| Weight Applied | Reading | | Error | | | Weight Applied | Reading | | Error | | |
| 0 | lb | 0 | lb | 0 | lb | 0 | lb | 0 | lb | 0 | lb |
| 6000 | lb | 6000 | lb | 0 | lb | 2000 | lb | 2000 | lb | 0 | lb |
| 0 | lb | 0 | lb | 0 | lb | 4000 | lb | 4000 | lb | 0 | lb |
| | lb | | lb | | lb | 6000 | lb | 6000 | lb | 0 | lb |
| | lb | | lb | | lb | 0 | lb | 0 | lb | 0 | lb |
| | lb | | lb | | lb | | lb | | lb | | lb |
| | lb | | lb | | lb | | lb | | lb | | lb |
| | lb | | lb | | lb | | lb | | lb | | lb |
| | lb | | lb | | lb | | lb | | lb | | lb |
| | lb | | lb | | lb | | lb | | lb | | lb |
| | lb | | lb | | lb | | lb | | lb | | lb |

AS LEFT BUILD TEST

| Test Weight Applied | | Material Weight Applied | | Reading | | Error | |
|---------------------|----|-------------------------|----|---------|----|-------|----|
| 6000 | lb | 8000 | lb | 14000 | lb | 0 | lb |
| 6000 | lb | 12240 | lb | 18240 | lb | 0 | lb |
| 6000 | lb | 18080 | lb | 24080 | lb | 0 | lb |
| 6000 | lb | 24000 | lb | 30000 | lb | 0 | lb |
| 6000 | lb | 30100 | lb | 36100 | lb | 0 | lb |
| | lb | | lb | | lb | | lb |
| | lb | | lb | | lb | | lb |
| | lb | | lb | | lb | | lb |
| | lb | | lb | | lb | | lb |
| | lb | | lb | | lb | | lb |
| | lb | | lb | | lb | | lb |
| | lb | | lb | | lb | | lb |
| | lb | | lb | | lb | | lb |
| | lb | | lb | | lb | | lb |
| | lb | | lb | | lb | | lb |
| | lb | | lb | | lb | | lb |
| | lb | | lb | | lb | | lb |
| | lb | | lb | | lb | | lb |

OVERALL TEST RESULTS PASS

Customer: SUPERIOR MATERIALS 4912
 SUP0992 10147 GRAND RIVER

BRIGHTON, MI 48116

Certificate ID: 23635-b230308081435

Test Date: 02/27/2023

ISO Number: MS-FM-CEM-001

TEST INFORMATION

| | |
|---|----------------------|
| TEST WEIGHT CLASSIFICATION | EXPANDED UNCERTAINTY |
| F | 35 |
| ENVIRONMENTAL CONDITIONS | TEMPERATURE |
| Acceptable | 33 |
| WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY? | |
| Yes | |
| TEST LOCATION | OVERALL RESULT |
| Onsite | Pass w/o adjustment |

Comments:

Scale on zero, load tested, build tested and sealed.

TRACEABILITY CERTIFICATE NUMBERS:

| Kit | NIST | Date | Date Due |
|--------------------|----------------|----------|----------|
| Description | | | |
| GL1501 | MI-03-21-15156 | 3/5/2021 | 3/5/2023 |
| TRUCK 15 - 1000 LB | | | |
| GL1502 | MI-03-21-15156 | 3/5/2021 | 3/5/2023 |
| TRUCK 15 - 1000 LB | | | |
| GL1503 | MI-03-21-15156 | 3/5/2021 | 3/5/2023 |
| TRUCK 15 - 1000 LB | | | |
| GL1504 | MI-03-21-15156 | 3/5/2021 | 3/5/2023 |
| TRUCK 15 - 1000 LB | | | |
| GL1505 | MI-03-21-15156 | 3/5/2021 | 3/5/2023 |
| TRUCK 15 - 1000 LB | | | |
| GL1506 | MI-03-21-15156 | 3/5/2021 | 3/5/2023 |
| TRUCK 15 - 1000 LB | | | |

Technician(s): LUCAS SULLIVAN;ROB DUBAY;

Signature: Lucas Sullivan  Digitally signed by Lucas Sullivan

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.