

WORK ORDER

Great Lakes Scale

35380 UNION LAKE RD
 HARRISON TOWNSHIP, MI 48045
 Phone: (586)776-0517
 Fax: (586)776-3169
 Fed Tax ID: 38-3206225

Work Order# 0034466-1
Request Date 4/18/2023 3:45:00 PM
Schedule Date 4/19/2023 2:14:00 PM

Bill To SUP0896

 SUP0896
 SUPERIOR MATERIALS HOLDINGS LLC
 PO BOX 2900
 FARMINGTON HILLS, MI, 48336
 (248) 788-8000

Ship To SUP1173

 SUP1173
 SUPERIOR MATERIALS 4901 (32)
 8911 WEST JEFFERSON
 DETROIT, MI,
 (248) 788-8000

Customer P.O
Requester DAN THOMAS
Phone (248) 788-8000
Email dan.hargrave-
 thomas@vcimentos.com

CONTRACT TYPE: No Contract

DESCRIPTION:

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK WITH WAYNE COUNTY GLS ONSITE AT 1:30P	

Service Date 4/19/2023

2 hours minimum portal to portal

WORK PERFORMED
 Inspected tested calibrated and sealed 3 hopper scales. Certificates to follow.
 CALL COMPLETE

TECHNICIAN(S)
 ED GORHAM, LUCAS SULLIVAN

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
40,000 x 20 lb	662821 AGG A	662821 AGG A	ALKON SCALE/INDICATOR / COMMAND	AGG A
10,000 x 5 lb	662821 CEMENT	662821 CEMENT	ALKON SCALE/INDICATOR / COMMAND	
6000 x 1	662821 WATER	662821 WATER	ALKON SCALE/INDICATOR / COMMAND	

PARTS LIST

--

Customer Signature

 Name:

WORK ORDER

35380 UNION LAKE RD
 HARRISON TOWNSHIP, MI 48045
 Phone: (586)776-0517
 Fax: (586)776-3169
 Fed Tax ID: 38-3206225

Work Order# 0034466-1
Request Date 4/18/2023 3:45:00 PM
Schedule Date 4/19/2023 2:14:00 PM

Bill To SUP0896
 SUP0896
 SUPERIOR MATERIALS HOLDINGS LLC
 PO BOX 2900
 FARMINGTON HILLS, MI, 48336
 (248) 788-8000

Ship To SUP1173
 SUP1173
 SUPERIOR MATERIALS 4901 (32)
 8911 WEST JEFFERSON
 DETROIT, MI,
 (248) 788-8000

Customer P.O
Requester DAN THOMAS
Phone (248) 788-8000
Email dan.hargrave-thomas@vcimentos.com

CONTRACT TYPE: No Contract

DESCRIPTION:

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK WITH WAYNE COUNTY GLS ONSITE AT 1:30P	

Service Date 4/19/2023

2 hours minimum portal to portal

WORK PERFORMED
 Inspected tested calibrated and sealed 3 hopper scales. Certificates to follow.
 CALL COMPLETE

TECHNICIAN(S)
 ED GORHAM, LUCAS SULLIVAN

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
40,000 x 20 lb	662821 AGG A	662821 AGG A	ALKON SCALE/INDICATOR / COMMAND	AGG A
10,000 x 5 lb	662821 CEMENT	662821 CEMENT	ALKON SCALE/INDICATOR / COMMAND	
6000 x 1	662821 WATER	662821 WATER	ALKON SCALE/INDICATOR / COMMAND	

PARTS LIST

--

Customer Signature



Walton

Customer: SUPERIOR MATERIALS 4901 (32)
 SUP1173 8911 WEST JEFFERSON

DETROIT, MI

Certificate ID: 662821 WATER230420072546

Test Date: 04/19/2023

ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	1.8
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	45
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

Comments:

Scale on zero, load tested, calibrated, retested and sealed.

TRACEABILITY CERTIFICATE NUMBERS:				
Kit	NIST	Date	Date Due	
Description				
GL1501	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15 - 1000 LB				
GL1502	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15 - 1000 LB				
GL1503	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15 - 1000 LB				
GL1504	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15 - 1000 LB				

Technician(s): ED GORHAM;LUCAS SULLIVAN;



Signature:

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.

Customer: SUPERIOR MATERIALS 4901 (32)
 SUP1173 8911 WEST JEFFERSON

DETROIT, MI

Certificate ID: 662821 CEMENT230420072925

Test Date: 04/19/2023

ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	8.7
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	45
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass adjustment needed


Comments:

Scale on zero, load tested, Calibrated, retested, build tested and sealed.

TRACEABILITY CERTIFICATE NUMBERS:

Kit	NIST	Date	Date Due
Description			
GL1501	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			
GL1502	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			
GL1503	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			
GL1504	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			
GL1505	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			
GL1506	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			

Technician(s): ED GORHAM;LUCAS SULLIVAN;



Signature:

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.



GREAT LAKES SCALE COMPANY
 35380 UNION LAKE RD.
 HARRISON TOWNSHIP, MI 48045
 Phone: (586) 776-0517

Certificate of Calibration

Customer: SUPERIOR MATERIALS 4901 (32)
 SUP1173 8911 WEST JEFFERSON

Certificate ID: 662821 AGG A230420073049

Test Date: 04/20/2023

DETROIT, MI

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: 662821 AGG A Capacity: 40000 lb Graduation: 20 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
04/20/2023	10/20/2023	Semi-Annually	662821 AGG A	III L
SCALE LOCATION		EQUIPMENT CONDITIONS		
AGG A		Working	Clean	
SELECT NUMBER OF PM DAYS:		CAPACITY	40000 UOM lb	INCREMENT 20

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS						AS LEFT TEST USING TEST WEIGHTS					
Weight Applied	Reading		Error			Weight Applied	Reading		Error		
0	lb	0	lb	0	lb	0	lb	0	lb	0	lb
1000	lb	1000	lb	0	lb	1000	lb	1000	lb	0	lb
2000	lb	2000	lb	0	lb	2000	lb	2000	lb	0	lb
3000	lb	3000	lb	0	lb	3000	lb	3000	lb	0	lb
4000	lb	4000	lb	0	lb	4000	lb	4000	lb	0	lb
5000	lb	5000	lb	0	lb	5000	lb	5000	lb	0	lb
6000	lb	6000	lb	0	lb	6000	lb	6000	lb	0	lb
0	lb	0	lb	0	lb	0	lb	0	lb	0	lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb

AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
2000	lb 6000	lb 8000	0 lb
4000	lb 6000	lb 10000	0 lb
6000	lb 6000	lb 12000	0 lb
2000	lb 12000	lb 14000	0 lb
4000	lb 12000	lb 16000	0 lb
6000	lb 12000	lb 18000	0 lb
2000	lb 18000	lb 20000	0 lb
4000	lb 18000	lb 22000	0 lb
6000	lb 18000	lb 24000	0 lb
2000	lb 24000	lb 26000	0 lb
4000	lb 24000	lb 28000	0 lb
6000	lb 24000	lb 30000	0 lb
2000	lb 30000	lb 32000	0 lb
4000	lb 30000	lb 34000	0 lb
6000	lb 30000	lb 36000	0 lb
	lb	lb	lb
	lb	lb	lb
	lb	lb	lb
	lb	lb	lb

OVERALL TEST RESULTS PASS

Customer: SUPERIOR MATERIALS 4901 (32)
 SUP1173 8911 WEST JEFFERSON

Certificate ID: 662821 AGG A230420073049

DETROIT, MI

Test Date: 04/20/2023

ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	35
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	55
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment


Comments:

Scale on zero load tested, build tested and sealed.

TRACEABILITY CERTIFICATE NUMBERS:

Kit	NIST	Date	Date Due
Description			
GL1501	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			
GL1502	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			
GL1503	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			
GL1504	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			
GL1505	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			
GL1506	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			

Technician(s): ED GORHAM;LUCAS SULLIVAN;



Signature:

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.