



WORK ORDER

Great Lakes Scale

35380 UNION LAKE RD
 HARRISON TOWNSHIP, MI 48045
 Phone: (586)776-0517
 Fax: (586)776-3169
 Fed Tax ID: 38-3206225

Work Order# 0034333-1
Request Date 3/27/2023 9:40:00 AM
Schedule Date 3/28/2023 1:00:00 PM

Bill To SUP0896
 SUP0896
 SUPERIOR MATERIALS HOLDINGS LLC
 PO BOX 2900
 FARMINGTON HILLS, MI, 48336
 (248) 788-8000

Ship To SUP1104
 SUP1104
 SUPERIOR MATERIALS 4906 (35)
 39001 W HURON RIVER DR
 ROMULUS, MI, 48174
 (248) 788-8000

Customer P.O 6300425997
Requester DAN THOMAS
Phone (248) 788-8048
Email dehthomas@superiormaterials.net

CONTRACT TYPE: No Contract

DESCRIPTION:

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK WITH WAYNE COUNTY	

Service Date _____

2 hours minimum portal to portal

WORK PERFORMED
 CALL COMPLETE
 CALIBRATIONS COMPLETE tested agg and cement. Calibrated cement. Certs to follow.

TECHNICIAN(S)
 LUCAS SULLIVAN, ROB DUBAY

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
10,000 x 10 LB	62821-CEMENT	62821-CEMENT	ALKON SCALE/INDICATOR / COMMAND	CEMENT
40,000 x 20 LB	62821-AGG	62821-AGG	ALKON SCALE/INDICATOR / COMMAND	AGG

PARTS LIST

--

Customer Signature

Name:



WORK ORDER

35380 UNION LAKE RD
HARRISON TOWNSHIP, MI 48045
Phone: (586)776-0517
Fax: (586)776-3169
Fed Tax ID: 38-3206225

Work Order# 0034333-1
Request Date 3/27/2023 9:40:00 AM
Schedule Date 3/28/2023 1:00:00 PM

Bill To SUP0896
SUP0896
SUPERIOR MATERIALS HOLDINGS LLC
PO BOX 2900
FARMINGTON HILLS, MI, 48336
(248) 788-8000

Ship To SUP1104
SUP1104
SUPERIOR MATERIALS 4906 (35)
39001 W HURON RIVER DR
ROMULUS, MI, 48174
(248) 788-8000

Customer P.O 6300425997
Requester DAN THOMAS
Phone (248) 788-8048
Email dehthomas@superiormaterials.net

CONTRACT TYPE: No Contract

DESCRIPTION:

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK WITH WAYNE COUNTY	

Service Date _____

2 hours minimum portal to portal

WORK PERFORMED
CALL COMPLETE
CALIBRATIONS COMPLETE tested agg and cement. Calibrated cement. Certs to follow.

TECHNICIAN(S)
LUCAS SULLIVAN, ROB DUBAY

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
10,000 x 10 LB	62821-a	62821-a	ALKON SCALE/INDICATOR / COMMAND	
40,000 x 20 LB	62821-b	62821-b	ALKON SCALE/INDICATOR / COMMAND	

PARTS LIST

Customer Signature

Jeremy



GREAT LAKES SCALE COMPANY
 35380 UNION LAKE RD.
 HARRISON TOWNSHIP, MI 48045
 Phone: (586) 776-0517

Certificate of Calibration

Certificate ID: 62821-a230404024157

Customer: SUPERIOR MATERIALS 4906 (35)
 SUP1104 39001 W HURON RIVER DR

Test Date: 03/28/2023

ROMULUS, MI 48174

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: 62821-a Capacity: 10000 lb Graduation: 10 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
03/28/2023	09/27/2023	Semi-Annually	62821-a	III
SCALE LOCATION		EQUIPMENT CONDITIONS		
CEMENT		Working	Clean	
SELECT NUMBER OF PM DAYS:	CAPACITY	10000 UOM	lb	INCREMENT 10

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

Weight Applied	Reading	Error	Weight Applied	Reading	Error
0	0	0	0	0	0
1000	1000	0	1000	1000	0
2000	2000	0	2000	2000	0
3000	3000	0	3000	3000	0
4000	4010	10	4000	4000	0
5000	5010	10	5000	5000	0
6000	6020	20	6000	6000	0

AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
6000	0	6000	0

OVERALL TEST RESULTS PASS

Customer: SUPERIOR MATERIALS 4906 (35)
 SUP1104 39001 W HURON RIVER DR

Certificate ID: 62821-a230404024157

Test Date: 03/28/2023

ROMULUS, MI 48174

ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	17
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	45
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

Comments:

TRACEABILITY CERTIFICATE NUMBERS:			
Kit	NIST	Date	Date Due
Description			
GL1501	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			
GL1502	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			
GL1503	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			
GL1504	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			
GL1505	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			
GL1506	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			

Technician(s): LUCAS SULLIVAN;ROB DUBAY;

Signature: Lucas Sullivan  Digitally signed by Lucas Sullivan

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.



GREAT LAKES SCALE COMPANY
 35380 UNION LAKE RD.
 HARRISON TOWNSHIP, MI 48045
 Phone: (586) 776-0517

Certificate of Calibration

Customer: SUPERIOR MATERIALS 4906 (35)
 SUP1104 39001 W HURON RIVER DR

Certificate ID: 62821-b230404024050

Test Date: 03/28/2023

ROMULUS, MI 48174

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: 62821-b Capacity: 40000 lb Graduation: 20 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
03/28/2023	09/27/2023	Semi-Annually	62821-b	III L
SCALE LOCATION		EQUIPMENT CONDITIONS		
AGG		Working	Clean	
SELECT NUMBER OF PM DAYS:	CAPACITY	40000 UOM	lb	INCREMENT 20

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS						AS LEFT TEST USING TEST WEIGHTS					
Weight Applied	Reading		Error			Weight Applied	Reading		Error		
0	lb	0	lb	0	lb	0	lb	0	lb	0	lb
1000	lb	1000	lb	0	lb	1000	lb	1000	lb	0	lb
2000	lb	2000	lb	0	lb	2000	lb	2000	lb	0	lb
3000	lb	3000	lb	0	lb	3000	lb	3000	lb	0	lb
4000	lb	4000	lb	0	lb	4000	lb	4000	lb	0	lb
5000	lb	5000	lb	0	lb	5000	lb	5000	lb	0	lb
6000	lb	6000	lb	0	lb	6000	lb	6000	lb	0	lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb

AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
2000	lb 6000	lb 8000	0 lb
4000	lb 6000	lb 10000	0 lb
6000	lb 6000	lb 12000	0 lb
2000	lb 12000	lb 14000	0 lb
4000	lb 12000	lb 16000	0 lb
6000	lb 12000	lb 18000	0 lb
2000	lb 18000	lb 20000	0 lb
4000	lb 18000	lb 22000	0 lb
6000	lb 18000	lb 24000	0 lb
2000	lb 24000	lb 26000	0 lb
4000	lb 24000	lb 28000	0 lb
6000	lb 24000	lb 30000	0 lb
2000	lb 30000	lb 32000	0 lb
4000	lb 30000	lb 34000	0 lb
6000	lb 30000	lb 36000	0 lb
	lb	lb	lb
	lb	lb	lb
	lb	lb	lb
	lb	lb	lb

OVERALL TEST RESULTS PASS

Customer: SUPERIOR MATERIALS 4906 (35)
 SUP1104 39001 W HURON RIVER DR

Certificate ID: 62821-b230404024050

Test Date: 03/28/2023

ROMULUS, MI 48174

ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	35
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	40
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

Comments:

TRACEABILITY CERTIFICATE NUMBERS:				
Kit	NIST	Date	Date Due	
Description				
GL1501	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15 - 1000 LB				
GL1502	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15 - 1000 LB				
GL1503	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15 - 1000 LB				
GL1504	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15 - 1000 LB				
GL1505	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15 - 1000 LB				
GL1506	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15 - 1000 LB				

Technician(s): LUCAS SULLIVAN;ROB DUBAY;

Signature: Lucas Sullivan  Digitally signed by Lucas Sullivan

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.