

# WORK ORDER

**Great Lakes Scale**

35380 UNION LAKE RD  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586)776-0517  
 Fax: (586)776-3169  
 Fed Tax ID: 38-3206225

**Work Order#** 0029762-1

**Request Date** 1/27/2021 12:34:00 PM

**Schedule Date** 1/28/2021 1:00:00 PM

**Bill To** SUP0896

SUP0896  
 SUPERIOR MATERIALS HOLDINGS LLC  
 PO BOX 2900  
 FARMINGTON HILLS, MI, 48336  
 (248) 788-8000

**Ship To** SUP1148

SUP1148  
 SUPERIOR MATERIALS 39  
 6750 SIMS DR  
 STERLING HEIGHTS, MI, 48313  
 (248) 788-8000

**Customer P.O**

**Requester** DAN THOMAS

**Phone** (248) 788-8048

**Email** dehthomas@superiormaterials.net

**CONTRACT TYPE:** No Contract

**DESCRIPTION:**

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK	

Service Date 1/29/2021

**2 hours minimum portal to portal**

WORK PERFORMED	TECHNICIAN(S)
CALL COMPLETE CALIBRATIONS COMPLETE	JOE BIELARZ, JON MEYERS

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
10,000 x 5 lb	23635-aa	23635-aa	ALKON SCALE/INDICATOR / COMMAND	CEMENT
40,000 x 20 lb	23635-bb	23635-bb	ALKON SCALE/INDICATOR / COMMAND	AGG
4000 lbs x 2 lbs	23635-cc	23635-cc	ALKON SCALE/INDICATOR / COMMAND	WATER

PARTS LIST

Customer Signature

\_\_\_\_\_  
 Name:



**Customer:** SUPERIOR MATERIALS 39  
 SUP1148 6750 SIMS DR

**Certificate ID:** 23635-aa210129030039

**Test Date:** 01/29/2021

STERLING HEIGHTS, MI 48313

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	3.68
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	40
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass adjustment needed

**Comments:**

TRACEABILITY CERTIFICATE NUMBERS:			
Kit	NIST	Date	Date Due
Description			
GL1502	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1503	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1504	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1505	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1506	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1517	MI-06-19-14306	6/26/2019	6/26/2021
TRUCK 15 - 1000 LB			

**Technician(s):** JOE BIELARZ;JON MEYERS;

**Signature:**



Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.



**Customer:** SUPERIOR MATERIALS 39  
 SUP1148 6750 SIMS DR

**Certificate ID:** 23635-cc210129030306

**Test Date:** 01/29/2021

STERLING HEIGHTS, MI 48313

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	1.59
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	40
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

**Comments:**

TRACEABILITY CERTIFICATE NUMBERS:			
Kit	NIST	Date	Date Due
Description			
GL1502	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1503	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1504	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1505	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1506	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1517	MI-06-19-14306	6/26/2019	6/26/2021
TRUCK 15 - 1000 LB			

**Technician(s):** JOE BIELARZ;JON MEYERS;

**Signature:**



Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.





GREAT LAKES SCALE COMPANY  
 35380 UNION LAKE RD.  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586) 776-0517

**Customer:** SUPERIOR MATERIALS 39  
 SUP1148 6750 SIMS DR

STERLING HEIGHTS, MI 48313

**Certificate ID:** 23635-cc210129030306

**Test Date:** 1/29/2021

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	1.59
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	40
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

Comments:

TRACEABILITY CERTIFICATE NUMBERS:					
Kit	NIST	Date	Date Due	Description	
GL1502	MI-03-19-14172	3/19/2019	3/19/2021	TRUCK 15 - 1000 LB	
GL1503	MI-03-19-14172	3/19/2019	3/19/2021	TRUCK 15 - 1000 LB	
GL1504	MI-03-19-14172	3/19/2019	3/19/2021	TRUCK 15 - 1000 LB	
GL1505	MI-03-19-14172	3/19/2019	3/19/2021	TRUCK 15 - 1000 LB	
GL1506	MI-03-19-14172	3/19/2019	3/19/2021	TRUCK 15 - 1000 LB	
GL1517	MI-06-19-14306	6/26/2019	6/26/2021	TRUCK 15 - 1000 LB	

**Technician(s):** JOE BIELARZ;JON MEYERS;

**Signature:**

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.

# WORK ORDER

**Great Lakes Scale**

35380 UNION LAKE RD  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586)776-0517  
 Fax: (586)776-3169  
 Fed Tax ID: 38-3206225

**Work Order#** 0029762-1

**Request Date** 1/27/2021 12:34:00 PM

**Schedule Date** 1/28/2021 1:00:00 PM

**Bill To** SUP0896

SUP0896  
 SUPERIOR MATERIALS HOLDINGS LLC  
 PO BOX 2900  
 FARMINGTON HILLS, MI, 48336  
 (248) 788-8000

**Ship To** SUP1148

SUP1148  
 SUPERIOR MATERIALS 39  
 6750 SIMS DR  
 STERLING HEIGHTS, MI, 48313  
 (248) 788-8000

**Customer P.O**

**Requester** DAN THOMAS

**Phone** (248) 788-8048

**Email** dehthomas@superiormaterials.net

**CONTRACT TYPE:** No Contract

**DESCRIPTION:**

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK	

Service Date 1/29/2021

**2 hours minimum portal to portal**

WORK PERFORMED	TECHNICIAN(S)
CALL COMPLETE CALIBRATIONS COMPLETE	JOE BIELARZ, JON MEYERS

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
10,000 x 5 lb	23635-aa	23635-aa	ALKON SCALE/INDICATOR / COMMAND	CEMENT
40,000 x 20 lb	23635-bb	23635-bb	ALKON SCALE/INDICATOR / COMMAND	AGG
4000 lbs x 2 lbs	23635-cc	23635-cc	ALKON SCALE/INDICATOR / COMMAND	WATER

PARTS LIST

Customer Signature



Per james