



# WORK ORDER

**Great Lakes Scale**

35380 UNION LAKE RD  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586)776-0517  
 Fax: (586)776-3169  
 Fed Tax ID: 38-3206225

**Work Order#** 0029385-1  
**Request Date** 11/13/2020 2:28:00 PM  
**Schedule Date** 11/17/2020 3:00:00 PM

**Bill To** SUP0896  
 SUP0896  
 SUPERIOR MATERIALS HOLDINGS LLC  
 PO BOX 2900  
 FARMINGTON HILLS, MI, 48336  
 (248) 788-8000

**Ship To** SUP1192  
 SUP1192  
 SUPERIOR MATERIALS 36  
 2470 AUBURN RD  
 AUBURN HILLS, MI,  
 (248) 788-8000

**Customer P.O**  
**Requester** DAN THOMAS  
**Phone** (248) 788-8048  
**Email** dehthomas@superiormaterials.net

**CONTRACT TYPE:** No Contract

**DESCRIPTION:**

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK	

Service Date 11/16/2020

**2 hours minimum portal to portal**

WORK PERFORMED	TECHNICIAN(S)
CALL COMPLETE CALIBRATIONS COMPLETE	JOE BIELARZ, MIKE GLADSTONE

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
38,000 x 20 LB	CBS62803A	CBS62803A	ALKON SCALE/INDICATOR / COMMAND	Agg A
38000 lbs x 20 lbs	CBS62803B	CBS62803B	ALKON SCALE/INDICATOR / COMMAND	AGG B
6700 lbs x 5 lbs	CBS62803C	CBS62803C	ALKON SCALE/INDICATOR / COMMAND	CEMENT

PARTS LIST

Customer Signature

\_\_\_\_\_  
Name:



GREAT LAKES SCALE COMPANY  
 35380 UNION LAKE RD.  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586) 776-0517

# Certificate of Calibration

**Customer:** SUPERIOR MATERIALS 36  
 SUP1192 2470 AUBURN RD

**Certificate ID:** CBS62803A201116051102

**Test Date:** 11/16/2020

AUBURN HILLS, MI

**ISO Number:** MS-FM-CEM-001

(\*)Make: ALKON Model : COMMAND SN#: CBS62803A Capacity: 38000 LB Graduation: 20 LB

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
11/16/2020	05/18/2021	Semi-Annually	CBS62803A	III L
SCALE LOCATION		EQUIPMENT CONDITIONS		
Agg A		Working	Clean	
SELECT NUMBER OF PM DAYS:	CAPACITY	38000 UOM	LB	INCREMENT
				20

### AS FOUND TEST USING TEST WEIGHTS

### AS LEFT TEST USING TEST WEIGHTS

Weight Applied	Reading	Error	Weight Applied	Reading	Error
0	LB 0	LB 0		LB	LB
1000	LB 1000	LB 0		LB	LB
2000	LB 2000	LB 0		LB	LB
3000	LB 3000	LB 0		LB	LB
4000	LB 4000	LB 0		LB	LB
5000	LB 5000	LB 0		LB	LB
6000	LB 6000	LB 0		LB	LB
0	LB 0	LB 0		LB	LB
	LB	LB		LB	LB
	LB	LB		LB	LB

### AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
2000	LB 6100	LB 8100	LB 0
4000	LB 6100	LB 10100	LB 0
6000	LB 6100	LB 12100	LB 0
2000	LB 12020	LB 14020	LB 0
4000	LB 12020	LB 16020	LB 0
6000	LB 12020	LB 18020	LB 0
2000	LB 18060	LB 20060	LB 0
4000	LB 18060	LB 22060	LB 0
6000	LB 18060	LB 24060	LB 0
2000	LB 24000	LB 26000	LB 0
4000	LB 24000	LB 28000	LB 0
6000	LB 24000	LB 30000	LB 0
2000	LB 31000	LB 33000	LB 0
4000	LB 31000	LB 35000	LB 0
6000	LB 31000	LB 37000	LB 0
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB

**OVERALL TEST RESULTS PASS**

**Customer:** SUPERIOR MATERIALS 36  
 SUP1192 2470 AUBURN RD

AUBURN HILLS, MI

**Certificate ID:** CBS62803A201116051102

**Test Date:** 11/16/2020

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	15
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	55
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

**Comments:**

TRACEABILITY CERTIFICATE NUMBERS:			
Kit	NIST	Date	Date Due
Description			
GL1502	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1503	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1504	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1505	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1506	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1517	MI-06-19-14306	6/26/2019	6/26/2021
TRUCK 15 - 1000 LB			

**Technician(s):** JOE BIELARZ;MIKE GLADSTONE;

**Signature:**



Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.



GREAT LAKES SCALE COMPANY  
 35380 UNION LAKE RD.  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586) 776-0517

# Certificate of Calibration

**Customer:** SUPERIOR MATERIALS 36  
 SUP1192 2470 AUBURN RD

**Certificate ID:** CBS62803B201116051212

**Test Date:** 11/16/2020

AUBURN HILLS, MI

**ISO Number:** MS-FM-CEM-001

(\*)Make: ALKON Model : COMMAND SN#: CBS62803B Capacity: 38000 lb Graduation: 20 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
11/16/2020	05/18/2021	Semi-Annually	CBS62803B	III L
SCALE LOCATION		EQUIPMENT CONDITIONS		
AGG B		Working	Clean	
SELECT NUMBER OF PM DAYS:	CAPACITY	38000	UOM	lb
			INCREMENT	20

## AS FOUND TEST USING TEST WEIGHTS

## AS LEFT TEST USING TEST WEIGHTS

Weight Applied	Reading	Error	Weight Applied	Reading	Error
0	lb 0	lb 0		lb	lb
1000	lb 1000	lb 0		lb	lb
2000	lb 2000	lb 0		lb	lb
3000	lb 3000	lb 0		lb	lb
4000	lb 4000	lb 0		lb	lb
5000	lb 5000	lb 0		lb	lb
6000	lb 6000	lb 0		lb	lb
0	lb 0	lb 0		lb	lb
	lb	lb		lb	lb
	lb	lb		lb	lb

## AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
0	lb 0	lb 0	lb 0
2000	lb 8000	lb 10000	lb 0
4000	lb 8000	lb 12000	lb 0
6000	lb 8000	lb 14000	lb 0
2000	lb 16000	lb 18000	lb 0
4000	lb 16000	lb 20000	lb 0
6000	lb 16000	lb 22000	lb 0
2000	lb 24000	lb 25980	lb -20
4000	lb 24000	lb 28000	lb 0
6000	lb 24000	lb 29980	lb -20
2000	lb 30000	lb 32000	lb 0
4000	lb 30000	lb 34000	lb 0
6000	lb 30000	lb 36000	lb 0
	lb	lb	lb
	lb	lb	lb
	lb	lb	lb
	lb	lb	lb
	lb	lb	lb
	lb	lb	lb

**OVERALL TEST RESULTS PASS**

**Customer:** SUPERIOR MATERIALS 36  
 SUP1192 2470 AUBURN RD

AUBURN HILLS, MI

**Certificate ID:** CBS62803B201116051212

**Test Date:** 11/16/2020

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	15
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	55
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

**Comments:**

TRACEABILITY CERTIFICATE NUMBERS:			
Kit	NIST	Date	Date Due
Description			
GL1502	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1503	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1504	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1505	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1506	MI-03-19-14172	3/19/2019	3/19/2021
TRUCK 15 - 1000 LB			
GL1517	MI-06-19-14306	6/26/2019	6/26/2021
TRUCK 15 - 1000 LB			

**Technician(s):** JOE BIELARZ;MIKE GLADSTONE;

**Signature:**



Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.



GREAT LAKES SCALE COMPANY  
 35380 UNION LAKE RD.  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586) 776-0517

# Certificate of Calibration

**Customer:** SUPERIOR MATERIALS 36  
 SUP1192 2470 AUBURN RD

**Certificate ID:** CBS62803C20116051449

**Test Date:** 11/16/2020

AUBURN HILLS, MI

**ISO Number:** MS-FM-CEM-001

(\*)Make: ALKON Model : COMMAND SN#: CBS62803C Capacity: 6700 lb Graduation: 5 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
11/16/2020	05/18/2021	Semi-Annually	CBS62803C	III
SCALE LOCATION		EQUIPMENT CONDITIONS		
CEMENT		Working	Clean	
SELECT NUMBER OF PM DAYS:	CAPACITY	6700	UOM lb	INCREMENT 5

### AS FOUND TEST USING TEST WEIGHTS

### AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS						AS LEFT TEST USING TEST WEIGHTS					
Weight Applied		Reading		Error		Weight Applied		Reading		Error	
0	lb	0	lb	0	lb	0	lb	0	lb	0	lb
6000	lb	5995	lb	-5	lb	6000	lb	5995	lb	-5	lb
	lb	0	lb	0	lb	0	lb	0	lb	0	lb
	lb		lb		lb	2000	lb	2000	lb	0	lb
	lb		lb		lb	4000	lb	4000	lb	0	lb
	lb		lb		lb	6000	lb	6000	lb	0	lb
	lb		lb		lb	0	lb	0	lb	0	lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb

### AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb

OVERALL TEST RESULTS Pass



GREAT LAKES SCALE COMPANY  
 35380 UNION LAKE RD.  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586) 776-0517

**Customer:** SUPERIOR MATERIALS 36  
 SUP1192 2470 AUBURN RD

AUBURN HILLS, MI

**Certificate ID:** CBS62803C201116051449

**Test Date:** 11/16/2020

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	3.68
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	55
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

Comments:

TRACEABILITY CERTIFICATE NUMBERS:					
Kit	NIST	Date	Date Due	Description	
GL1502	MI-03-19-14172	3/19/2019	3/19/2021	TRUCK 15 - 1000 LB	
GL1503	MI-03-19-14172	3/19/2019	3/19/2021	TRUCK 15 - 1000 LB	
GL1504	MI-03-19-14172	3/19/2019	3/19/2021	TRUCK 15 - 1000 LB	
GL1505	MI-03-19-14172	3/19/2019	3/19/2021	TRUCK 15 - 1000 LB	
GL1506	MI-03-19-14172	3/19/2019	3/19/2021	TRUCK 15 - 1000 LB	
GL1517	MI-06-19-14306	6/26/2019	6/26/2021	TRUCK 15 - 1000 LB	

**Technician(s):** JOE BIELARZ;MIKE GLADSTONE;

**Signature:**

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.