

WORK ORDER

Great Lakes Scale

35380 UNION LAKE RD
 HARRISON TOWNSHIP, MI 48045
 Phone: (586)776-0517
 Fax: (586)776-3169
 Fed Tax ID: 38-3206225

Work Order# 0030391-1

Request Date 5/7/2021 9:25:00 AM

Schedule Date 5/12/2021 9:00:00 AM

Bill To SUP0896

SUP0896
 SUPERIOR MATERIALS HOLDINGS LLC
 PO BOX 2900
 FARMINGTON HILLS, MI, 48336
 (248) 788-8000

Ship To SUP1112

SUP1112
 SUPERIOR MATERIALS 102
 44922 GRAND RIVER AVE
 NOVI, MI, 48375
 (248) 788-8000

Customer P.O

Requester DAN THOMAS

Phone (248) 788-8048

Email dehthomas@superiormaterials.net

CONTRACT TYPE: No Contract

DESCRIPTION:

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK WITH WAYNE COUNTY 9AM	

Service Date 5/12/2021

2 hours minimum portal to portal

WORK PERFORMED
 Inspected tested calibrated and sealed 4 hopper scales. Certificates to follow.
 CALL COMPLETE

TECHNICIAN(S)
 ED GORHAM, MIKE GLADSTONE

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
40,000 x 20 lb	105134 AGG	105134 AGG	ALKON SCALE/INDICATOR / COMMAND	
30,000 x 20 lb	1802300105	1802300105	Rice Lake Scale/Indicator / 920i	
10,000 5 lb	105134 CEMENT	105134 CEMENT	ALKON SCALE/INDICATOR / COMMAND	
3,000 x 5 lb	105134 WATER	105134 WATER	ALKON SCALE/INDICATOR / COMMAND	

PARTS LIST

--	--	--	--	--

Customer Signature

 Name:



GREAT LAKES SCALE COMPANY
 35380 UNION LAKE RD.
 HARRISON TOWNSHIP, MI 48045
 Phone: (586) 776-0517

Certificate of Calibration

Customer: SUPERIOR MATERIALS 102
 SUP1112 44922 GRAND RIVER AVE

Certificate ID: 105134 CEMENT210512124252

Test Date: 05/12/2021

NOVI, MI 48375

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: 105134 CEMENT Capacity: 10000 lb Graduation: 5 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
05/12/2021			105134 CEMENT	III
SCALE LOCATION			EQUIPMENT CONDITIONS	
			Working	
SELECT NUMBER OF PM DAYS:		CAPACITY	10000 UOM	INCREMENT
			lb	5

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS						AS LEFT TEST USING TEST WEIGHTS					
Weight Applied	Reading		Error			Weight Applied	Reading		Error		
0	lb	0	lb	0	lb		lb	lb	0	lb	lb
1000	lb	1000	lb	0	lb		lb	lb	0	lb	lb
2000	lb	2000	lb	0	lb		lb	lb	0	lb	lb
3000	lb	3000	lb	0	lb		lb	lb	0	lb	lb
4000	lb	4000	lb	0	lb		lb	lb	0	lb	lb
5000	lb	5000	lb	0	lb		lb	lb	0	lb	lb
6000	lb	6000	lb	0	lb		lb	lb	0	lb	lb
	lb		lb	0	lb		lb	lb	0	lb	lb
	lb		lb	0	lb		lb	lb	0	lb	lb
	lb		lb	0	lb		lb	lb	0	lb	lb

AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
lb	lb	lb	0 lb
lb	lb	lb	0 lb
lb	lb	lb	0 lb
lb	lb	lb	0 lb
lb	lb	lb	0 lb
lb	lb	lb	0 lb
lb	lb	lb	0 lb
lb	lb	lb	0 lb
lb	lb	lb	0 lb
lb	lb	lb	0 lb
lb	lb	lb	0 lb
lb	lb	lb	0 lb
lb	lb	lb	0 lb
lb	lb	lb	0 lb
lb	lb	lb	0 lb
lb	lb	lb	0 lb
lb	lb	lb	0 lb
lb	lb	lb	0 lb
lb	lb	lb	0 lb
lb	lb	lb	0 lb

OVERALL TEST RESULTS PASS

Customer: SUPERIOR MATERIALS 102
 SUP1112 44922 GRAND RIVER AVE
 NOVI, MI 48375

Certificate ID: 105134 CEMENT210512124252
Test Date: 05/12/2021
ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	3.68
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	61
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

Comments:

Scale on zero, load tested and sealed.

TRACEABILITY CERTIFICATE NUMBERS:				
Kit	NIST	Date	Date Due	
Description				
GL1501	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1502	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1503	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1504	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1505	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1506	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				

Technician(s): ED GORHAM;MIKE GLADSTONE;



Signature:

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.



GREAT LAKES SCALE COMPANY
 35380 UNION LAKE RD.
 HARRISON TOWNSHIP, MI 48045
 Phone: (586) 776-0517

Certificate of Calibration

Customer: SUPERIOR MATERIALS 102
 SUP1112 44922 GRAND RIVER AVE

Certificate ID: 105134 AGG210512124358

Test Date: 05/12/2021

NOVI, MI 48375

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: 105134 AGG Capacity: 40000 lb Graduation: 20 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
05/12/2021			105134 AGG	III L
SCALE LOCATION		EQUIPMENT CONDITIONS		
		Working		
SELECT NUMBER OF PM DAYS:		CAPACITY	40000 UOM lb	INCREMENT 20

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS						AS LEFT TEST USING TEST WEIGHTS					
Weight Applied	Reading		Error			Weight Applied	Reading		Error		
0	lb	0	lb	0	lb	0	lb	0	lb	0	lb
6000	lb	6000	lb	0	lb	1000	lb	1000	lb	0	lb
0	lb	0	lb	0	lb	2000	lb	2000	lb	0	lb
	lb		lb	0	lb	3000	lb	3000	lb	0	lb
	lb		lb	0	lb	4000	lb	4000	lb	0	lb
	lb		lb	0	lb	5000	lb	5000	lb	0	lb
	lb		lb	0	lb	6000	lb	6000	lb	0	lb
	lb		lb	0	lb	0	lb	0	lb	0	lb
	lb		lb	0	lb		lb		lb	0	lb
	lb		lb	0	lb		lb		lb	0	lb

AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
2000	6000	8000	0
4000	6000	10000	0
6000	6000	12000	0
2000	12000	14000	0
4000	12000	16000	0
6000	12000	18000	0
2000	18000	20000	0
4000	18000	22000	0
6000	18000	24000	0
2000	24000	26000	0
4000	24000	28000	0
6000	24000	30000	0
2000	30000	32000	0
4000	30000	34000	0
6000	30000	36000	0
			0
			0
			0
			0

OVERALL TEST RESULTS PASS

Customer: SUPERIOR MATERIALS 102
 SUP1112 44922 GRAND RIVER AVE

 NOVI, MI 48375

Certificate ID: 105134 AGG210512124358
Test Date: 05/12/2021
ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	15
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	61
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

Comments:

Test aggregate scale, sealed.

TRACEABILITY CERTIFICATE NUMBERS:				
Kit	NIST	Date	Date Due	
Description				
GL1501	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1502	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1503	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1504	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1505	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1506	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				

Technician(s): ED GORHAM; MIKE GLADSTONE;

Signature:



Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.



Certificate of Calibration

Customer: SUPERIOR MATERIALS 102
 SUP1112 44922 GRAND RIVER AVE

Certificate ID: 105134 WATER210512124647

Test Date: 05/12/2021

NOVI, MI 48375

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: 105134 WATER Capacity: 3000 lb Graduation: 5 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
05/12/2021			105134 WATER	III
SCALE LOCATION		EQUIPMENT CONDITIONS		
		Working		
SELECT NUMBER OF PM DAYS:		CAPACITY	3000 UOM lb	INCREMENT 5

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS						AS LEFT TEST USING TEST WEIGHTS					
Weight Applied		Reading		Error		Weight Applied		Reading		Error	
0	lb	0	lb	0	lb		lb		lb	0	lb
1000	lb	1000	lb	0	lb		lb		lb	0	lb
2000	lb	2000	lb	0	lb		lb		lb	0	lb
3000	lb	3000	lb	0	lb		lb		lb	0	lb
	lb		lb	0	lb		lb		lb	0	lb
	lb		lb	0	lb		lb		lb	0	lb
	lb		lb	0	lb		lb		lb	0	lb
	lb		lb	0	lb		lb		lb	0	lb
	lb		lb	0	lb		lb		lb	0	lb
	lb		lb	0	lb		lb		lb	0	lb
	lb		lb	0	lb		lb		lb	0	lb

AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
	lb	lb	0 lb
	lb	lb	0 lb
	lb	lb	0 lb
	lb	lb	0 lb
	lb	lb	0 lb
	lb	lb	0 lb
	lb	lb	0 lb
	lb	lb	0 lb
	lb	lb	0 lb
	lb	lb	0 lb
	lb	lb	0 lb
	lb	lb	0 lb
	lb	lb	0 lb
	lb	lb	0 lb
	lb	lb	0 lb
	lb	lb	0 lb
	lb	lb	0 lb
	lb	lb	0 lb
	lb	lb	0 lb
	lb	lb	0 lb
	lb	lb	0 lb
	lb	lb	0 lb

OVERALL TEST RESULTS PASS

Customer: SUPERIOR MATERIALS 102
 SUP1112 44922 GRAND RIVER AVE

 NOVI, MI 48375

Certificate ID: 105134 WATER210512124647
Test Date: 05/12/2021
ISO Number: MS-FM-CEM-001

TEST INFORMATION


TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	3.39
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	61
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

Comments:

Scale on zero, load tested and sealed.

TRACEABILITY CERTIFICATE NUMBERS:				
Kit	NIST	Date	Date Due	
Description				
GL1501	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1502	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1503	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				

Technician(s): ED GORHAM;MIKE GLADSTONE;



Signature:

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.

WORK ORDER

Great Lakes Scale

35380 UNION LAKE RD
 HARRISON TOWNSHIP, MI 48045
 Phone: (586)776-0517
 Fax: (586)776-3169
 Fed Tax ID: 38-3206225

Work Order# 0030391-1

Request Date 5/7/2021 9:25:00 AM

Schedule Date 5/12/2021 9:00:00 AM

Bill To SUP0896

SUP0896
 SUPERIOR MATERIALS HOLDINGS LLC
 PO BOX 2900
 FARMINGTON HILLS, MI, 48336
 (248) 788-8000

Ship To SUP1112

SUP1112
 SUPERIOR MATERIALS 102
 44922 GRAND RIVER AVE
 NOVI, MI, 48375
 (248) 788-8000

Customer P.O

Requester DAN THOMAS

Phone (248) 788-8048

Email dehthomas@superiormaterials.net

CONTRACT TYPE: No Contract

DESCRIPTION:

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK WITH WAYNE COUNTY 9AM	

Service Date 5/12/2021

2 hours minimum portal to portal

WORK PERFORMED
 Inspected tested calibrated and sealed 4 hopper scales. Certificates to follow.
 CALL COMPLETE

TECHNICIAN(S)
 ED GORHAM, MIKE GLADSTONE

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
40,000 x 20 lb	105134 AGG	105134 AGG	ALKON SCALE/INDICATOR / COMMAND	
30,000 x 20 lb	1802300105	1802300105	Rice Lake Scale/Indicator / 920i	
10,000 5 lb	105134 CEMENT	105134 CEMENT	ALKON SCALE/INDICATOR / COMMAND	
3,000 x 5 lb	105134 WATER	105134 WATER	ALKON SCALE/INDICATOR / COMMAND	

PARTS LIST

--	--	--	--	--

Customer Signature



Ratliff