

WORK ORDER

Great Lakes Scale

35380 UNION LAKE RD
 HARRISON TOWNSHIP, MI 48045
 Phone: (586)776-0517
 Fax: (586)776-3169
 Fed Tax ID: 38-3206225

Work Order# 0030312-1
Request Date 4/22/2021 3:01:00 PM
Schedule Date 4/26/2021 1:00:00 PM

Bill To SUP0896
 SUP0896
 SUPERIOR MATERIALS HOLDINGS LLC
 PO BOX 2900
 FARMINGTON HILLS, MI, 48336
 (248) 788-8000

Ship To SUP1142
 SUP1142
 SUPERIOR MATERIALS 20
 G 5300 NORTH DORT HWY
 FLINT , MI, 48505
 (248) 788-8000

Customer P.O
Requester DAN THOMAS
Phone (248) 788-8048
Email dehthomas@superiormaterials.net

CONTRACT TYPE: No Contract

DESCRIPTION:

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK APROX 1PM	

Service Date 4/26/2021

2 hours minimum portal to portal

WORK PERFORMED	TECHNICIAN(S)
CALL COMPLETE Test cement and aggregate scales.	ED GORHAM, MIKE GLADSTONE

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
40000 lb x 20 lb	62822A	62822A	ALKON SCALE/INDICATOR / COMMAND	Aggregate
10000 lb x 5 lb	62822C	62822C	ALKON SCALE/INDICATOR / COMMAND	Cement

PARTS LIST

Customer Signature

 Name:



GREAT LAKES SCALE COMPANY
 35380 UNION LAKE RD.
 HARRISON TOWNSHIP, MI 48045
 Phone: (586) 776-0517

Certificate of Calibration

Customer: SUPERIOR MATERIALS 20
 SUP1142 G 5300 NORTH DORT HWY

Certificate ID: 62822C210426024752

Test Date: 04/26/2021

FLINT , MI 48505

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: 62822C Capacity: 10000 LB Graduation: 5 LB

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
04/26/2021			62822C	III
SCALE LOCATION		EQUIPMENT CONDITIONS		
Cement		Working		
SELECT NUMBER OF PM DAYS:	CAPACITY	10000 UOM	LB	INCREMENT
				5

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS						AS LEFT TEST USING TEST WEIGHTS					
Weight Applied		Reading		Error		Weight Applied		Reading		Error	
1000	LB	1000	LB	0	LB		LB		LB	0	LB
2000	LB	2000	LB	0	LB		LB		LB	0	LB
3000	LB	3000	LB	0	LB		LB		LB	0	LB
4000	LB	4000	LB	0	LB		LB		LB	0	LB
5000	LB	5000	LB	0	LB		LB		LB	0	LB
6000	LB	6000	LB	0	LB		LB		LB	0	LB
	LB		LB	0	LB		LB		LB	0	LB
	LB		LB	0	LB		LB		LB	0	LB
	LB		LB	0	LB		LB		LB	0	LB
	LB		LB	0	LB		LB		LB	0	LB

AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
LB	LB	LB	0
LB	LB	LB	0
LB	LB	LB	0
LB	LB	LB	0
LB	LB	LB	0
LB	LB	LB	0
LB	LB	LB	0
LB	LB	LB	0
LB	LB	LB	0
LB	LB	LB	0
LB	LB	LB	0
LB	LB	LB	0
LB	LB	LB	0
LB	LB	LB	0
LB	LB	LB	0
LB	LB	LB	0
LB	LB	LB	0
LB	LB	LB	0
LB	LB	LB	0
LB	LB	LB	0
LB	LB	LB	0
LB	LB	LB	0

OVERALL TEST RESULTS PASS

Customer: SUPERIOR MATERIALS 20
 SUP1142 G 5300 NORTH DORT HWY

FLINT , MI 48505

Certificate ID: 62822C210426024752

Test Date: 04/26/2021

ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	3.68
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	45
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

Comments:

Scale on zero, load tested and sealed

TRACEABILITY CERTIFICATE NUMBERS:			
Kit	NIST	Date	Date Due
Description			
GL0601	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0602	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0603	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0604	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0605	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0606	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			

Technician(s): ED GORHAM;MIKE GLADSTONE;



Signature:

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.



GREAT LAKES SCALE COMPANY
 35380 UNION LAKE RD.
 HARRISON TOWNSHIP, MI 48045
 Phone: (586) 776-0517

Certificate of Calibration

Customer: SUPERIOR MATERIALS 20
 SUP1142 G 5300 NORTH DORT HWY

Certificate ID: 62822A210426024208

Test Date: 04/26/2021

FLINT , MI 48505

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: 62822A Capacity: 40000 LB Graduation: 20 LB

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
04/26/2021			62822A	III
SCALE LOCATION		EQUIPMENT CONDITIONS		
Aggregate		Working		
SELECT NUMBER OF PM DAYS:		CAPACITY	40000 UOM LB	INCREMENT 20

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS					AS LEFT TEST USING TEST WEIGHTS				
Weight Applied	Reading	Error			Weight Applied	Reading	Error		
1000	LB 1000	LB 0	LB			LB	LB 0	LB	
2000	LB 2000	LB 0	LB			LB	LB 0	LB	
3000	LB 3000	LB 0	LB			LB	LB 0	LB	
	LB	LB 0	LB			LB	LB 0	LB	
	LB	LB 0	LB			LB	LB 0	LB	
	LB	LB 0	LB			LB	LB 0	LB	
	LB	LB 0	LB			LB	LB 0	LB	
	LB	LB 0	LB			LB	LB 0	LB	
	LB	LB 0	LB			LB	LB 0	LB	
	LB	LB 0	LB			LB	LB 0	LB	
	LB	LB 0	LB			LB	LB 0	LB	

AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
3000	LB 3000	LB 6000	LB 0
3000	LB 6000	LB 9000	LB 0
3000	LB 9000	LB 12000	LB 0
3000	LB 12000	LB 15000	LB 0
3000	LB 15000	LB 18000	LB 0
3000	LB 18000	LB 21000	LB 0
3000	LB 21000	LB 24000	LB 0
3000	LB 24000	LB 27000	LB 0
3000	LB 27000	LB 30000	LB 0
3000	LB 30000	LB 33000	LB 0
3000	LB 33000	LB 36000	LB 0
	LB	LB	LB 0
	LB	LB	LB 0
	LB	LB	LB 0
	LB	LB	LB 0
	LB	LB	LB 0
	LB	LB	LB 0
	LB	LB	LB 0
	LB	LB	LB 0

OVERALL TEST RESULTS PASS

Customer: SUPERIOR MATERIALS 20
 SUP1142 G 5300 NORTH DORT HWY

FLINT , MI 48505

Certificate ID: 62822A210426024208

Test Date: 04/26/2021

ISO Number: MS-FM-CEM-001

TEST INFORMATION


TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	¹⁵
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	45
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

Comments:

Scale on zero, load tested, build tested and sealed

TRACEABILITY CERTIFICATE NUMBERS:			
Kit	NIST	Date	Date Due
Description			
GL0601	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0602	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0603	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0604	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0605	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0606	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			

Technician(s): ED GORHAM;MIKE GLADSTONE;



Signature:

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.