



# WORK ORDER

**Great Lakes Scale**

35380 UNION LAKE RD  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586)776-0517  
 Fax: (586)776-3169  
 Fed Tax ID: 38-3206225

**Work Order#** 0030311-1  
**Request Date** 4/22/2021 3:00:00 PM  
**Schedule Date** 4/26/2021 9:00:00 AM

**Bill To** SUP0896  
 SUP0896  
 SUPERIOR MATERIALS HOLDINGS LLC  
 PO BOX 2900  
 FARMINGTON HILLS, MI, 48336  
 (248) 788-8000

**Ship To** SUP1173  
 SUP1173  
 SUPERIOR MATERIALS 32  
 8911 WEST JEFFERSON  
 DETROIT, MI,  
 (248) 788-8000

**Customer P.O**  
**Requester** DAN THOMAS  
**Phone** (248) 788-8048  
**Email** dehthomas@superiormaterials.net

**CONTRACT TYPE:** No Contract

**DESCRIPTION:**

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK	

Service Date 4/26/2021

**2 hours minimum portal to portal**

WORK PERFORMED	TECHNICIAN(S)
CALL COMPLETE Test and calibrate scales with Wayne County,all scales approved by Wayne County.	ED GORHAM, MIKE GLADSTONE

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
10,000 x 5 lb	62821 CEMENT	62821 CEMENT	ALKON SCALE/INDICATOR / COMMAND	
6000 x 1	662821 WATER	662821 WATER	ALKON SCALE/INDICATOR / COMMAND	
40,000 x 20 lb	662821 AGG A	662821 AGG A	ALKON SCALE/INDICATOR / COMMAND	AGG A

PARTS LIST

Customer Signature

\_\_\_\_\_  
Name:



**Customer:** SUPERIOR MATERIALS 32  
 SUP1173 8911 WEST JEFFERSON

DETROIT, MI

**Certificate ID:** 62821 CEMENT210426112954

**Test Date:** 04/26/2021

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	3.39
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	45
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass adjustment needed

**Comments:**

Test and calibrate scale,sealed.

**TRACEABILITY CERTIFICATE NUMBERS:**

Kit	NIST	Date	Date Due
Description			
GL0601	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0602	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0603	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0604	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0605	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0606	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			

**Technician(s):** ED GORHAM;MIKE GLADSTONE;

**Signature:**



Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.



**Customer:** SUPERIOR MATERIALS 32  
 SUP1173 8911 WEST JEFFERSON

DETROIT, MI

**Certificate ID:** 662821 WATER210426113409

**Test Date:** 04/26/2021

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	1.28
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	45
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

**Comments:**

Test scale,sealed.

TRACEABILITY CERTIFICATE NUMBERS:			
Kit	NIST	Date	Date Due
Description			
GL0601	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0602	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0603	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0604	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0605	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0606	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			

**Technician(s):** ED GORHAM;MIKE GLADSTONE;

**Signature:**



Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.



GREAT LAKES SCALE COMPANY  
 35380 UNION LAKE RD.  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586) 776-0517

# Certificate of Calibration

**Customer:** SUPERIOR MATERIALS 32  
 SUP1173 8911 WEST JEFFERSON

**Certificate ID:** 662821 AGG A210426113125

**Test Date:** 04/26/2021

DETROIT, MI

**ISO Number:** MS-FM-CEM-001

(\*)Make: ALKON Model : COMMAND SN#: 662821 AGG A Capacity: 40000 LB Graduation: 20 LB

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
04/26/2021			662821 AGG A	III L
SCALE LOCATION		EQUIPMENT CONDITIONS		
AGG A		Working		
SELECT NUMBER OF PM DAYS:		CAPACITY	40000 UOM LB	INCREMENT 20

### AS FOUND TEST USING TEST WEIGHTS

### AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS						AS LEFT TEST USING TEST WEIGHTS					
Weight Applied		Reading		Error		Weight Applied		Reading		Error	
1000	LB	1000	LB	0	LB		LB		LB	0	LB
2000	LB	2000	LB	0	LB		LB		LB	0	LB
3000	LB	3000	LB	0	LB		LB		LB	0	LB
4000	LB	4000	LB	0	LB		LB		LB	0	LB
5000	LB	5000	LB	0	LB		LB		LB	0	LB
6000	LB	6000	LB	0	LB		LB		LB	0	LB
	LB		LB	0	LB		LB		LB	0	LB
	LB		LB	0	LB		LB		LB	0	LB
	LB		LB	0	LB		LB		LB	0	LB
	LB		LB	0	LB		LB		LB	0	LB

### AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
2000	LB 6000	LB 8000	LB 0
4000	LB 6000	LB 10000	LB 0
6000	LB 6000	LB 12000	LB 0
2000	LB 12000	LB 14000	LB 0
4000	LB 12000	LB 16000	LB 0
6000	LB 12000	LB 18000	LB 0
2000	LB 18000	LB 20000	LB 0
4000	LB 18000	LB 22000	LB 0
6000	LB 18000	LB 24000	LB 0
2000	LB 24000	LB 26000	LB 0
4000	LB 24000	LB 28000	LB 0
6000	LB 24000	LB 30000	LB 0
2000	LB 30000	LB 32000	LB 0
4000	LB 30000	LB 34000	LB 0
6000	LB 30000	LB 36000	LB 0
	LB	LB	LB 0
	LB	LB	LB 0
	LB	LB	LB 0
	LB	LB	LB 0

**OVERALL TEST RESULTS PASS**

**Customer:** SUPERIOR MATERIALS 32  
 SUP1173 8911 WEST JEFFERSON

DETROIT, MI

**Certificate ID:** 662821 AGG A210426113125

**Test Date:** 04/26/2021

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	15
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	45
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

**Comments:**

Scale on zero, load tested, build tested and sealed.

TRACEABILITY CERTIFICATE NUMBERS:			
Kit	NIST	Date	Date Due
Description			
GL0601	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0602	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0603	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0604	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0605	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			
GL0606	MI-06-19-14292	6/13/2019	6/13/2021
TRUCK 6 - 1000 LB			

**Technician(s):** ED GORHAM;MIKE GLADSTONE;



**Signature:**

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.