



GREAT LAKES SCALE COMPANY  
 35380 UNION LAKE RD.  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586) 776-0517

# Certificate of Calibration

**Customer:** SUPERIOR MATERIALS 101  
 SUP1422 40 Floral

**Certificate ID:** 23635agg211110031552

**Test Date:** 11/10/2021

MT CLEMENS, MI 48043

**ISO Number:** MS-FM-CEM-001

(\*)Make: ALKON Model : COMMAND SN#: 23635agg Capacity: 40000 lb Graduation: 20 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
11/10/2021	05/12/2022	Semi-Annually	23635agg	III L
SCALE LOCATION		EQUIPMENT CONDITIONS		
		Working	Clean	
SELECT NUMBER OF PM DAYS:	CAPACITY	40000 UOM	lb	INCREMENT 20

## AS FOUND TEST USING TEST WEIGHTS

## AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS						AS LEFT TEST USING TEST WEIGHTS					
Weight Applied		Reading		Error		Weight Applied		Reading		Error	
0	lb	0	lb	0	lb		lb		lb		lb
2000	lb	2000	lb	0	lb		lb		lb		lb
4000	lb	4000	lb	0	lb		lb		lb		lb
6000	lb	6000	lb	0	lb		lb		lb		lb
0	lb	0	lb	0	lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb

## AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
6000	10000	16000	0
6000	20000	26000	0
6000	32000	38000	0

**OVERALL TEST RESULTS PASS**

**Customer:** SUPERIOR MATERIALS 101  
 SUP1422 40 Floral

**Certificate ID:** 23635agg211110031552

**Test Date:** 11/10/2021

MT CLEMENS, MI 48043

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	<sup>35</sup>
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	55
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

**Comments:**

TRACEABILITY CERTIFICATE NUMBERS:				
Kit	NIST	Date	Date Due	
Description				
GL1502	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1501	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1503	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1504	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1505	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1506	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				

**Technician(s):** JOE BIELARZ;MIKE GLADSTONE;

**Signature:**



Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.



GREAT LAKES SCALE COMPANY  
35380 UNION LAKE RD.  
HARRISON TOWNSHIP, MI 48045  
Phone: (586) 776-0517

# Certificate of Calibration

Certificate ID: 23635cem21110031722

Customer: SUPERIOR MATERIALS 101  
SUP1422 40 Floral

Test Date: 11/10/2021

MT CLEMENS, MI 48043

ISO Number: MS-FM-CEM-001

(\*)Make: ALKON Model : COMMAND SN#: 23635cem Capacity: 10000 lb Graduation: 5 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS	
11/10/2021	05/12/2022	Semi-Annually	23635cem	III	
SCALE LOCATION		EQUIPMENT CONDITIONS			
		Working	Clean		
SELECT NUMBER OF PM DAYS:	CAPACITY	10000 UOM	lb	INCREMENT	5

## AS FOUND TEST USING TEST WEIGHTS

## AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS						AS LEFT TEST USING TEST WEIGHTS					
Weight Applied		Reading		Error		Weight Applied		Reading		Error	
0	lb	0	lb	0	lb	0	lb	0	lb	0	lb
2000	lb	2080	lb	80	lb	2000	lb	2000	lb	0	lb
4000	lb	4240	lb	240	lb	4000	lb	4000	lb	0	lb
6000	lb	6320	lb	320	lb	6000	lb	6000	lb	0	lb
0	lb	0	lb	0	lb	0	lb	0	lb	0	lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb

## AS LEFT BUILD TEST

Test Weight Applied		Material Weight Applied		Reading		Error	
6000	lb	3300	lb	9300	lb	0	lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb

OVERALL TEST RESULTS PASS

**Customer:** SUPERIOR MATERIALS 101  
 SUP1422 40 Floral

**Certificate ID:** 23635cem211110031722

**Test Date:** 11/10/2021

MT CLEMENS, MI 48043

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	8.7
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	55
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

**Comments:**

TRACEABILITY CERTIFICATE NUMBERS:				
Kit	NIST	Date	Date Due	
Description				
GL1502	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1501	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1503	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1504	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1505	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1506	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				

**Technician(s):** JOE BIELARZ;MIKE GLADSTONE;

**Signature:**



Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.

# WORK ORDER

35380 UNION LAKE RD  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586)776-0517  
 Fax: (586)776-3169  
 Fed Tax ID: 38-3206225

**Work Order#** 0031443-1  
**Request Date** 11/5/2021 3:49:00 PM  
**Schedule Date** 11/10/2021 2:19:00 PM

**Bill To** SUP0896  
 SUP0896  
 SUPERIOR MATERIALS HOLDINGS LLC  
 PO BOX 2900  
 FARMINGTON HILLS, MI, 48336  
 (248) 788-8000

**Ship To** SUP1422  
 SUP1422  
 SUPERIOR MATERIALS 101  
 40 Floral  
 MT CLEMENS, MI, 48043  
 (248) 788-8000

**Customer P.O**  
**Requester** Dan Thomas  
**Phone** (248) 788-8048  
**Email** dehthomas@superiormaterials.net

**CONTRACT TYPE:** No Contract

**DESCRIPTION:**

TASK DESCRIPTION	SERVICE TYPE: CAL
PLANT CHECK ON SITE AT 2PM	

Service Date 11/10/2021

**2 hours minimum portal to portal**

WORK PERFORMED	TECHNICIAN(S)
CALL COMPLETE CALIBRATIONS COMPLETE	JOE BIELARZ, MIKE GLADSTONE

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
40,000 x 20 lb	23635agg	23635agg	ALKON SCALE/INDICATOR / COMMAND	
10,000 x 5 lb	23635cem	23635cem	ALKON SCALE/INDICATOR / COMMAND	

PARTS LIST

Customer Signature



Art d

# WORK ORDER

**Great Lakes Scale**

35380 UNION LAKE RD  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586)776-0517  
 Fax: (586)776-3169  
 Fed Tax ID: 38-3206225

**Work Order#** 0031443-1  
**Request Date** 11/5/2021 3:49:00 PM  
**Schedule Date** 11/10/2021 2:19:00 PM

**Bill To** SUP0896  
 SUP0896  
 SUPERIOR MATERIALS HOLDINGS LLC  
 PO BOX 2900  
 FARMINGTON HILLS, MI, 48336  
 (248) 788-8000

**Ship To** SUP1422  
 SUP1422  
 SUPERIOR MATERIALS 101  
 40 Floral  
 MT CLEMENS, MI, 48043  
 (248) 788-8000

**Customer P.O**  
**Requester** Dan Thomas  
**Phone** (248) 788-8048  
**Email** dehthomas@superiormaterials.net

**CONTRACT TYPE:** No Contract

**DESCRIPTION:**

TASK DESCRIPTION	SERVICE TYPE: CAL
PLANT CHECK ON SITE AT 2PM	

Service Date 11/10/2021

**2 hours minimum portal to portal**

WORK PERFORMED	TECHNICIAN(S)
CALL COMPLETE CALIBRATIONS COMPLETE	JOE BIELARZ, MIKE GLADSTONE

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
40,000 x 20 lb	23635agg	23635agg	ALKON SCALE/INDICATOR / COMMAND	
10,000 x 5 lb	23635cem	23635cem	ALKON SCALE/INDICATOR / COMMAND	

PARTS LIST

Customer Signature

\_\_\_\_\_  
 Name: