

# WORK ORDER

**Great Lakes Scale**

35380 UNION LAKE RD  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586)776-0517  
 Fax: (586)776-3169  
 Fed Tax ID: 38-3206225

**Work Order#** 0031240-1  
**Request Date** 10/1/2021 8:46:00 AM  
**Schedule Date** 10/7/2021 2:00:00 PM

**Bill To** SUP0896  
 SUP0896  
 SUPERIOR MATERIALS HOLDINGS LLC  
 PO BOX 2900  
 FARMINGTON HILLS, MI, 48336  
 (248) 788-8000

**Ship To** SUP1019  
 SUP1019  
 SUPERIOR MATERIALS 14  
 1384 LAKE LANSING RD  
 LANSING, MI,  
 (248) 788-8000

**Customer P.O**  
**Requester** DAN THOMAS  
**Phone** (248) 788-8048  
**Email** dehthomas@superiormaterials.net

**CONTRACT TYPE:** No Contract

**DESCRIPTION:**

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK 2PM	

Service Date 10/7/2021

**2 hours minimum portal to portal**

**WORK PERFORMED**  
 Inspected tested and sealed 2 hopper scales. Certificates to follow.  
 CALL COMPLETE

**TECHNICIAN(S)**  
 ED GORHAM, ROB DUBAY

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
40,000 lbs x 20 lbs	341agg	341agg	ALKON SCALE/INDICATOR / COMMAND	
10,000 lbs x 10 lbs	341cem	341cem	ALKON SCALE/INDICATOR / COMMAND	

**PARTS LIST**

--

Customer Signature

\_\_\_\_\_  
 Name:



GREAT LAKES SCALE COMPANY  
 35380 UNION LAKE RD.  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586) 776-0517

# Certificate of Calibration

**Customer:** SUPERIOR MATERIALS 14  
 SUP1019 1384 LAKE LANSING RD

**Certificate ID:** 341cem211007025541

**Test Date:** 10/07/2021

LANSING, MI

**ISO Number:** MS-FM-CEM-001

(\*)Make: ALKON Model : COMMAND SN#: 341cem Capacity: 10000 lb Graduation: 10 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
10/07/2021	04/08/2022	Semi-Annually	341cem	III
SCALE LOCATION		EQUIPMENT CONDITIONS		
		Working	Clean	
SELECT NUMBER OF PM DAYS:	CAPACITY	10000 UOM	lb	INCREMENT 10

### AS FOUND TEST USING TEST WEIGHTS

### AS LEFT TEST USING TEST WEIGHTS

Weight Applied	Reading	Error	Weight Applied	Reading	Error
0	lb 0	lb 0		lb	lb
2000	lb 2000	lb 0		lb	lb
4000	lb 4000	lb 0		lb	lb
6000	lb 6000	lb 0		lb	lb
0	lb 0	lb 0		lb	lb
	lb	lb		lb	lb
	lb	lb		lb	lb
	lb	lb		lb	lb
	lb	lb		lb	lb
	lb	lb		lb	lb

### AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
6000	3500	9500	0
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb
		lb	lb

**OVERALL TEST RESULTS PASS**

**Customer:** SUPERIOR MATERIALS 14  
 SUP1019 1384 LAKE LANSING RD

LANSING, MI

**Certificate ID:** 341cem211007025541

**Test Date:** 10/07/2021

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	6.68
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	65
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

**Comments:**

Scale on zero load tested build tested and sealed

TRACEABILITY CERTIFICATE NUMBERS:

Kit	NIST	Date	Date Due
Description			
GL1501	MI-03-21-15156	3/5/2021	3/5/2023
TRUCK 15 - 1000 LB			
GL1502	MI-03-21-15156	3/5/2021	3/5/2023
TRUCK 15 - 1000 LB			
GL1503	MI-03-21-15156	3/5/2021	3/5/2023
TRUCK 15 - 1000 LB			
GL1504	MI-03-21-15156	3/5/2021	3/5/2023
TRUCK 15 - 1000 LB			
GL1505	MI-03-21-15156	3/5/2021	3/5/2023
TRUCK 15 - 1000 LB			
GL1506	MI-03-21-15156	3/5/2021	3/5/2023
TRUCK 15 - 1000 LB			

**Technician(s):** ED GORHAM;ROB DUBAY;

**Signature:**



Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.



GREAT LAKES SCALE COMPANY  
 35380 UNION LAKE RD.  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586) 776-0517

# Certificate of Calibration

**Customer:** SUPERIOR MATERIALS 14  
 SUP1019 1384 LAKE LANSING RD

**Certificate ID:** 341agg211007030003

**Test Date:** 10/07/2021

LANSING, MI

**ISO Number:** MS-FM-CEM-001

(\*)Make: ALKON Model : COMMAND SN#: 341agg Capacity: 40000 lb Graduation: 20 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
10/07/2021	04/08/2022	Semi-Annually	341agg	III L
SCALE LOCATION			EQUIPMENT CONDITIONS	
			Working	Clean
SELECT NUMBER OF PM DAYS:		CAPACITY	40000 UOM	INCREMENT 20

## AS FOUND TEST USING TEST WEIGHTS

## AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS						AS LEFT TEST USING TEST WEIGHTS					
Weight Applied		Reading		Error		Weight Applied		Reading		Error	
0	lb	0	lb	0	lb		lb		lb		lb
2000	lb	2000	lb	0	lb		lb		lb		lb
4000	lb	4000	lb	0	lb		lb		lb		lb
6000	lb	6000	lb	0	lb		lb		lb		lb
0	lb	0	lb	0	lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb

## AS LEFT BUILD TEST

Test Weight Applied		Material Weight Applied		Reading		Error	
6000	lb	8000	lb	14000	lb	0	lb
6000	lb	16000	lb	22000	lb	0	lb
6000	lb	24000	lb	30000	lb	0	lb
6000	lb	32000	lb	38000	lb	0	lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb

**OVERALL TEST RESULTS PASS**

**Customer:** SUPERIOR MATERIALS 14  
 SUP1019 1384 LAKE LANSING RD  
  
 LANSING, MI

**Certificate ID:** 341agg211007030003  
**Test Date:** 10/07/2021  
**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	15
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	65
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

**Comments:**

Scale on zero load tested, build tested and sealed

TRACEABILITY CERTIFICATE NUMBERS:				
Kit	NIST	Date	Date Due	
Description				
GL1501	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1502	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1503	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1504	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1505	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1506	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				

**Technician(s):** ED GORHAM;ROB DUBAY;

**Signature:** 

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.

# WORK ORDER

35380 UNION LAKE RD  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586)776-0517  
 Fax: (586)776-3169  
 Fed Tax ID: 38-3206225

**Work Order#** 0031240-1  
**Request Date** 10/1/2021 8:46:00 AM  
**Schedule Date** 10/7/2021 2:00:00 PM

**Bill To** SUP0896  
 SUP0896  
 SUPERIOR MATERIALS HOLDINGS LLC  
 PO BOX 2900  
 FARMINGTON HILLS, MI, 48336  
 (248) 788-8000

**Ship To** SUP1019  
 SUP1019  
 SUPERIOR MATERIALS 14  
 1384 LAKE LANSING RD  
 LANSING, MI,  
 (248) 788-8000

**Customer P.O**  
**Requester** DAN THOMAS  
**Phone** (248) 788-8048  
**Email** dehthomas@superiormaterials.net

**CONTRACT TYPE:** No Contract

**DESCRIPTION:**

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK 2PM	

Service Date 10/7/2021

**2 hours minimum portal to portal**

**WORK PERFORMED**  
 Inspected tested and sealed 2 hopper scales. Certificates to follow.  
 CALL COMPLETE

**TECHNICIAN(S)**  
 ED GORHAM, ROB DUBAY

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
40,000 lbs x 20 lbs	341agg	341agg	ALKON SCALE/INDICATOR / COMMAND	
10,000 lbs x 10 lbs	341cem	341cem	ALKON SCALE/INDICATOR / COMMAND	

**PARTS LIST**

--

Customer Signature



Preadmore