

Customer: SUPERIOR MATERIALS 20
 SUP1142 G 5300 NORTH DORT HWY

FLINT , MI 48505

Certificate ID: 62822C211022032215

Test Date: 10/22/2021

ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	3.68
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	45
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

Comments:

TRACEABILITY CERTIFICATE NUMBERS:			
Kit	NIST	Date	Date Due
Description			
GL0606	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0604	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0603	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0602	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0601	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0605	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			

Technician(s): JOE AGY;ROB DUBAY;

Signature:



Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.



WORK ORDER

35380 UNION LAKE RD
HARRISON TOWNSHIP, MI 48045
Phone: (586)776-0517
Fax: (586)776-3169
Fed Tax ID: 38-3206225

Work Order# 0031354-1
Request Date 10/22/2021 7:32:00 AM
Schedule Date 10/22/2021 7:32:00 AM

Bill To SUP0896
SUP0896
SUPERIOR MATERIALS HOLDINGS LLC
PO BOX 2900
FARMINGTON HILLS, MI, 48336
(248) 788-8000

Ship To SUP1142
SUP1142
SUPERIOR MATERIALS 20
G 5300 NORTH DORT HWY
FLINT , MI, 48505
(248) 788-8000

Customer P.O
Requester DAN THOMAS
Phone (248) 788-8048
Email dehthomas@superiormaterials.net

CONTRACT TYPE: No Contract

DESCRIPTION:

TASK DESCRIPTION	SERVICE TYPE: CAL
Scale Check on Site at 2PM	

Service Date 10/22/2021

2 hours minimum portal to portal

WORK PERFORMED	TECHNICIAN(S)
CALL COMPLETE CALIBRATIONS COMPLETE 10/22/21 Joe A. / Rob D. Inspection went good, no issues found. Calibration done.	JOE AGY, ROB DUBAY

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
40000 lb x 20 lb	62822A	62822A	ALKON SCALE/INDICATOR / COMMAND	Aggregate
10000 lb x 5 lb	62822C	62822C	ALKON SCALE/INDICATOR / COMMAND	Cement

PARTS LIST

Customer Signature

Barber



GREAT LAKES SCALE COMPANY
 35380 UNION LAKE RD.
 HARRISON TOWNSHIP, MI 48045
 Phone: (586) 776-0517

Certificate of Calibration

Customer: SUPERIOR MATERIALS 20
 SUP1142 G 5300 NORTH DORT HWY

Certificate ID: 62822C211022032215

Test Date: 10/22/2021

FLINT , MI 48505

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: 62822C Capacity: 10000 LB Graduation: 5 LB

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
10/22/2021			62822C	III
SCALE LOCATION		EQUIPMENT CONDITIONS		
Cement		Working	Clean	
SELECT NUMBER OF PM DAYS:	CAPACITY	10000 UOM	LB	INCREMENT
				5

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS					AS LEFT TEST USING TEST WEIGHTS				
Weight Applied	Reading	Error	Weight Applied	Reading	Error				
1000	LB 1000	LB 0	1000	LB 1000	LB 0				
2000	LB 2000	LB 0	2000	LB 2000	LB 0				
3000	LB 3000	LB 0	3000	LB 3000	LB 0				
4000	LB 4000	LB 0	4000	LB 4000	LB 0				
5000	LB 5000	LB 0	5000	LB 5000	LB 0				
6000	LB 6000	LB 0	6000	LB 6000	LB 0				
	LB	LB		LB	LB				
	LB	LB		LB	LB				
	LB	LB		LB	LB				
	LB	LB		LB	LB				

AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
0	LB 0	LB 0	LB 0
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB

OVERALL TEST RESULTS PASS

Customer: SUPERIOR MATERIALS 20
 SUP1142 G 5300 NORTH DORT HWY

FLINT , MI 48505

Certificate ID: 62822C211022032215

Test Date: 10/22/2021

ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	3.68
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	45
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

Comments:

TRACEABILITY CERTIFICATE NUMBERS:			
Kit	NIST	Date	Date Due
Description			
GL0606	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0604	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0603	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0602	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0601	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0605	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			

Technician(s): JOE AGY;ROB DUBAY;

Signature:



Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.



GREAT LAKES SCALE COMPANY
 35380 UNION LAKE RD.
 HARRISON TOWNSHIP, MI 48045
 Phone: (586) 776-0517

Certificate of Calibration

Customer: SUPERIOR MATERIALS 20
 SUP1142 G 5300 NORTH DORT HWY

Certificate ID: 62822A211022032249

Test Date: 10/22/2021

FLINT , MI 48505

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: 62822A Capacity: 40000 LB Graduation: 20 LB

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
10/22/2021			62822A	III
SCALE LOCATION		EQUIPMENT CONDITIONS		
Aggregate		Working		
SELECT NUMBER OF PM DAYS:		CAPACITY	40000 UOM LB	INCREMENT 20

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS					AS LEFT TEST USING TEST WEIGHTS				
Weight Applied	Reading		Error		Weight Applied	Reading		Error	
1000	LB	1000	LB	0	1000	LB	1000	LB	0
2000	LB	2000	LB	0	2000	LB	2000	LB	0
3000	LB	3000	LB	0	3000	LB	3000	LB	0
	LB		LB			LB		LB	
	LB		LB			LB		LB	
	LB		LB			LB		LB	
	LB		LB			LB		LB	
	LB		LB			LB		LB	
	LB		LB			LB		LB	
	LB		LB			LB		LB	
	LB		LB			LB		LB	

AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied		Reading		Error	
3000	LB	3000	LB	6000	LB	0
3000	LB	6000	LB	9000	LB	0
3000	LB	9000	LB	12000	LB	0
3000	LB	12000	LB	15000	LB	0
3000	LB	15000	LB	18000	LB	0
3000	LB	18000	LB	21000	LB	0
3000	LB	21000	LB	24000	LB	0
3000	LB	24000	LB	27000	LB	0
3000	LB	27000	LB	30000	LB	0
3000	LB	30000	LB	33000	LB	0
3000	LB	33000	LB	36000	LB	0
	LB		LB		LB	
	LB		LB		LB	
	LB		LB		LB	
	LB		LB		LB	
	LB		LB		LB	
	LB		LB		LB	
	LB		LB		LB	
	LB		LB		LB	

OVERALL TEST RESULTS PASS



GREAT LAKES SCALE COMPANY
 35380 UNION LAKE RD.
 HARRISON TOWNSHIP, MI 48045
 Phone: (586) 776-0517

Customer: SUPERIOR MATERIALS 20
 SUP1142 G 5300 NORTH DORT HWY

FLINT , MI 48505

Certificate ID: 62822A211022032249

Test Date: 10/22/2021

ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	15
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	45
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

Comments:

TRACEABILITY CERTIFICATE NUMBERS:			
Kit	NIST	Date	Date Due
Description			
GL0606	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0604	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0603	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			

Technician(s): JOE AGY;ROB DUBAY;

Signature:

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.