

# WORK ORDER

**Great Lakes Scale**

35380 UNION LAKE RD  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586)776-0517  
 Fax: (586)776-3169  
 Fed Tax ID: 38-3206225

**Work Order#** 0031253-1  
**Request Date** 10/1/2021 3:39:00 PM  
**Schedule Date** 10/4/2021 2:09:00 PM

**Bill To** SUP0896  
 SUP0896  
 SUPERIOR MATERIALS HOLDINGS LLC  
 PO BOX 2900  
 FARMINGTON HILLS, MI, 48336  
 (248) 788-8000

**Ship To** SUP1104  
 SUP1104  
 SUPERIOR MATERIALS 35  
 39001 W HURON RD  
 ROMULUS, MI, 48174  
 (248) 788-8000

**Customer P.O**  
**Requester** DAN THOMAS  
**Phone** (248) 788-8048  
**Email** dehthomas@superiormaterials.net

**CONTRACT TYPE:** No Contract

**DESCRIPTION:**

TASK DESCRIPTION	SERVICE TYPE: CAL
Scale check with Wayne County GLS arrival at 1:30PM	

Service Date 10/4/2021

**2 hours minimum portal to portal**

WORK PERFORMED	TECHNICIAN(S)
Inspected tested calibrated and sealed 2 hopper scales with the county inspector. Certificates to follow. CALL COMPLETE	ED GORHAM, ROB DUBAY

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
10,000 x 10 LB	62821-a	62821-a	ALKON SCALE/INDICATOR / COMMAND	
40,000 x 20 LB	62821-b	62821-b	ALKON SCALE/INDICATOR / COMMAND	

PARTS LIST

Customer Signature

\_\_\_\_\_  
 Name:



GREAT LAKES SCALE COMPANY  
 35380 UNION LAKE RD.  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586) 776-0517

# Certificate of Calibration

**Customer:** SUPERIOR MATERIALS 35  
 SUP1104 39001 W HURON RD

**Certificate ID:** 62821-a211005072850

**Test Date:** 10/04/2021

ROMULUS, MI 48174

**ISO Number:** MS-FM-CEM-001

(\*)Make: ALKON Model : COMMAND SN#: 62821-a Capacity: 10000 LB Graduation: 10 LB

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
10/04/2021	4-04-2022		62821-a	III
SCALE LOCATION		EQUIPMENT CONDITIONS		
		Working		
SELECT NUMBER OF PM DAYS:	CAPACITY	10000 UOM	LB	INCREMENT
				10

### AS FOUND TEST USING TEST WEIGHTS

### AS LEFT TEST USING TEST WEIGHTS

Weight Applied	Reading	Error	Weight Applied	Reading	Error
0	LB 0	LB 0		LB	LB
1000	LB 1000	LB 0		LB	LB
2000	LB 2000	LB 0		LB	LB
3000	LB 3000	LB 0		LB	LB
4000	LB 4000	LB 0		LB	LB
5000	LB 5000	LB 0		LB	LB
6000	LB 6000	LB 0		LB	LB
	LB	LB		LB	LB
	LB	LB		LB	LB
	LB	LB		LB	LB

### AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
6000	LB 3130	LB 9130	LB 0
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB

**OVERALL TEST RESULTS PASS**



GREAT LAKES SCALE COMPANY  
 35380 UNION LAKE RD.  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586) 776-0517

**Customer:** SUPERIOR MATERIALS 35  
 SUP1104 39001 W HURON RD

ROMULUS, MI 48174

**Certificate ID:** 62821-a211005072850

**Test Date:** 10/04/2021

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	6.68
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	70
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

**Comments:**

Scale on zero, load tested, build tested and sealed.

TRACEABILITY CERTIFICATE NUMBERS:				
Kit	NIST	Date	Date Due	
Description				
GL1501	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1502	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1503	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1504	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1505	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1506	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				

**Technician(s):** ED GORHAM;ROB DUBAY;

**Signature:**

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.



GREAT LAKES SCALE COMPANY  
 35380 UNION LAKE RD.  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586) 776-0517

# Certificate of Calibration

**Customer:** SUPERIOR MATERIALS 35  
 SUP1104 39001 W HURON RD

**Certificate ID:** 62821-b211005073640

**Test Date:** 10/04/2021

ROMULUS, MI 48174

**ISO Number:** MS-FM-CEM-001

(\*)Make: ALKON Model : COMMAND SN#: 62821-b Capacity: 40000 LB Graduation: 20 LB

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
10/04/2021	4-04-2022		62821-b	III L
SCALE LOCATION		EQUIPMENT CONDITIONS		
		Working		
SELECT NUMBER OF PM DAYS:	CAPACITY	40000 UOM	LB	INCREMENT 20

## AS FOUND TEST USING TEST WEIGHTS

## AS LEFT TEST USING TEST WEIGHTS

Weight Applied	Reading	Error	Weight Applied	Reading	Error
0	LB 0	LB 0		LB	LB
1000	LB 1000	LB 0		LB	LB
2000	LB 2000	LB 0		LB	LB
3000	LB 3000	LB 0		LB	LB
4000	LB 4000	LB 0		LB	LB
5000	LB 5000	LB 0		LB	LB
6000	LB 6000	LB 0		LB	LB
	LB	LB		LB	LB
	LB	LB		LB	LB
	LB	LB		LB	LB

## AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
2000	LB 6000	LB 8000	LB 0
4000	LB 6000	LB 10000	LB 0
6000	LB 6000	LB 12000	LB 0
2000	LB 12000	LB 14000	LB 0
4000	LB 12000	LB 16000	LB 0
6000	LB 12000	LB 18000	LB 0
2000	LB 18000	LB 20000	LB 0
4000	LB 18000	LB 22000	LB 0
6000	LB 18000	LB 24000	LB 0
2000	LB 24000	LB 26000	LB 0
4000	LB 24000	LB 28000	LB 0
6000	LB 24000	LB 30000	LB 0
2000	LB 30000	LB 32000	LB 0
4000	LB 30000	LB 34000	LB 0
6000	LB 30000	LB 36000	LB 0
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB

**OVERALL TEST RESULTS PASS**



GREAT LAKES SCALE COMPANY  
 35380 UNION LAKE RD.  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586) 776-0517

**Customer:** SUPERIOR MATERIALS 35  
 SUP1104 39001 W HURON RD

ROMULUS, MI 48174

**Certificate ID:** 62821-b211005073640

**Test Date:** 10/04/2021

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	15
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	70
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

**Comments:**

Scale on zero, load tested, build tested and sealed.

**TRACEABILITY CERTIFICATE NUMBERS:**

Kit	NIST	Date	Date Due
Description			
GL1501	MI-03-21-15156	3/5/2021	3/5/2023
TRUCK 15 - 1000 LB			
GL1502	MI-03-21-15156	3/5/2021	3/5/2023
TRUCK 15 - 1000 LB			
GL1503	MI-03-21-15156	3/5/2021	3/5/2023
TRUCK 15 - 1000 LB			
GL1504	MI-03-21-15156	3/5/2021	3/5/2023
TRUCK 15 - 1000 LB			
GL1505	MI-03-21-15156	3/5/2021	3/5/2023
TRUCK 15 - 1000 LB			
GL1506	MI-03-21-15156	3/5/2021	3/5/2023
TRUCK 15 - 1000 LB			

**Technician(s):** ED GORHAM;ROB DUBAY;

**Signature:**

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.

# WORK ORDER

35380 UNION LAKE RD  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586)776-0517  
 Fax: (586)776-3169  
 Fed Tax ID: 38-3206225

**Work Order#** 0031253-1  
**Request Date** 10/1/2021 3:39:00 PM  
**Schedule Date** 10/4/2021 2:09:00 PM

**Bill To** SUP0896  
 SUP0896  
 SUPERIOR MATERIALS HOLDINGS LLC  
 PO BOX 2900  
 FARMINGTON HILLS, MI, 48336  
 (248) 788-8000

**Ship To** SUP1104  
 SUP1104  
 SUPERIOR MATERIALS 35  
 39001 W HURON RD  
 ROMULUS, MI, 48174  
 (248) 788-8000

**Customer P.O**  
**Requester** DAN THOMAS  
**Phone** (248) 788-8048  
**Email** dehthomas@superiormaterials.net

**CONTRACT TYPE:** No Contract

**DESCRIPTION:**

TASK DESCRIPTION	SERVICE TYPE: CAL
Scale check with Wayne County GLS arrival at 1:30PM	

Service Date 10/4/2021

**2 hours minimum portal to portal**

WORK PERFORMED	TECHNICIAN(S)
Inspected tested calibrated and sealed 2 hopper scales with the county inspector. Certificates to follow. CALL COMPLETE	ED GORHAM, ROB DUBAY

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
10,000 x 10 LB	62821-a	62821-a	ALKON SCALE/INDICATOR / COMMAND	
40,000 x 20 LB	62821-b	62821-b	ALKON SCALE/INDICATOR / COMMAND	

PARTS LIST

Customer Signature



Murray