



# WORK ORDER

**Great Lakes Scale**

35380 UNION LAKE RD  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586)776-0517  
 Fax: (586)776-3169  
 Fed Tax ID: 38-3206225

**Work Order#** 0031949-1  
**Request Date** 2/11/2022 3:58:00 PM  
**Schedule Date** 2/14/2022 8:28:00 AM

**Bill To** SUP0896

SUP0896  
 SUPERIOR MATERIALS HOLDINGS LLC  
 PO BOX 2900  
 FARMINGTON HILLS, MI, 48336  
 (248) 788-8000

**Ship To** SUP1193

SUP1193  
 SUPERIOR MATERIALS 38  
 901 S WAGNER RD  
 ANN ARBOR, MI, 48103  
 (248) 788-8000

**Customer P.O**

**Requester** DAN THOMAS  
**Phone** (248) 788-8048  
**Email** dehthomas@superiormaterials.net

**CONTRACT TYPE:** No Contract**DESCRIPTION:**

TASK DESCRIPTION	SERVICE TYPE: CAL
relined agg bin, wants scale check 2/14/22 - 9:30am on site	

Service Date 2/14/2022

**2 hours minimum portal to portal**

WORK PERFORMED	TECHNICIAN(S)
CALL COMPLETE CALIBRATIONS COMPLETE tested agg hopper since being re-lined. Tested and calibrated cement as well.	ED GORHAM, ROB DUBAY

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
40,000 lbs X 20 lbs	102559agg	102559agg	ALKON SCALE/INDICATOR / COMMAND	
10000 lbs x 5 lbs	102559cem	102559cem	ALKON SCALE/INDICATOR / COMMAND	

PARTS LIST

Customer Signature

Name:



GREAT LAKES SCALE COMPANY  
 35380 UNION LAKE RD.  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586) 776-0517

## Certificate of Calibration

**Certificate ID:** 102559cem220214114009

**Customer:** SUPERIOR MATERIALS 38  
 SUP1193 901 S WAGNER RD

**Test Date:** 02/14/2022

ANN ARBOR, MI 48103

**ISO Number:** MS-FM-CEM-001

(\*)Make: ALKON Model : COMMAND SN#: 102559cem Capacity: 10000 lb Graduation: 5 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
02/14/2022	08/16/2022	Semi-Annually	102559cem	III
SCALE LOCATION		EQUIPMENT CONDITIONS		
		Working	Clean	
SELECT NUMBER OF PM DAYS:	CAPACITY	10000 UOM	lb	INCREMENT
				5

### AS FOUND TEST USING TEST WEIGHTS

### AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS			AS LEFT TEST USING TEST WEIGHTS		
Weight Applied	Reading	Error	Weight Applied	Reading	Error
0	lb 0	lb 0	0	lb 0	lb 0
6000	lb 6010	lb 10	2000	lb 2000	lb 0
0	lb 0	lb 0	4000	lb 4000	lb 0
	lb	lb	6000	lb 6000	lb 0
	lb	lb	0	lb 0	lb 0
	lb	lb		lb	lb
	lb	lb		lb	lb
	lb	lb		lb	lb
	lb	lb		lb	lb
	lb	lb		lb	lb
	lb	lb		lb	lb

### AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
6000	3030	9030	0

**OVERALL TEST RESULTS PASS**



GREAT LAKES SCALE COMPANY  
 35380 UNION LAKE RD.  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586) 776-0517

# Certificate of Calibration

**Customer:** SUPERIOR MATERIALS 38  
 SUP1193 901 S WAGNER RD

**Certificate ID:** 102559agg220214114450

**Test Date:** 02/14/2022

ANN ARBOR, MI 48103

**ISO Number:** MS-FM-CEM-001

(\*)Make: ALKON Model : COMMAND SN#: 102559agg Capacity: 40000 lb Graduation: 20 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
02/14/2022	08/16/2022	Semi-Annually	102559agg	III L
SCALE LOCATION		EQUIPMENT CONDITIONS		
		Working	Clean	
SELECT NUMBER OF PM DAYS:	CAPACITY	40000	UOM	lb
			INCREMENT	20

### AS FOUND TEST USING TEST WEIGHTS

### AS LEFT TEST USING TEST WEIGHTS

Weight Applied	Reading	Error	Weight Applied	Reading	Error
0	lb 0	lb 0		lb	lb
2000	lb 2000	lb 0		lb	lb
4000	lb 4000	lb 0		lb	lb
6000	lb 6000	lb 0		lb	lb
0	lb 0	lb 0		lb	lb
	lb	lb		lb	lb
	lb	lb		lb	lb
	lb	lb		lb	lb
	lb	lb		lb	lb
	lb	lb		lb	lb

### AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
6000	lb 6000	lb 12000	lb 0
6000	lb 12000	lb 18000	lb 0
6000	lb 18000	lb 24000	lb 0
6000	lb 24000	lb 30000	lb 0
6000	lb 30000	lb 36000	lb 0
	lb	lb	lb
	lb	lb	lb
	lb	lb	lb
	lb	lb	lb
	lb	lb	lb
	lb	lb	lb
	lb	lb	lb
	lb	lb	lb
	lb	lb	lb
	lb	lb	lb
	lb	lb	lb
	lb	lb	lb
	lb	lb	lb

**OVERALL TEST RESULTS PASS**

**Customer:** SUPERIOR MATERIALS 38  
 SUP1193 901 S WAGNER RD

**Certificate ID:** 102559agg220214114450

**Test Date:** 02/14/2022

ANN ARBOR, MI 48103

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	<sup>35</sup>
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	14
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

**Comments:**

Scale on zero, load tested, build tested and sealed

**TRACEABILITY CERTIFICATE NUMBERS:**

Kit	NIST	Date	Date Due
Description			
GLSP1	MI-12-20-15021	12/8/2020	12/8/2022
1000 LB TRUCK 110 CLASS F			
GLSP2	MI-12-20-15021	12/8/2020	12/8/2022
1000 LB TRUCK 110 CLASS F			
GLSP3	MI-12-20-15021	12/8/2020	12/8/2022
1000 LB TRUCK 110 CLASS F			
GLSP4	MI-12-20-15021	12/8/2020	12/8/2022
1000 LB TRUCK 110 CLASS F			
GLSP5	MI-12-20-15021	12/8/2020	12/8/2022
1000 LB TRUCK 110 CLASS F			
GLSP6	MI-12-20-15021	12/8/2020	12/8/2022
1000 LB TRUCK 110 CLASS F			

**Technician(s):** ED GORHAM; ROB DUBAY;



**Signature:**

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.

**Customer:** SUPERIOR MATERIALS 38  
 SUP1193 901 S WAGNER RD

**Certificate ID:** 102559cem220214114009

**Test Date:** 02/14/2022

ANN ARBOR, MI 48103

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	8.7
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	14
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment


**Comments:**

Scale on zero, load tested calibrated, retested and sealed.

**TRACEABILITY CERTIFICATE NUMBERS:**

Kit	NIST	Date	Date Due
Description			
GLSP1	MI-12-20-15021	12/8/2020	12/8/2022
1000 LB TRUCK 110 CLASS F			
GLSP2	MI-12-20-15021	12/8/2020	12/8/2022
1000 LB TRUCK 110 CLASS F			
GLSP3	MI-12-20-15021	12/8/2020	12/8/2022
1000 LB TRUCK 110 CLASS F			
GLSP4	MI-12-20-15021	12/8/2020	12/8/2022
1000 LB TRUCK 110 CLASS F			
GLSP5	MI-12-20-15021	12/8/2020	12/8/2022
1000 LB TRUCK 110 CLASS F			
GLSP6	MI-12-20-15021	12/8/2020	12/8/2022
1000 LB TRUCK 110 CLASS F			

**Technician(s):** ED GORHAM;ROB DUBAY;



**Signature:**

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.