



GREAT LAKES SCALE COMPANY
 35380 UNION LAKE RD.
 HARRISON TOWNSHIP, MI 48045
 Phone: (586) 776-0517

Certificate of Calibration

Customer: SUPERIOR MATERIALS 2
 SUP1048 20565 HOOVER

Certificate ID: 72113-a220124105646

Test Date: 01/24/2022

DETROIT, MI

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: 72113-a Capacity: 10000 LB Graduation: 5 LB

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
01/24/2022			72113-a	III
SCALE LOCATION		EQUIPMENT CONDITIONS		
CEMENT SCALE		Working		
SELECT NUMBER OF PM DAYS:		CAPACITY	10000 UOM	LB INCREMENT 5

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS					AS LEFT TEST USING TEST WEIGHTS				
Weight Applied	Reading	Error	Weight Applied	Reading	Error				
0	LB 0	LB 0	0	LB 0	LB 0				
1000	LB 995	LB -5	1000	LB 1000	LB 0				
2000	LB 1990	LB -10	2000	LB 2000	LB 0				
3000	LB 2985	LB -15	3000	LB 3000	LB 0				
4000	LB 3980	LB -20	4000	LB 4000	LB 0				
5000	LB 4975	LB -25	5000	LB 5000	LB 0				
6000	LB 5970	LB -30	6000	LB 6000	LB 0				
	LB	LB		LB	LB				
	LB	LB		LB	LB				
	LB	LB		LB	LB				

AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
0	LB 0	LB 0	LB 0
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB

OVERALL TEST RESULTS PASS



GREAT LAKES SCALE COMPANY
 35380 UNION LAKE RD.
 HARRISON TOWNSHIP, MI 48045
 Phone: (586) 776-0517

Certificate of Calibration

Customer: SUPERIOR MATERIALS 2
 SUP1048 20565 HOOVER

Certificate ID: 72113-b220124110126

Test Date: 01/24/2022

DETROIT, MI

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: 72113-b Capacity: 40000 LB Graduation: 40 LB

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
01/24/2022			72113-b	III L
SCALE LOCATION		EQUIPMENT CONDITIONS		
AGG SCALE		Working		
SELECT NUMBER OF PM DAYS:	CAPACITY	40000 UOM	LB	INCREMENT 40

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

Weight Applied	Reading	Error	Weight Applied	Reading	Error
0	LB 0	LB 0	0	LB 0	LB 0
1000	LB 1000	LB 0	1000	LB 1000	LB 0
2000	LB 2000	LB 0	2000	LB 2000	LB 0
3000	LB 3000	LB 0	3000	LB 3000	LB 0
4000	LB 4000	LB 0	4000	LB 4000	LB 0
5000	LB 5000	LB 0	5000	LB 5000	LB 0
6000	LB 6000	LB 0	6000	LB 6000	LB 0
	LB	LB		LB	LB
	LB	LB		LB	LB
	LB	LB		LB	LB

AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
2000	LB 6000	LB 8000	LB 0
4000	LB 6000	LB 10000	LB 0
6000	LB 6000	LB 12000	LB 0
2000	LB 12000	LB 14000	LB 0
4000	LB 12000	LB 16000	LB 0
6000	LB 12000	LB 18000	LB 0
2000	LB 18000	LB 20000	LB 0
4000	LB 18000	LB 22000	LB 0
6000	LB 18000	LB 24000	LB 0
2000	LB 24000	LB 26000	LB 0
4000	LB 24000	LB 28000	LB 0
6000	LB 24000	LB 30000	LB 0
2000	LB 30000	LB 32000	LB 0
4000	LB 30000	LB 34000	LB 0
6000	LB 30000	LB 36000	LB 0
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB

OVERALL TEST RESULTS PASS

Customer: SUPERIOR MATERIALS 2
 SUP1048 20565 HOOVER

Certificate ID: 72113-b220124110126

Test Date: 01/24/2022

DETROIT, MI

ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	69
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	75
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass adjustment needed

Comments:

TRACEABILITY CERTIFICATE NUMBERS:

Kit	NIST	Date	Date Due
Description			
GL0602	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0601	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0603	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0604	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0606	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0605	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			

Technician(s): JOE AGY;ROB DUBAY;

Signature:



Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.

Customer: SUPERIOR MATERIALS 2
 SUP1048 20565 HOOVER

Certificate ID: 72113-a220124105646

Test Date: 01/24/2022

DETROIT, MI

ISO Number: MS-FM-CEM-001

TEST INFORMATION


TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	8.7
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	19
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass adjustment needed

Comments:

TRACEABILITY CERTIFICATE NUMBERS:

Kit	NIST	Date	Date Due
Description			
GL0602	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0601	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0603	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0604	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0606	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0605	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			

Technician(s): JOE AGY;ROB DUBAY;



Signature:

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.



WORK ORDER

35380 UNION LAKE RD
HARRISON TOWNSHIP, MI 48045
Phone: (586)776-0517
Fax: (586)776-3169
Fed Tax ID: 38-3206225

Work Order# 0031835-1
Request Date 1/21/2022 3:20:00 PM
Schedule Date 1/24/2022 1:49:00 PM

Bill To SUP0896
SUP0896
SUPERIOR MATERIALS HOLDINGS LLC
PO BOX 2900
FARMINGTON HILLS, MI, 48336
(248) 788-8000

Ship To SUP1048
SUP1048
SUPERIOR MATERIALS 2
20565 HOOVER
DETROIT, MI,
(248) 788-8000

Customer P.O
Requester DAN THOMAS
Phone (248) 788-8048
Email dehthomas@superiormaterials.net

CONTRACT TYPE: No Contract

DESCRIPTION:

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK WITH WAYNE COUNTY ON SITE AT 8:30AM	

Service Date 1/24/2022

2 hours minimum portal to portal


WORK PERFORMED
CALL COMPLETE CALIBRATIONS COMPLETE 01/24/22 joe a / rob d
Scale with Wayne County checked good, binding materials up on boot, adjusted chains around boot, cleaned around boot, scale retested good only 3 grads heavy after. Agg tested good. Wayne County showed up a little late. All set

TECHNICIAN(S)
JOE AGY, ROB DUBAY

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
10,000 x 5 LB	72113-a	72113-a	ALKON SCALE/INDICATOR / COMMAND	CEMENT SCALE
40,000 x 40 LB	72113-b	72113-b	ALKON SCALE/INDICATOR / COMMAND	AGG SCALE

PARTS LIST

Customer Signature



pete

WORK ORDER

Great Lakes Scale

35380 UNION LAKE RD
 HARRISON TOWNSHIP, MI 48045
 Phone: (586)776-0517
 Fax: (586)776-3169
 Fed Tax ID: 38-3206225

Work Order# 0031835-1
Request Date 1/21/2022 3:20:00 PM
Schedule Date 1/24/2022 1:49:00 PM

Bill To SUP0896
 SUP0896
 SUPERIOR MATERIALS HOLDINGS LLC
 PO BOX 2900
 FARMINGTON HILLS, MI, 48336
 (248) 788-8000

Ship To SUP1048
 SUP1048
 SUPERIOR MATERIALS 2
 20565 HOOVER
 DETROIT, MI,
 (248) 788-8000

Customer P.O
Requester DAN THOMAS
Phone (248) 788-8048
Email dehthomas@superiormaterials.net

CONTRACT TYPE: No Contract

DESCRIPTION:

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK WITH WAYNE COUNTY ON SITE AT 8:30AM	

Service Date 1/24/2022

2 hours minimum portal to portal

WORK PERFORMED
CALL COMPLETE CALIBRATIONS COMPLETE 01/24/22 joe a / rob d Scale with Wayne County checked good, binding materials up on boot, adjusted chains around boot, cleaned around boot, scale retested good only 3 grads heavy after. Agg tested good. Wayne County showed up a little late. All set

TECHNICIAN(S)
JOE AGY, ROB DUBAY

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
10,000 x 5 LB	72113-a	72113-a	ALKON SCALE/INDICATOR / COMMAND	CEMENT SCALE
40,000 x 40 LB	72113-b	72113-b	ALKON SCALE/INDICATOR / COMMAND	AGG SCALE

PARTS LIST

Customer Signature

 Name: