

# WORK ORDER

35380 UNION LAKE RD  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586)776-0517  
 Fax: (586)776-3169  
 Fed Tax ID: 38-3206225

**Work Order#** 0032056-1  
**Request Date** 3/2/2022 2:53:00 PM  
**Schedule Date** 3/4/2022 8:30:00 AM

**Bill To** SUP0896  
 SUP0896  
 SUPERIOR MATERIALS HOLDINGS LLC  
 PO BOX 2900  
 FARMINGTON HILLS, MI, 48336  
 (248) 788-8000

**Ship To** SUP0992  
 SUP0992  
 SUPERIOR MATERIALS 103  
 10147 GRAND RIVER  
 BRIGHTON, MI, 48116  
 (248) 788-8000

**Customer P.O**  
**Requester** DAN THOMAS  
**Phone**  
**Email** dehthomas@superiormaterials.net

**CONTRACT TYPE:** No Contract

**DESCRIPTION:**

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK	

Service Date 3/4/2022

**2 hours minimum portal to portal**

**WORK PERFORMED**  
 CALL COMPLETE  
 CALIBRATIONS COMPLETE set up and tested agg, cement, and water. Customer had to bring in a lift to cut away foam from an access hole for the agg scale. All tested good.

**TECHNICIAN(S)**  
 JOE BIELARZ, ROB DUBAY

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
4,000 x 5 LB	23635-c	23625-c	ALKON SCALE/INDICATOR / COMMAND	
10,000 x 5 LB	23635-a	23635-a	ALKON SCALE/INDICATOR / COMMAND	
45,000 x 20 LB	23635-b	23635-b	ALKON SCALE/INDICATOR / COMMAND	

**PARTS LIST**

--

Customer Signature



kyle



**Customer:** SUPERIOR MATERIALS 103  
 SUP0992 10147 GRAND RIVER

**Certificate ID:** 23635-c220304111735

**Test Date:** 03/04/2022

BRIGHTON, MI 48116

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	3.39
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	35
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

**Comments:**

TRACEABILITY CERTIFICATE NUMBERS:				
Kit	NIST	Date	Date Due	
Description				
GL1501	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1502	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1503	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1504	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1505	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1506	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				

**Technician(s):** JOE BIELARZ;ROB DUBAY;

**Signature:**



Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.



GREAT LAKES SCALE COMPANY  
 35380 UNION LAKE RD.  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586) 776-0517

# Certificate of Calibration

**Customer:** SUPERIOR MATERIALS 103  
 SUP0992 10147 GRAND RIVER

**Certificate ID:** 23635-b220304112253

**Test Date:** 03/04/2022

BRIGHTON, MI 48116

**ISO Number:** MS-FM-CEM-001

(\*)Make: ALKON Model : COMMAND SN#: 23635-b Capacity: 45000 LB Graduation: 20 LB

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
03/04/2022	09/03/2022	Semi-Annually	23635-b	III L
SCALE LOCATION			EQUIPMENT CONDITIONS	
			Working	Clean
<b>SELECT NUMBER OF PM DAYS:</b>		<b>CAPACITY</b>	45000 UOM	<b>INCREMENT</b> 20

### AS FOUND TEST USING TEST WEIGHTS

### AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS						AS LEFT TEST USING TEST WEIGHTS					
Weight Applied		Reading		Error		Weight Applied		Reading		Error	
0	LB	0	LB	0	LB	0	LB	0	LB	0	LB
6000	LB	6000	LB	0	LB	2000	LB	2000	LB	0	LB
0	LB	0	LB	0	LB	4000	LB	4000	LB	0	LB
	LB		LB		LB	6000	LB	6000	LB	0	LB
	LB		LB		LB	0	LB	0	LB	0	LB
	LB		LB		LB		LB		LB		LB
	LB		LB		LB		LB		LB		LB
	LB		LB		LB		LB		LB		LB
	LB		LB		LB		LB		LB		LB
	LB		LB		LB		LB		LB		LB
	LB		LB		LB		LB		LB		LB

### AS LEFT BUILD TEST

Test Weight Applied		Material Weight Applied		Reading		Error	
6000	LB	8000	LB	14000	LB	0	LB
6000	LB	16000	LB	22000	LB	0	LB
6000	LB	24000	LB	30000	LB	0	LB
6000	LB	30000	LB	36000	LB	0	LB
	LB		LB		LB		LB
	LB		LB		LB		LB
	LB		LB		LB		LB
	LB		LB		LB		LB
	LB		LB		LB		LB
	LB		LB		LB		LB
	LB		LB		LB		LB
	LB		LB		LB		LB
	LB		LB		LB		LB
	LB		LB		LB		LB
	LB		LB		LB		LB
	LB		LB		LB		LB
	LB		LB		LB		LB

**OVERALL TEST RESULTS PASS**

**Customer:** SUPERIOR MATERIALS 103  
 SUP0992 10147 GRAND RIVER

**Certificate ID:** 23635-c220304111735

**Test Date:** 03/04/2022

BRIGHTON, MI 48116

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	3.39
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	35
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

**Comments:**

TRACEABILITY CERTIFICATE NUMBERS:				
Kit	NIST	Date	Date Due	
Description				
GL1501	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1502	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1503	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1504	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1505	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1506	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				

**Technician(s):** JOE BIELARZ;ROB DUBAY;

**Signature:**



Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.



**Customer:** SUPERIOR MATERIALS 103  
 SUP0992 10147 GRAND RIVER

**Certificate ID:** 23635-a220304112044

**Test Date:** 03/04/2022

BRIGHTON, MI 48116

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	8.7
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	35
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

**Comments:**

TRACEABILITY CERTIFICATE NUMBERS:				
Kit	NIST	Date	Date Due	
Description				
GL1501	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1502	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1503	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1504	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1505	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1506	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				

**Technician(s):** JOE BIELARZ;ROB DUBAY;

**Signature:**



Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.