

# WORK ORDER

**Great Lakes Scale**

35380 UNION LAKE RD  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586)776-0517  
 Fax: (586)776-3169  
 Fed Tax ID: 38-3206225

**Work Order#** 0032254-1  
**Request Date** 3/31/2022 3:43:00 PM  
**Schedule Date** 4/1/2022 9:00:00 AM

**Bill To** SUP0896  
 SUP0896  
 SUPERIOR MATERIALS HOLDINGS LLC  
 PO BOX 2900  
 FARMINGTON HILLS, MI, 48336  
 (248) 788-8000

**Ship To** SUP1104  
 SUP1104  
 SUPERIOR MATERIALS 35  
 39001 W HURON RD  
 ROMULUS, MI, 48174  
 (248) 788-8000

**Customer P.O**  
**Requester** DAN THOMAS  
**Phone** (248) 788-8048  
**Email** dehthomas@superiormaterials.net

**CONTRACT TYPE:** No Contract

**DESCRIPTION:**

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK WITH WAYNE COUNTY	

Service Date 4/1/2022

**2 hours minimum portal to portal**

WORK PERFORMED
CALL COMPLETE CALIBRATIONS COMPLETE 04/01/22 joe a / joe b
Scale check with county all set. Agg scale had material build up to top binding and made weights light. Cleaned up weights were good. Cements weights were light as well. One boot was hanging off. Customer fixed. Checked all around found no issues. Calibrated all set.

TECHNICIAN(S)
JOE AGY, JOE BIELARZ

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
10,000 x 10 LB	62821-a	62821-a	ALKON SCALE/INDICATOR / COMMAND	
40,000 x 20 LB	62821-b	62821-b	ALKON SCALE/INDICATOR / COMMAND	

PARTS LIST

Customer Signature

\_\_\_\_\_  
 Name:





GREAT LAKES SCALE COMPANY  
 35380 UNION LAKE RD.  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586) 776-0517

# Certificate of Calibration

**Customer:** SUPERIOR MATERIALS 35  
 SUP1104 39001 W HURON RD

ROMULUS, MI 48174

**Certificate ID:** 62821-b220401123255

**Test Date:** 04/01/2022

**ISO Number:** MS-FM-CEM-001

(\*)Make: ALKON Model : COMMAND SN#: 62821-b Capacity: 40000 LB Graduation: 20 LB

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
04/01/2022	10/01/2022	Semi-Annually	62821-b	III L
SCALE LOCATION		EQUIPMENT CONDITIONS		
		Working	Clean	
SELECT NUMBER OF PM DAYS:	CAPACITY	40000	UOM	LB
			INCREMENT	20

## AS FOUND TEST USING TEST WEIGHTS

Weight Applied	Reading	Error
0	LB 0	LB 0
1000	LB 1000	LB 0
2000	LB 2000	LB 0
3000	LB 3000	LB 0
4000	LB 4000	LB 0
5000	LB 5000	LB 0
6000	LB 6000	LB 0
	LB	LB
	LB	LB
	LB	LB

## AS LEFT TEST USING TEST WEIGHTS

Weight Applied	Reading	Error
	LB	LB
	LB	LB
	LB	LB
	LB	LB
	LB	LB
	LB	LB
	LB	LB
	LB	LB
	LB	LB
	LB	LB

## AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
2000	LB 6000	LB 8000	LB 0
4000	LB 6000	LB 10000	LB 0
6000	LB 6000	LB 12000	LB 0
2000	LB 12000	LB 14000	LB 0
4000	LB 12000	LB 16000	LB 0
6000	LB 12000	LB 18000	LB 0
2000	LB 18000	LB 20000	LB 0
4000	LB 18000	LB 22000	LB 0
6000	LB 18000	LB 24000	LB 0
2000	LB 24000	LB 26000	LB 0
4000	LB 24000	LB 28000	LB 0
6000	LB 24000	LB 30000	LB 0
2000	LB 30000	LB 32000	LB 0
4000	LB 30000	LB 34000	LB 0
6000	LB 30000	LB 36000	LB 0
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB

**OVERALL TEST RESULTS PASS**

# WORK ORDER

**Great Lakes Scale**

35380 UNION LAKE RD  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586)776-0517  
 Fax: (586)776-3169  
 Fed Tax ID: 38-3206225

**Work Order#** 0032254-1  
**Request Date** 3/31/2022 3:43:00 PM  
**Schedule Date** 4/1/2022 9:00:00 AM

**Bill To** SUP0896  
  
 SUP0896  
 SUPERIOR MATERIALS HOLDINGS LLC  
 PO BOX 2900  
 FARMINGTON HILLS, MI, 48336  
 (248) 788-8000

**Ship To** SUP1104  
  
 SUP1104  
 SUPERIOR MATERIALS 35  
 39001 W HURON RD  
 ROMULUS, MI, 48174  
 (248) 788-8000

**Customer P.O**  
**Requester** DAN THOMAS  
**Phone** (248) 788-8048  
**Email** dehthomas@superiormaterials.net

**CONTRACT TYPE:** No Contract

**DESCRIPTION:**

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK WITH WAYNE COUNTY	

Service Date 4/1/2022

**2 hours minimum portal to portal**

WORK PERFORMED
CALL COMPLETE CALIBRATIONS COMPLETE 04/01/22 joe a / joe b  Scale check with county all set. Agg scale had material build up to top binding and made weights light. Cleaned up weights were good. Cements weights were light as well. One boot was hanging off. Customer fixed. Checked all around found no issues. Calibrated all set.

TECHNICIAN(S)
JOE AGY, JOE BIELARZ

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
10,000 x 10 LB	62821-a	62821-a	ALKON SCALE/INDICATOR / COMMAND	
40,000 x 20 LB	62821-b	62821-b	ALKON SCALE/INDICATOR / COMMAND	

PARTS LIST

Customer Signature

\_\_\_\_\_  
 Name:

**Customer:** SUPERIOR MATERIALS 35  
 SUP1104 39001 W HURON RD

ROMULUS, MI 48174

**Certificate ID:** 62821-b220401123255

**Test Date:** 04/01/2022

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	<sup>35</sup>
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	35
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

**Comments:**

TRACEABILITY CERTIFICATE NUMBERS:				
Kit	NIST	Date	Date Due	
Description				
GL1501	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1502	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1503	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1504	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1505	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1506	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				

**Technician(s):** JOE AGY;JOE BIELARZ;



**Signature:**

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.

**Customer:** SUPERIOR MATERIALS 35  
 SUP1104 39001 W HURON RD

ROMULUS, MI 48174

**Certificate ID:** 62821-a220401123444

**Test Date:** 04/01/2022

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	17
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	35
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

**Comments:**

TRACEABILITY CERTIFICATE NUMBERS:

Kit	NIST	Date	Date Due
Description			
GL1501	MI-03-21-15156	3/5/2021	3/5/2023
TRUCK 15 - 1000 LB			
GL1502	MI-03-21-15156	3/5/2021	3/5/2023
TRUCK 15 - 1000 LB			
GL1503	MI-03-21-15156	3/5/2021	3/5/2023
TRUCK 15 - 1000 LB			
GL1504	MI-03-21-15156	3/5/2021	3/5/2023
TRUCK 15 - 1000 LB			
GL1505	MI-03-21-15156	3/5/2021	3/5/2023
TRUCK 15 - 1000 LB			
GL1506	MI-03-21-15156	3/5/2021	3/5/2023
TRUCK 15 - 1000 LB			

**Technician(s):** JOE AGY;JOE BIELARZ;

**Signature:**



Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.