



# Certificate of Calibration

**Customer:** SUPERIOR MATERIALS 101  
 SUP1422 40 Floral

**Certificate ID:** 23635agg221108092736

**Test Date:** 11/02/2022

MT CLEMENS, MI 48043

**ISO Number:** MS-FM-CEM-001

(\*)Make: ALKON Model : COMMAND SN#: 23635agg Capacity: 40000 lb Graduation: 20 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
11/02/2022	05/04/2023	Semi-Annually	23635agg	III L
SCALE LOCATION			EQUIPMENT CONDITIONS	
			Working	Clean
SELECT NUMBER OF PM DAYS:		CAPACITY	40000 UOM	INCREMENT 20

**AS FOUND TEST USING TEST WEIGHTS**

**AS LEFT TEST USING TEST WEIGHTS**

AS FOUND TEST USING TEST WEIGHTS						AS LEFT TEST USING TEST WEIGHTS					
Weight Applied		Reading		Error		Weight Applied		Reading		Error	
0	lb	0	lb	0	lb	0	lb	0	lb	0	lb
2000	lb	2000	lb	0	lb	2000	lb	2000	lb	0	lb
4000	lb	4000	lb	0	lb	4000	lb	4000	lb	0	lb
6000	lb	6000	lb	0	lb	6000	lb	6000	lb	0	lb
0	lb	0	lb	0	lb	0	lb	0	lb	0	lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb

**AS LEFT BUILD TEST**

Test Weight Applied		Material Weight Applied		Reading		Error	
6000	lb	10000	lb	16000	lb	0	lb
6000	lb	20000	lb	26000	lb	0	lb
6000	lb	32000	lb	38000	lb	0	lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb

**OVERALL TEST RESULTS PASS**

**Customer:** SUPERIOR MATERIALS 101  
 SUP1422 40 Floral

**Certificate ID:** 23635agg221108092736

**Test Date:** 11/02/2022

MT CLEMENS, MI 48043

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	35
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	60
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment


**Comments:**

Scale on zero, load tested, build tested and sealed.

**TRACEABILITY CERTIFICATE NUMBERS:**

Kit	NIST	Date	Date Due
Description			
GL0605	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0606	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0604	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0603	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0602	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0601	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			

**Technician(s):** ED GORHAM;ROB DUBAY;



**Signature:**

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.

# WORK ORDER

**Great Lakes Scale**

35380 UNION LAKE RD  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586)776-0517  
 Fax: (586)776-3169  
 Fed Tax ID: 38-3206225

**Work Order#** 0033540-1  
**Request Date** 11/1/2022 3:28:00 PM  
**Schedule Date** 11/2/2022 2:00:00 PM

**Bill To** SUP0896  
 SUP0896  
 SUPERIOR MATERIALS HOLDINGS LLC  
 PO BOX 2900  
 FARMINGTON HILLS, MI, 48336  
 (248) 788-8000

**Ship To** SUP1422  
 SUP1422  
 SUPERIOR MATERIALS 101  
 40 Floral  
 MT CLEMENS, MI, 48043  
 (248) 788-8000

**Customer P.O**  
**Requester** Dan Thomas  
**Phone** (248) 788-8048  
**Email** dehthomas@superiormaterials.net

**CONTRACT TYPE:** No Contract

**DESCRIPTION:**

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK	

Service Date 11/2/2022

**2 hours minimum portal to portal**

**WORK PERFORMED**  
 Inspected tested calibrated and sealed 2 hopper scales. Certificates to follow.  
 CALL COMPLETE

**TECHNICIAN(S)**  
 ED GORHAM, ROB DUBAY

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
40,000 x 20 lb	23635agg	23635agg	ALKON SCALE/INDICATOR / COMMAND	
10,000 x 5 lb	23635cem	23635cem	ALKON SCALE/INDICATOR / COMMAND	

**PARTS LIST**

--

Customer Signature

\_\_\_\_\_  
 Name:

# WORK ORDER

35380 UNION LAKE RD  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586)776-0517  
 Fax: (586)776-3169  
 Fed Tax ID: 38-3206225

**Work Order#** 0033540-1  
**Request Date** 11/1/2022 3:28:00 PM  
**Schedule Date** 11/2/2022 2:00:00 PM

**Bill To** SUP0896  
 SUP0896  
 SUPERIOR MATERIALS HOLDINGS LLC  
 PO BOX 2900  
 FARMINGTON HILLS, MI, 48336  
 (248) 788-8000

**Ship To** SUP1422  
 SUP1422  
 SUPERIOR MATERIALS 101  
 40 Floral  
 MT CLEMENS, MI, 48043  
 (248) 788-8000

**Customer P.O**  
**Requester** Dan Thomas  
**Phone** (248) 788-8048  
**Email** dehthomas@superiormaterials.net

**CONTRACT TYPE:** No Contract

**DESCRIPTION:**

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK	

Service Date 11/2/2022

**2 hours minimum portal to portal**

**WORK PERFORMED**  
 Inspected tested calibrated and sealed 2 hopper scales. Certificates to follow.  
 CALL COMPLETE

**TECHNICIAN(S)**  
 ED GORHAM, ROB DUBAY

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
40,000 x 20 lb	23635agg	23635agg	ALKON SCALE/INDICATOR / COMMAND	
10,000 x 5 lb	23635cem	23635cem	ALKON SCALE/INDICATOR / COMMAND	

**PARTS LIST**

--

Customer Signature



Daggett



GREAT LAKES SCALE COMPANY  
 35380 UNION LAKE RD.  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586) 776-0517

# Certificate of Calibration

**Customer:** SUPERIOR MATERIALS 101  
 SUP1422 40 Floral

**Certificate ID:** 23635cem221103043043

**Test Date:** 11/02/2022

MT CLEMENS, MI 48043

**ISO Number:** MS-FM-CEM-001

(\*)Make: ALKON Model : COMMAND SN#: 23635cem Capacity: 10000 lb Graduation: 5 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
11/02/2022	05/04/2023	Semi-Annually	23635cem	III
SCALE LOCATION			EQUIPMENT CONDITIONS	
			Working	Clean
SELECT NUMBER OF PM DAYS:	CAPACITY	10000 UOM	lb	INCREMENT
				5

### AS FOUND TEST USING TEST WEIGHTS

### AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS						AS LEFT TEST USING TEST WEIGHTS					
Weight Applied		Reading		Error		Weight Applied		Reading		Error	
0	lb	0	lb	0	lb	0	lb	0	lb	0	lb
2000	lb	2000	lb	0	lb	2000	lb	2000	lb	0	lb
4000	lb	4000	lb	0	lb	4000	lb	4000	lb	0	lb
6000	lb	6000	lb	0	lb	6000	lb	6000	lb	0	lb
0	lb	0	lb	0	lb	0	lb	0	lb	0	lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb

### AS LEFT BUILD TEST

Test Weight Applied		Material Weight Applied		Reading		Error	
6000	lb	5000	lb	11000	lb	0	lb
6000	lb	10000	lb	16000	lb	0	lb
6000	lb	15000	lb	21000	lb	0	lb
6000	lb	20000	lb	26000	lb	0	lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb

**OVERALL TEST RESULTS PASS**

**Customer:** SUPERIOR MATERIALS 101  
 SUP1422 40 Floral

**Certificate ID:** 23635cem221103043043

**Test Date:** 11/02/2022

MT CLEMENS, MI 48043

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**


TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	8.7
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	60
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

**Comments:**

Scale on zero, load tested, build tested and sealed.

TRACEABILITY CERTIFICATE NUMBERS:				
Kit	NIST	Date	Date Due	
Description				
GL0605	MI-06-21-15309	6/7/2021	6/7/2023	
1000 LB TRUCK 6 CLASS F				
GL0606	MI-06-21-15309	6/7/2021	6/7/2023	
1000 LB TRUCK 6 CLASS F				
GL0604	MI-06-21-15309	6/7/2021	6/7/2023	
1000 LB TRUCK 6 CLASS F				
GL0603	MI-06-21-15309	6/7/2021	6/7/2023	
1000 LB TRUCK 6 CLASS F				
GL0602	MI-06-21-15309	6/7/2021	6/7/2023	
1000 LB TRUCK 6 CLASS F				
GL0601	MI-06-21-15309	6/7/2021	6/7/2023	
1000 LB TRUCK 6 CLASS F				

**Technician(s):** ED GORHAM;ROB DUBAY;



**Signature:**

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.