

WORK ORDER

Great Lakes Scale

35380 UNION LAKE RD
 HARRISON TOWNSHIP, MI 48045
 Phone: (586)776-0517
 Fax: (586)776-3169
 Fed Tax ID: 38-3206225

Work Order# 0033444-1
Request Date 10/13/2022 3:16:00 PM
Schedule Date 10/14/2022 1:15:00 PM

Bill To SUP0896
 SUP0896
 SUPERIOR MATERIALS HOLDINGS LLC
 PO BOX 2900
 FARMINGTON HILLS, MI, 48336
 (248) 788-8000

Ship To SUP1142
 SUP1142
 SUPERIOR MATERIALS 20
 G 5300 NORTH DORT HWY
 FLINT , MI, 48505
 (248) 788-8000

Customer P.O
Requester DAN THOMAS
Phone (248) 788-8048
Email dehthomas@superiormaterials.net

CONTRACT TYPE: No Contract

DESCRIPTION:

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK	
Brady Glomski 810-845-3373	

Service Date 10/14/2022

2 hours minimum portal to portal

WORK PERFORMED
 Inspected tested calibrated and sealed 2 hopper scales as per contract. Certificates to follow.
 CALL COMPLETE

TECHNICIAN(S)
 ED GORHAM, TREVOR CONAWAY-MAGGA

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
40000 lb x 20 lb	62822A	62822A	ALKON SCALE/INDICATOR / COMMAND	Aggregate
10000 lb x 5 lb	62822C	62822C	ALKON SCALE/INDICATOR / COMMAND	Cement

PARTS LIST

Customer Signature

 Name:



GREAT LAKES SCALE COMPANY
 35380 UNION LAKE RD.
 HARRISON TOWNSHIP, MI 48045
 Phone: (586) 776-0517

Certificate of Calibration

Customer: SUPERIOR MATERIALS 20
 SUP1142 G 5300 NORTH DORT HWY

Certificate ID: 62822A221014013618

Test Date: 10/14/2022

FLINT , MI 48505

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: 62822A Capacity: 40000 LB Graduation: 20 LB

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
10/14/2022	04/15/2023	Semi-Annually	62822A	III
SCALE LOCATION		EQUIPMENT CONDITIONS		
Aggregate		Working		
SELECT NUMBER OF PM DAYS:		CAPACITY	40000 UOM LB	INCREMENT 20

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS					AS LEFT TEST USING TEST WEIGHTS				
Weight Applied	Reading		Error		Weight Applied	Reading		Error	
1000	LB	1000	LB	0	1000	LB	1000	LB	0
2000	LB	2000	LB	0	2000	LB	2000	LB	0
3000	LB	3000	LB	0	3000	LB	3000	LB	0
	LB		LB			LB		LB	
	LB		LB			LB		LB	
	LB		LB			LB		LB	
	LB		LB			LB		LB	
	LB		LB			LB		LB	
	LB		LB			LB		LB	
	LB		LB			LB		LB	
	LB		LB			LB		LB	

AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
3000	3000	6000	0
3000	6000	9000	0
3000	9000	12000	0
3000	12000	15000	0
3000	15000	18000	0
3000	18000	21000	0
3000	21000	24000	0
3000	24000	27000	0
3000	27000	30000	0
3000	30000	33000	0
3000	33000	36000	0

OVERALL TEST RESULTS PASS

Customer: SUPERIOR MATERIALS 20
 SUP1142 G 5300 NORTH DORT HWY

FLINT , MI 48505

Certificate ID: 62822A221014013618

Test Date: 10/14/2022

ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	³⁵
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	51
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

Comments:

scale on zero, load tested and sealed.

TRACEABILITY CERTIFICATE NUMBERS:				
Kit	NIST	Date	Date Due	
Description				
GL1501	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1502	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				
GL1503	MI-03-21-15156	3/5/2021	3/5/2023	
TRUCK 15 - 1000 LB				

Technician(s): ED GORHAM;TREVOR CONAWAY-MAGGA;



Signature:

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.



GREAT LAKES SCALE COMPANY
 35380 UNION LAKE RD.
 HARRISON TOWNSHIP, MI 48045
 Phone: (586) 776-0517

Certificate of Calibration

Customer: SUPERIOR MATERIALS 20
 SUP1142 G 5300 NORTH DORT HWY

Certificate ID: 62822C221014013814

Test Date: 10/14/2022

FLINT , MI 48505

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: 62822C Capacity: 10000 LB Graduation: 5 LB

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
10/14/2022	04/15/2023	Semi-Annually	62822C	III
SCALE LOCATION		EQUIPMENT CONDITIONS		
Cement		Working	Clean	
SELECT NUMBER OF PM DAYS:		CAPACITY	10000 UOM	LB INCREMENT 5

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS			AS LEFT TEST USING TEST WEIGHTS				
Weight Applied	Reading		Error	Weight Applied	Reading		Error
0	LB	0	LB 0	1000	LB	1000	LB 0
6000	LB	6640	LB 640	2000	LB	2000	LB 0
0	LB	0	LB 0	3000	LB	3000	LB 0
	LB		LB	4000	LB	4000	LB 0
	LB		LB	5000	LB	5000	LB 0
	LB		LB	6000	LB	6000	LB 0
	LB		LB		LB		LB
	LB		LB		LB		LB
	LB		LB		LB		LB
	LB		LB		LB		LB

AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied		Reading		Error
0	LB	0	LB	0	LB 0
	LB		LB		LB
	LB		LB		LB
	LB		LB		LB
	LB		LB		LB
	LB		LB		LB
	LB		LB		LB
	LB		LB		LB
	LB		LB		LB
	LB		LB		LB
	LB		LB		LB
	LB		LB		LB
	LB		LB		LB
	LB		LB		LB
	LB		LB		LB
	LB		LB		LB
	LB		LB		LB
	LB		LB		LB
	LB		LB		LB
	LB		LB		LB
	LB		LB		LB

OVERALL TEST RESULTS PASS

Customer: SUPERIOR MATERIALS 20
 SUP1142 G 5300 NORTH DORT HWY

FLINT , MI 48505

Certificate ID: 62822C221014013814

Test Date: 10/14/2022

ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	8.7
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	51
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

Comments:

Scale on zero, load tested, calibrated, retested and sealed.

TRACEABILITY CERTIFICATE NUMBERS:

Kit	NIST	Date	Date Due
Description			
GL1501	MI-03-21-15156	3/5/2021	3/5/2023
TRUCK 15 - 1000 LB			
GL1502	MI-03-21-15156	3/5/2021	3/5/2023
TRUCK 15 - 1000 LB			
GL1505	MI-03-21-15156	3/5/2021	3/5/2023
TRUCK 15 - 1000 LB			
GL1506	MI-03-21-15156	3/5/2021	3/5/2023
TRUCK 15 - 1000 LB			
GL1503	MI-03-21-15156	3/5/2021	3/5/2023
TRUCK 15 - 1000 LB			
GL1504	MI-03-21-15156	3/5/2021	3/5/2023
TRUCK 15 - 1000 LB			

Technician(s): ED GORHAM;TREVOR CONAWAY-MAGGA;



Signature:

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.

WORK ORDER

35380 UNION LAKE RD
 HARRISON TOWNSHIP, MI 48045
 Phone: (586)776-0517
 Fax: (586)776-3169
 Fed Tax ID: 38-3206225

Work Order# 0033444-1
Request Date 10/13/2022 3:16:00 PM
Schedule Date 10/14/2022 1:15:00 PM

Bill To SUP0896
 SUP0896
 SUPERIOR MATERIALS HOLDINGS LLC
 PO BOX 2900
 FARMINGTON HILLS, MI, 48336
 (248) 788-8000

Ship To SUP1142
 SUP1142
 SUPERIOR MATERIALS 20
 G 5300 NORTH DORT HWY
 FLINT , MI, 48505
 (248) 788-8000

Customer P.O
Requester DAN THOMAS
Phone (248) 788-8048
Email dehthomas@superiormaterials.net

CONTRACT TYPE: No Contract

DESCRIPTION:

TASK DESCRIPTION SERVICE TYPE: CAL
 SCALE CHECK
 Brady Glomski 810-845-3373

Service Date 10/14/2022

2 hours minimum portal to portal

WORK PERFORMED
 Inspected tested calibrated and sealed 2 hopper scales as per contract. Certificates to follow.
 CALL COMPLETE

TECHNICIAN(S)
 ED GORHAM, TREVOR CONAWAY-MAGGA

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
40000 lb x 20 lb	62822A	62822A	ALKON SCALE/INDICATOR / COMMAND	Aggregate
10000 lb x 5 lb	62822C	62822C	ALKON SCALE/INDICATOR / COMMAND	Cement

PARTS LIST

Customer Signature



Barber