

# WORK ORDER

**Great Lakes Scale**

35380 UNION LAKE RD  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586)776-0517  
 Fax: (586)776-3169  
 Fed Tax ID: 38-3206225

**Work Order#** 0033362-1

**Request Date** 9/29/2022 2:28:00 PM

**Schedule Date** 9/30/2022 12:58:00 PM

**Bill To** SUP0896

SUP0896  
 SUPERIOR MATERIALS HOLDINGS LLC  
 PO BOX 2900  
 FARMINGTON HILLS, MI, 48336  
 (248) 788-8000

**Ship To** SUP1104

SUP1104  
 SUPERIOR MATERIALS 35  
 39001 W HURON RD  
 ROMULUS, MI, 48174  
 (248) 788-8000

**Customer P.O**

**Requester** DAN THOMAS

**Phone** (248) 788-8048

**Email** dehthomas@superiormaterials.net

**CONTRACT TYPE:** No Contract

**DESCRIPTION:**

TASK DESCRIPTION	SERVICE TYPE: CAL
PLANT CHECK WITH WAYNE COUNTY ONSITE AT 9:30AM	

Service Date 9/30/2022

**2 hours minimum portal to portal**

WORK PERFORMED
CALL COMPLETE CALIBRATIONS COMPLETE 9/30/22 Lucas and Rob
Had to wait for them to finish loading trucks to start. Found the agg scale to be off once it reached the higher thresholds of weight. Had to empty hopper, cal and retest. Cement scale checked out good. Ran the tests with Wayne county.

TECHNICIAN(S)
LUCAS SULLIVAN, ROB DUBAY

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
10,000 x 10 LB	62821-a	62821-a	ALKON SCALE/INDICATOR / COMMAND	
40,000 x 20 LB	62821-b	62821-b	ALKON SCALE/INDICATOR / COMMAND	

PARTS LIST

Customer Signature

\_\_\_\_\_  
 Name:

# WORK ORDER

35380 UNION LAKE RD  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586)776-0517  
 Fax: (586)776-3169  
 Fed Tax ID: 38-3206225

**Work Order#** 0033362-1  
**Request Date** 9/29/2022 2:28:00 PM  
**Schedule Date** 9/30/2022 12:58:00 PM

**Bill To** SUP0896  
 SUP0896  
 SUPERIOR MATERIALS HOLDINGS LLC  
 PO BOX 2900  
 FARMINGTON HILLS, MI, 48336  
 (248) 788-8000

**Ship To** SUP1104  
 SUP1104  
 SUPERIOR MATERIALS 35  
 39001 W HURON RD  
 ROMULUS, MI, 48174  
 (248) 788-8000

**Customer P.O**  
**Requester** DAN THOMAS  
**Phone** (248) 788-8048  
**Email** dehthomas@superiormaterials.net

**CONTRACT TYPE:** No Contract

**DESCRIPTION:**

TASK DESCRIPTION	SERVICE TYPE: CAL
PLANT CHECK WITH WAYNE COUNTY ONSITE AT 9:30AM	

Service Date 9/30/2022

**2 hours minimum portal to portal**

WORK PERFORMED
CALL COMPLETE CALIBRATIONS COMPLETE 9/30/22 Lucas and Rob
Had to wait for them to finish loading trucks to start. Found the agg scale to be off once it reached the higher thresholds of weight. Had to empty hopper, cal and retest. Cement scale checked out good. Ran the tests with Wayne county.

TECHNICIAN(S)
LUCAS SULLIVAN, ROB DUBAY

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
10,000 x 10 LB	62821-a	62821-a	ALKON SCALE/INDICATOR / COMMAND	
40,000 x 20 LB	62821-b	62821-b	ALKON SCALE/INDICATOR / COMMAND	

PARTS LIST

Customer Signature



Anthony

# Certificate of Calibration

**Customer:** SUPERIOR MATERIALS 35  
 SUP1104 39001 W HURON RD

**Certificate ID:** 62821-a221005074628

**Test Date:** 09/30/2022

ROMULUS, MI 48174

**ISO Number:** MS-FM-CEM-001

(\*)Make: ALKON Model : COMMAND SN#: 62821-a Capacity: 10000 LB Graduation: 10 LB

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
09/30/2022	04/01/2023	Semi-Annually	62821-a	III
SCALE LOCATION		EQUIPMENT CONDITIONS		
		Working	Clean	
<b>SELECT NUMBER OF PM DAYS:</b>	<b>CAPACITY</b>	10000	<b>UOM</b>	<b>INCREMENT</b>
			LB	10

### AS FOUND TEST USING TEST WEIGHTS

Weight Applied	Reading	Error
0	LB 0	LB 0
1000	LB 1000	LB 0
2000	LB 2000	LB 0
3000	LB 3000	LB 0
4000	LB 4000	LB 0
5000	LB 5000	LB 0
6000	LB 6000	LB 0
	LB	LB
	LB	LB
	LB	LB

### AS LEFT TEST USING TEST WEIGHTS

Weight Applied	Reading	Error
	LB	LB
	LB	LB
	LB	LB
	LB	LB
	LB	LB
	LB	LB
	LB	LB
	LB	LB
	LB	LB
	LB	LB
	LB	LB

### AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
6000	LB 0	LB 6000	LB 0
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB

**OVERALL TEST RESULTS PASS**

**Customer:** SUPERIOR MATERIALS 35  
 SUP1104 39001 W HURON RD  
 ROMULUS, MI 48174

**Certificate ID:** 62821-a221005074628  
**Test Date:** 09/30/2022  
**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	17
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	60
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

**Comments:**

TRACEABILITY CERTIFICATE NUMBERS:				
Kit	NIST	Date	Date Due	
Description				
GL0601	MI-06-21-15309	6/7/2021	6/7/2023	
1000 LB TRUCK 6 CLASS F				
GL0605	MI-06-21-15309	6/7/2021	6/7/2023	
1000 LB TRUCK 6 CLASS F				
GL0606	MI-06-21-15309	6/7/2021	6/7/2023	
1000 LB TRUCK 6 CLASS F				
GL0604	MI-06-21-15309	6/7/2021	6/7/2023	
1000 LB TRUCK 6 CLASS F				
GL0603	MI-06-21-15309	6/7/2021	6/7/2023	
1000 LB TRUCK 6 CLASS F				
GL0602	MI-06-21-15309	6/7/2021	6/7/2023	
1000 LB TRUCK 6 CLASS F				

**Technician(s):** LUCAS SULLIVAN;ROB DUBAY;

**Signature:** Lucas Sullivan  Digitally signed by Lucas Sullivan

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.



GREAT LAKES SCALE COMPANY  
 35380 UNION LAKE RD.  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586) 776-0517

# Certificate of Calibration

**Customer:** SUPERIOR MATERIALS 35  
 SUP1104 39001 W HURON RD

**Certificate ID:** 62821-b221005075154

**Test Date:** 09/30/2022

ROMULUS, MI 48174

**ISO Number:** MS-FM-CEM-001

(\*)Make: ALKON Model : COMMAND SN#: 62821-b Capacity: 40000 LB Graduation: 20 LB

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
09/30/2022	04/01/2023	Semi-Annually	62821-b	III L
SCALE LOCATION		EQUIPMENT CONDITIONS		
		Working	Clean	
SELECT NUMBER OF PM DAYS:	CAPACITY	40000	UOM	LB
			INCREMENT	20

## AS FOUND TEST USING TEST WEIGHTS

## AS LEFT TEST USING TEST WEIGHTS

Weight Applied	Reading	Error	Weight Applied	Reading	Error
0	LB 0	LB 0	0	LB 0	LB 0
1000	LB 1000	LB 0	1000	LB 1000	LB 0
2000	LB 2000	LB 0	2000	LB 2000	LB 0
3000	LB 3000	LB 0	3000	LB 3000	LB 0
4000	LB 4000	LB 0	4000	LB 4000	LB 0
5000	LB 4980	LB -20	5000	LB 5000	LB 0
6000	LB 5960	LB -40	6000	LB 6000	LB 0
	LB	LB		LB	LB
	LB	LB		LB	LB
	LB	LB		LB	LB

## AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
2000	LB 6000	LB 8000	LB 0
4000	LB 6000	LB 10000	LB 0
6000	LB 6000	LB 12000	LB 0
2000	LB 12000	LB 14000	LB 0
4000	LB 12000	LB 16000	LB 0
6000	LB 12000	LB 18000	LB 0
2000	LB 18000	LB 20000	LB 0
4000	LB 18000	LB 22000	LB 0
6000	LB 18000	LB 24000	LB 0
2000	LB 24000	LB 26000	LB 0
4000	LB 24000	LB 28000	LB 0
6000	LB 24000	LB 30000	LB 0
2000	LB 30000	LB 32000	LB 0
4000	LB 30000	LB 34000	LB 0
6000	LB 30000	LB 36000	LB 0
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB
	LB	LB	LB

**OVERALL TEST RESULTS PASS**

**Customer:** SUPERIOR MATERIALS 35  
 SUP1104 39001 W HURON RD  
 ROMULUS, MI 48174

**Certificate ID:** 62821-b221005075154  
**Test Date:** 09/30/2022  
**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	35
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	60
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

**Comments:**

TRACEABILITY CERTIFICATE NUMBERS:			
Kit	NIST	Date	Date Due
Description			
GL0601	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0605	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0606	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0604	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0603	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			
GL0602	MI-06-21-15309	6/7/2021	6/7/2023
1000 LB TRUCK 6 CLASS F			

**Technician(s):** LUCAS SULLIVAN;ROB DUBAY;

**Signature:** Lucas Sullivan  Digitally signed by Lucas Sullivan

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.