

# WORK ORDER

35380 UNION LAKE RD  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586)776-0517  
 Fax: (586)776-3169  
 Fed Tax ID: 38-3206225

**Work Order#** 0035963-1  
**Request Date** 1/5/2024 7:24:00 AM  
**Schedule Date** 1/5/2024 7:23:00 AM

**Bill To** SUP0896  
 SUP0896  
 SUPERIOR MATERIALS HOLDINGS LLC  
 PO BOX 2900  
 FARMINGTON HILLS, MI, 48336  
 (248) 788-8000

**Ship To** SUP1148  
 SUP1148  
 SUPERIOR MATERIALS 4909 (39)  
 6750 SIMS DR  
 STERLING HEIGHTS, MI, 48313  
 (248) 788-8000

**Customer P.O**  
**Requester** DAN THOMAS  
**Phone** (248) 788-8000  
**Email** dan.hargrave-thomas@vcimentos.com

**CONTRACT TYPE:** No Contract

**DESCRIPTION:**

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK	

Service Date 1/5/2024

**2 hours minimum portal to portal**

**WORK PERFORMED**  
 CALL COMPLETE  
 CALIBRATIONS COMPLETE set up tested and signed off stickers. cement, agg, and water tested good

**TECHNICIAN(S)**  
 LUCAS SULLIVAN, ROB DUBAY

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
10,000 x 5 lb	23635-aa	23635-aa	ALKON SCALE/INDICATOR / COMMAND	CEMENT
40,000 x 20 lb	23635-bb	23635-bb	ALKON SCALE/INDICATOR / COMMAND	AGG
4000 lbs x 2 lbs	23635-cc	23635-cc	ALKON SCALE/INDICATOR / COMMAND	WATER

**PARTS LIST**

--

Customer Signature



jason



GREAT LAKES SCALE COMPANY  
 35380 UNION LAKE RD.  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586) 776-0517

# Certificate of Calibration

**Customer:** SUPERIOR MATERIALS 4909 (39)  
 SUP1148 6750 SIMS DR

**Certificate ID:** 23635-aa240105101552

**Test Date:** 01/05/2024

STERLING HEIGHTS, MI 48313

**ISO Number:** MS-FM-CEM-001

(\*)Make: ALKON Model : COMMAND SN#: 23635-aa Capacity: 10000 lb Graduation: 5 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
01/05/2024	07/06/2024	Semi-Annually	23635-aa	III
SCALE LOCATION		EQUIPMENT CONDITIONS		
CEMENT		Working	Clean	
SELECT NUMBER OF PM DAYS:	CAPACITY	10000 UOM	lb	INCREMENT
				5

## AS FOUND TEST USING TEST WEIGHTS

## AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS						AS LEFT TEST USING TEST WEIGHTS					
Weight Applied	Reading		Error			Weight Applied	Reading		Error		
0	lb	0	lb	0	lb	0	lb	0	lb	0	lb
1000	lb	1000	lb	0	lb	1000	lb	1000	lb	0	lb
2000	lb	2000	lb	0	lb	2000	lb	2000	lb	0	lb
3000	lb	3000	lb	0	lb	3000	lb	3000	lb	0	lb
4000	lb	4000	lb	0	lb	4000	lb	4000	lb	0	lb
5000	lb	5000	lb	0	lb	5000	lb	5000	lb	0	lb
6000	lb	6000	lb	0	lb	6000	lb	6000	lb	0	lb
0	lb	0	lb	0	lb	0	lb	0	lb	0	lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb

## AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied		Reading		Error	
6000	lb	0	lb	6000	lb	0
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	
	lb		lb		lb	

**OVERALL TEST RESULTS PASS**

**Customer:** SUPERIOR MATERIALS 4909 (39)  
 SUP1148 6750 SIMS DR

**Certificate ID:** 23635-aa240105101552

**Test Date:** 01/05/2024

STERLING HEIGHTS, MI 48313

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	8.7
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	60
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

**Comments:**

TRACEABILITY CERTIFICATE NUMBERS:				
Kit	NIST	Date	Date Due	
Description				
GL1501	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15 - 1000 LB				
GL1502	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15 - 1000 LB				
GL1503	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15 - 1000 LB				
GL1504	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15 - 1000 LB				
GL1505	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15 - 1000 LB				
GL1506	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15 - 1000 LB				

**Technician(s):** LUCAS SULLIVAN;ROB DUBAY;

**Signature:** ROBERT DUBAY  Digitally signed by ROBERT DUBAY

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.



GREAT LAKES SCALE COMPANY  
 35380 UNION LAKE RD.  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586) 776-0517

# Certificate of Calibration

**Customer:** SUPERIOR MATERIALS 4909 (39)  
 SUP1148 6750 SIMS DR

**Certificate ID:** 23635-bb240105101727

**Test Date:** 01/05/2024

STERLING HEIGHTS, MI 48313

**ISO Number:** MS-FM-CEM-001

(\*)Make: ALKON Model : COMMAND SN#: 23635-bb Capacity: 40000 lb Graduation: 20 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
01/05/2024	07/06/2024	Semi-Annually	23635-bb	III L
SCALE LOCATION		EQUIPMENT CONDITIONS		
AGG		Working		
SELECT NUMBER OF PM DAYS:		CAPACITY	40000 UOM lb	INCREMENT 20

## AS FOUND TEST USING TEST WEIGHTS

## AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS						AS LEFT TEST USING TEST WEIGHTS					
Weight Applied	Reading		Error			Weight Applied	Reading		Error		
0	lb	0	lb	0	lb	0	lb	0	lb	0	lb
1000	lb	1000	lb	0	lb	1000	lb	1000	lb	0	lb
2000	lb	2000	lb	0	lb	2000	lb	2000	lb	0	lb
3000	lb	3000	lb	0	lb	3000	lb	3000	lb	0	lb
4000	lb	4000	lb	0	lb	4000	lb	4000	lb	0	lb
5000	lb	5000	lb	0	lb	5000	lb	5000	lb	0	lb
6000	lb	6000	lb	0	lb	6000	lb	6000	lb	0	lb
0	lb	0	lb	0	lb	0	lb	0	lb	0	lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb

## AS LEFT BUILD TEST

Test Weight Applied		Material Weight Applied		Reading		Error	
2000	lb	10000	lb	12000	lb	0	lb
4000	lb	10000	lb	14000	lb	0	lb
6000	lb	10000	lb	16000	lb	0	lb
2000	lb	20000	lb	22000	lb	0	lb
4000	lb	20000	lb	24000	lb	0	lb
6000	lb	20000	lb	26000	lb	0	lb
2000	lb	30000	lb	32000	lb	0	lb
4000	lb	30000	lb	34000	lb	0	lb
6000	lb	30000	lb	36000	lb	0	lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb

**OVERALL TEST RESULTS PASS**

**Customer:** SUPERIOR MATERIALS 4909 (39)  
 SUP1148 6750 SIMS DR

**Certificate ID:** 23635-bb240105101727

**Test Date:** 01/05/2024

STERLING HEIGHTS, MI 48313

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	35
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	60
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

**Comments:**

TRACEABILITY CERTIFICATE NUMBERS:				
Kit	NIST	Date	Date Due	
Description				
GL1501	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15 - 1000 LB				
GL1502	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15 - 1000 LB				
GL1503	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15 - 1000 LB				
GL1504	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15 - 1000 LB				
GL1505	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15 - 1000 LB				
GL1506	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15 - 1000 LB				

**Technician(s):** LUCAS SULLIVAN;ROB DUBAY;

**Signature:** ROBERT DUBAY Digitally signed by ROBERT DUBAY

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.



GREAT LAKES SCALE COMPANY  
 35380 UNION LAKE RD.  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586) 776-0517

# Certificate of Calibration

**Customer:** SUPERIOR MATERIALS 4909 (39)  
 SUP1148 6750 SIMS DR

**Certificate ID:** 23635-cc240105102117

**Test Date:** 01/05/2024

STERLING HEIGHTS, MI 48313

**ISO Number:** MS-FM-CEM-001

(\*)Make: ALKON Model : COMMAND SN#: 23635-cc Capacity: 4000 lb Graduation: 2 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
01/05/2024	07/06/2024	Semi-Annually	23635-cc	III
SCALE LOCATION		EQUIPMENT CONDITIONS		
WATER		Working		
SELECT NUMBER OF PM DAYS:		CAPACITY	4000 UOM lb	INCREMENT 2

### AS FOUND TEST USING TEST WEIGHTS

### AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS						AS LEFT TEST USING TEST WEIGHTS					
Weight Applied		Reading		Error		Weight Applied		Reading		Error	
0	lb	0	lb	0	lb	0	lb	0	lb	0	lb
100	lb	100	lb	0	lb	100	lb	100	lb	0	lb
200	lb	200	lb	0	lb	200	lb	200	lb	0	lb
300	lb	300	lb	0	lb	300	lb	300	lb	0	lb
200	lb	200	lb	0	lb	200	lb	200	lb	0	lb
100	lb	100	lb	0	lb	100	lb	100	lb	0	lb
0	lb	0	lb	0	lb	0	lb	0	lb	0	lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb

### AS LEFT BUILD TEST

Test Weight Applied		Material Weight Applied		Reading		Error	
0	lb	0	lb	0	lb	0	lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb

**OVERALL TEST RESULTS PASS**

**Customer:** SUPERIOR MATERIALS 4909 (39)  
 SUP1148 6750 SIMS DR

**Certificate ID:** 23635-cc240105102117

**Test Date:** 01/05/2024

STERLING HEIGHTS, MI 48313

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	3.5
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	60
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

**Comments:**  
 N/A

TRACEABILITY CERTIFICATE NUMBERS:				
Kit	NIST	Date	Date Due	
Description				
GL1507	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15-50LB				
GL1508	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15-50LB				
GL1509	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15-50LB				
GL1510	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15-50LB				
GL1511	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15-50LB				
GL1512	MI--03-23-16216	3/9/2023	3/9/2025	
TRUCK 15-50LB				

**Technician(s):** LUCAS SULLIVAN;ROB DUBAY;

**Signature:** ROBERT DUBAY  Digitally signed by ROBERT DUBAY

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.