



# WORK ORDER

**Great Lakes Scale**

35380 UNION LAKE RD  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586)776-0517  
 Fax: (586)776-3169  
 Fed Tax ID: 38-3206225

**Work Order#** 0035713-1  
**Request Date** 11/8/2023 7:17:00 AM  
**Schedule Date** 11/8/2023 2:00:00 PM

**Bill To** SUP0896  
 SUP0896  
 SUPERIOR MATERIALS HOLDINGS LLC  
 PO BOX 2900  
 FARMINGTON HILLS, MI, 48336  
 (248) 788-8000

**Ship To** SUP1192  
 SUP1192  
 SUPERIOR MATERIALS 4907 (36)  
 2470 AUBURN RD  
 AUBURN HILLS, MI,  
 (248) 788-8000

**Customer P.O** 6300425995  
**Requester** DAN THOMAS  
**Phone** (248) 788-8000  
**Email** dan.hargrave-thomas@vcimentos.com

**CONTRACT TYPE:** No Contract

**DESCRIPTION:**

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK - 2pm	

Service Date 11/8/2023

**2 hours minimum portal to portal**

**WORK PERFORMED**  
 Inspected tested, calibrated and sealed 3 hopper scales. Certifications to follow.  
 CALL COMPLETE

**TECHNICIAN(S)**  
 ED GORHAM, MATT PLEVA

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
38,000 x 20 LB	CBS62803A	CBS62803A	ALKON SCALE/INDICATOR / COMMAND	Agg A
38000 lbs x 20 lbs	CBS62803B	CBS62803B	ALKON SCALE/INDICATOR / COMMAND	AGG B
6700 lbs x 5 lbs	CBS62803C	CBS62803C	ALKON SCALE/INDICATOR / COMMAND	CEMENT

**PARTS LIST**

Customer Signature

\_\_\_\_\_  
 Name:

# WORK ORDER

35380 UNION LAKE RD  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586)776-0517  
 Fax: (586)776-3169  
 Fed Tax ID: 38-3206225

**Work Order#** 0035713-1  
**Request Date** 11/8/2023 7:17:00 AM  
**Schedule Date** 11/8/2023 2:00:00 PM

**Bill To** SUP0896  
 SUP0896  
 SUPERIOR MATERIALS HOLDINGS LLC  
 PO BOX 2900  
 FARMINGTON HILLS, MI, 48336  
 (248) 788-8000

**Ship To** SUP1192  
 SUP1192  
 SUPERIOR MATERIALS 4907 (36)  
 2470 AUBURN RD  
 AUBURN HILLS, MI,  
 (248) 788-8000

**Customer P.O** 6300425995  
**Requester** DAN THOMAS  
**Phone** (248) 788-8000  
**Email** dan.hargrave-thomas@vcimentos.com

**CONTRACT TYPE:** No Contract

**DESCRIPTION:**

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK - 2pm	

Service Date 11/8/2023

**2 hours minimum portal to portal**

**WORK PERFORMED**  
 Inspected tested, calibrated and sealed 3 hopper scales. Certifications to follow.  
 CALL COMPLETE

**TECHNICIAN(S)**  
 ED GORHAM, MATT PLEVA

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
38,000 x 20 LB	CBS62803A	CBS62803A	ALKON SCALE/INDICATOR / COMMAND	Agg A
38000 lbs x 20 lbs	CBS62803B	CBS62803B	ALKON SCALE/INDICATOR / COMMAND	AGG B
6700 lbs x 5 lbs	CBS62803C	CBS62803C	ALKON SCALE/INDICATOR / COMMAND	CEMENT

**PARTS LIST**

--

Customer Signature



Hoban



**Customer:** SUPERIOR MATERIALS 4907 (36)  
 SUP1192 2470 AUBURN RD

**Certificate ID:** CBS62803A231109073310

**Test Date:** 11/08/2023

AUBURN HILLS, MI

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	35
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	52
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

**Comments:**

Scale on zero, load tested, build tested and sealed.

TRACEABILITY CERTIFICATE NUMBERS:

Kit	NIST	Date	Date Due
Description			
GL1501	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			
GL1502	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			
GL1503	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			
GL1504	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			
GL1505	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			
GL1506	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			

**Technician(s):** ED GORHAM;MATT PLEVA;



**Signature:**

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.





GREAT LAKES SCALE COMPANY  
 35380 UNION LAKE RD.  
 HARRISON TOWNSHIP, MI 48045  
 Phone: (586) 776-0517

**Customer:** SUPERIOR MATERIALS 4907 (36)  
 SUP1192 2470 AUBURN RD

**Certificate ID:** CBS62803B231109073358

**Test Date:** 11/08/2023

AUBURN HILLS, MI

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	35
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	52
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

TRACEABILITY CERTIFICATE NUMBERS:			
Kit	NIST	Date	Date Due
Description			
GL1501	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			
GL1502	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			
GL1503	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			
GL1504	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			
GL1505	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			
GL1506	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			

**Comments:**  
 Scale on zero, load tested, build tested and sealed.

**Technician(s):** ED GORHAM;MATT PLEVA;

**Signature:**

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.



**Customer:** SUPERIOR MATERIALS 4907 (36)  
 SUP1192 2470 AUBURN RD

**Certificate ID:** CBS62803C231109023040

**Test Date:** 11/08/2023

AUBURN HILLS, MI

**ISO Number:** MS-FM-CEM-001

**TEST INFORMATION**

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	8.7
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	52
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

**Comments:**

Scale on zero, load tested, calibrated, retested and sealed.

TRACEABILITY CERTIFICATE NUMBERS:

Kit	NIST	Date	Date Due
Description			
GL1501	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			
GL1502	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			
GL1503	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			
GL1504	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			
GL1505	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			
GL1506	MI--03-23-16216	3/9/2023	3/9/2025
TRUCK 15 - 1000 LB			

**Technician(s):** ED GORHAM;MATT PLEVA;



**Signature:**

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.