

WORK ORDER

Great Lakes Scale

35380 UNION LAKE RD
 HARRISON TOWNSHIP, MI 48045
 Phone: (586)776-0517
 Fax: (586)776-3169
 Fed Tax ID: 38-3206225

Work Order# 0035180-1
Request Date 8/10/2023 6:44:00 AM
Schedule Date 8/10/2023 2:30:00 PM

Bill To SUP0896
 SUP0896
 SUPERIOR MATERIALS HOLDINGS LLC
 PO BOX 2900
 FARMINGTON HILLS, MI, 48336
 (248) 788-8000

Ship To SUP1193
 SUP1193
 SUPERIOR MATERIALS 4908 (38)
 901 S WAGNER RD
 ANN ARBOR, MI, 48103
 (248) 788-8000

Customer P.O 6300425996
Requester DAN THOMAS
Phone (248) 788-8000
Email dan.hargrave-thomas@vcimentos.com

CONTRACT TYPE: No Contract

DESCRIPTION:

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK - ON SITE 2:30PM	

Service Date 8/10/2023

2 hours minimum portal to portal

WORK PERFORMED
 Inspected tested and sealed 2 hopper scales. Certifications to follow.
 CALL COMPLETE

TECHNICIAN(S)
 ED GORHAM, ROB DUBAY

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
40,000 lbs X 20 lbs	102559agg	102559agg	ALKON SCALE/INDICATOR / COMMAND	
10000 lbs x 5 lbs	102559cem	102559cem	ALKON SCALE/INDICATOR / COMMAND	

PARTS LIST

--

Customer Signature

 Name:



GREAT LAKES SCALE COMPANY
 35380 UNION LAKE RD.
 HARRISON TOWNSHIP, MI 48045
 Phone: (586) 776-0517

Certificate of Calibration

Customer: SUPERIOR MATERIALS 4908 (38)
 SUP1193 901 S WAGNER RD

Certificate ID: 102559cem230815085025

Test Date: 08/10/2023

ANN ARBOR, MI 48103

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: 102559cem Capacity: 10000 lb Graduation: 5 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
08/10/2023	02/09/2024	Semi-Annually	102559cem	III
SCALE LOCATION			EQUIPMENT CONDITIONS	
			Working	Clean
SELECT NUMBER OF PM DAYS:		CAPACITY	10000 UOM	INCREMENT
			lb	5

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS						AS LEFT TEST USING TEST WEIGHTS					
Weight Applied		Reading		Error		Weight Applied		Reading		Error	
0	lb	0	lb	0	lb	0	lb	0	lb	0	lb
2000	lb	2000	lb	0	lb	2000	lb	2000	lb	0	lb
4000	lb	4000	lb	0	lb	4000	lb	4000	lb	0	lb
6000	lb	6000	lb	0	lb	6000	lb	6000	lb	0	lb
0	lb	0	lb	0	lb	0	lb	0	lb	0	lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb

AS LEFT BUILD TEST

Test Weight Applied	Material Weight Applied	Reading	Error
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb
lb	lb	lb	lb

OVERALL TEST RESULTS N/A

Customer: SUPERIOR MATERIALS 4908 (38)
 SUP1193 901 S WAGNER RD

Certificate ID: 102559cem230815085025

Test Date: 08/10/2023

ANN ARBOR, MI 48103

ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	8.7
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	82
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

Comments:

Scale on zero, load tested calibrated, retested and sealed.

TRACEABILITY CERTIFICATE NUMBERS:

Kit	NIST	Date	Date Due	Description
GL0601	MI-06-23-16373	6/6/2023	6/6/2025	1000 LB TRUCK 6 CLASS F
GL0602	MI-06-23-16373	6/6/2023	6/6/2025	1000 LB TRUCK 6 CLASS F
GL0603	MI-06-23-16373	6/6/2023	6/6/2025	1000 LB TRUCK 6 CLASS F
GL0604	MI-06-23-16373	6/6/2023	6/6/2025	1000 LB TRUCK 6 CLASS F
GL0605	MI-06-23-16373	6/6/2023	6/6/2025	1000 LB TRUCK 6 CLASS F
GL0606	MI-06-23-16373	6/6/2023	6/6/2025	1000 LB TRUCK 6 CLASS F

Technician(s): ED GORHAM;ROB DUBAY;



Signature:

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.

WORK ORDER

35380 UNION LAKE RD
 HARRISON TOWNSHIP, MI 48045
 Phone: (586)776-0517
 Fax: (586)776-3169
 Fed Tax ID: 38-3206225

Work Order# 0035180-1
Request Date 8/10/2023 6:44:00 AM
Schedule Date 8/10/2023 2:30:00 PM

Bill To SUP0896
 SUP0896
 SUPERIOR MATERIALS HOLDINGS LLC
 PO BOX 2900
 FARMINGTON HILLS, MI, 48336
 (248) 788-8000

Ship To SUP1193
 SUP1193
 SUPERIOR MATERIALS 4908 (38)
 901 S WAGNER RD
 ANN ARBOR, MI, 48103
 (248) 788-8000

Customer P.O 6300425996
Requester DAN THOMAS
Phone (248) 788-8000
Email dan.hargrave-thomas@vcimentos.com

CONTRACT TYPE: No Contract

DESCRIPTION:

TASK DESCRIPTION	SERVICE TYPE: CAL
SCALE CHECK - ON SITE 2:30PM	

Service Date 8/10/2023

2 hours minimum portal to portal

WORK PERFORMED
 Inspected tested and sealed 2 hopper scales. Certifications to follow.
 CALL COMPLETE

TECHNICIAN(S)
 ED GORHAM, ROB DUBAY

EQUIPMENT SERVICED	ASSET/UNIT #	SERIAL #	MODEL	LOCATION
40,000 lbs X 20 lbs	102559agg	102559agg	ALKON SCALE/INDICATOR / COMMAND	
10000 lbs x 5 lbs	102559cem	102559cem	ALKON SCALE/INDICATOR / COMMAND	

PARTS LIST

--

Customer Signature



Colin



GREAT LAKES SCALE COMPANY
 35380 UNION LAKE RD.
 HARRISON TOWNSHIP, MI 48045
 Phone: (586) 776-0517

Certificate of Calibration

Customer: SUPERIOR MATERIALS 4908 (38)
 SUP1193 901 S WAGNER RD

Certificate ID: 102559agg230815085138

Test Date: 08/10/2023

ANN ARBOR, MI 48103

ISO Number: MS-FM-CEM-001

(*)Make: ALKON Model : COMMAND SN#: 102559agg Capacity: 40000 lb Graduation: 20 lb

CALIBRATION DATE	DUE DATE	TEST INTERVAL	SCALE ID	SCALE CLASS
08/10/2023	02/09/2024	Semi-Annually	102559agg	III L
SCALE LOCATION		EQUIPMENT CONDITIONS		
		Working	Clean	
SELECT NUMBER OF PM DAYS:		CAPACITY	40000 UOM lb	INCREMENT 20

AS FOUND TEST USING TEST WEIGHTS

AS LEFT TEST USING TEST WEIGHTS

AS FOUND TEST USING TEST WEIGHTS						AS LEFT TEST USING TEST WEIGHTS					
Weight Applied		Reading		Error		Weight Applied		Reading		Error	
0	lb	0	lb	0	lb	0	lb	0	lb	0	lb
2000	lb	2000	lb	0	lb	2000	lb	2000	lb	0	lb
4000	lb	4000	lb	0	lb	4000	lb	4000	lb	0	lb
6000	lb	6000	lb	0	lb	6000	lb	6000	lb	0	lb
0	lb	0	lb	0	lb	0	lb	0	lb	0	lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb
	lb		lb		lb		lb		lb		lb

AS LEFT BUILD TEST

Test Weight Applied		Material Weight Applied		Reading		Error	
6000	lb	6000	lb	12000	lb	0	lb
6000	lb	10000	lb	16000	lb	0	lb
6000	lb	20000	lb	26000	lb	0	lb
6000	lb	30000	lb	36000	lb	0	lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb
	lb		lb		lb		lb

OVERALL TEST RESULTS PASS

Customer: SUPERIOR MATERIALS 4908 (38)
 SUP1193 901 S WAGNER RD

Certificate ID: 102559agg230815085138

Test Date: 08/10/2023

ANN ARBOR, MI 48103

ISO Number: MS-FM-CEM-001

TEST INFORMATION

TEST WEIGHT CLASSIFICATION	EXPANDED UNCERTAINTY
F	35
ENVIRONMENTAL CONDITIONS	TEMPERATURE
Acceptable	82
WAS THE SCALE WITHIN CUSTOMERS REQUIRED ACCURACY?	
Yes	
TEST LOCATION	OVERALL RESULT
Onsite	Pass w/o adjustment

Comments:

Scale on zero, load tested, build tested and sealed

TRACEABILITY CERTIFICATE NUMBERS:

Kit	NIST	Date	Date Due	Description
GL0601	MI-06-23-16373	6/6/2023	6/6/2025	1000 LB TRUCK 6 CLASS F
GL0602	MI-06-23-16373	6/6/2023	6/6/2025	1000 LB TRUCK 6 CLASS F
GL0603	MI-06-23-16373	6/6/2023	6/6/2025	1000 LB TRUCK 6 CLASS F
GL0604	MI-06-23-16373	6/6/2023	6/6/2025	1000 LB TRUCK 6 CLASS F
GL0605	MI-06-23-16373	6/6/2023	6/6/2025	1000 LB TRUCK 6 CLASS F
GL0606	MI-06-23-16373	6/6/2023	6/6/2025	1000 LB TRUCK 6 CLASS F

Technician(s): ED GORHAM;ROB DUBAY;



Signature:

Devices were calibrated with the certified test weights. Adjustments made to restore and/or maintain the accuracy of the device conform to the tolerances established by NIST or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%.